


The following article is hosted by PTB https://doi.org/10.7795/810.20250619
It is provided for personal use only.

5th International DCC-Conference

2025-02-25 to 2025-02-27 Proceedings

DOI: https://doi.org/10.7795/810.20250619





5th International DCC-Conference

2025-02-25 to 2025-02-27

Proceedings

DOI: https://doi.org/10.7795/810.20250619

Table of contents

Table of contents1	
Chairs of the Program Committee4	Ļ
Members of the International Programme Committee4	ļ
Agenda and Chairpersons4	ļ
Day 1 2025-02-25 (Tuesday) Plenary Session (Chair: Martin Koval)7	,
Welcome address of the president of PTB, Prof. Dr. Cornelia Denz, Germany7 P01 Interoperable conformity data exchange in a digital quality infrastructure - update from UNECE	5
Day 1 (2025-02-25 Tuesday) Parallel Session "Key Aspects of DCC I" (Chair:	
Shanna Schönhals)45	;
001 Harmonisation of DCCs in DKD - an update45 002 Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability	
003 Implementation of Digital Calibration Certificates for Force According to	
ISO 37663 004A Taxonomy of Radiometric-Measurand Metadata72	
005A Matter of Trust85	
Day 1 (2025-02-25 Tuesday) Parallel Session "Applications NMI I" (Chair:	
Brett Hyland)93	}
006 Concepts and frameworks fostering interoperability within a digital quality infrastructure: DCC development in Colombia and QI infrastructure interoperability under an Enterprise Architecture Framework)
009 Generating Mass DCC based on DKD expert report using a spreadsheet 117 010 Real-World Applications of Digital Calibration Certificates (DCC): Use Cases and Implementation Stories	7



Day 2 (202	25-02-26 Wednesday) Plenary Session (Chair: Hugo Gasca)138
P04	Advances in the DCC enablement at Beamex
Day 2 (20)	25-02-26 Wednesday) Parallel Session "Tools & Software" (Chair: Anjali
Sha	arma)
Inte 012 GEI	ital Calibration Exchange (DCX): Software Tools, Interfaces, and Workflow egration
013 For Frai 014 Cor	ol
	ramework for digitising dimensional metrology226
• ,	25-02-26 Wednesday) Parallel Session "Industry Applications" (Chair:
Car	los Galvan)239
for 9 017 Aut	tructions for using the DCC scheme to create a digital calibration certificate gauge blocks239 omation in calibration process using DCC, primary to secondary labs254 mantic registration of DCC and DCoC onto Information Model of
Inte	ernational Data Spaces
020 Pro	gress Towards Producing DCC with Standard Structured Measurand prmation282
• •	25-02-26 Wednesday) Parallel Session "Applications NMI II" (Chair: Jariya ajarern)283
022Trai 023A se 024Pro 025Pra	ital Calibration Certificates for Radioactivity and FAIR Data Exchange283 nsport of DCC with cryptographic signature in PDF/A
Day 3 (202	25-02-27 Thursday) Plenary Session (Chair: Hector Laiz)
P07Mod to A	Motivation for using DCC



Day 3 (2025-02-27 Thursday) Parallel Session: Schema & Semantics (chair:	Robert
Hanisch)	373
026Augmenting the DCC Framework with the DSI Ontology: Enabling Sema	
Web Integration in Calibration Data Management	
Presentation of Muhammed-Ali Demir	
028An In-Depth Look at the DCX Schema	
029 Compatible sub schemata of the DCC Schema	
Day 3 (2025-02-27 Thursday) Parallel Session "Key Aspects of DCC II" (Chair	: Mark
Kuster)	406
030 Calibration and metrological traceability	406
031 Traceability chains using the DCC	
032 Applying FAIR Principles to Quality Infrastructure Data for Economic Analysis Approaching Challenges in Data Availability, Accessibility, and Interoper	alysis - ability
033A new data structure for improved scientific data reporting in a FAIR con manner	npliant
Day 3 (2025-02-27 Thursday) Parallel Session "DCR & DRMC" (Chair: Alexis '	Valqui)
	444
035 Implementation of Digital Calibration Request, DCR	444
036Benefits of developing DCR with a focus on machine readability and	
automation.	
037 The Digital Reference Material Document Derived from the DCC	
Closing Session & Statistical information:	483



Chairs of the Program Committee

- Carlos Galván, Centro Nacional de Metrologia CENAM, Mexico
- Shanna Schönhals, PTB Physikalisch-Technische Bundesanstalt; Germany (Vice-Chair)

Members of the International Programme Committee

The preparation of the scientific programme and review of submitted abstracts was carried out by the International Programme Committee (IPC). The IPC combines a broad knowledge and commitment in digitalisation to ensure a high-quality event. Members of the IPC are:

- David Balslev-Harder, Team Leader at Danish Fundamental Metrology (DFM),
 Denmark
- Jariya Buajarem, Head of Digital Centre and Materials Metrology, Quality System and International Cooperation Office, NIMT, Thailand
- Carlos Galván, Dirección Metrología Dimensional, CENAM, Mexico
- Hugo Gasca Aragon, Co-Chair of IMEKO TC6 Digitalization
- Blair Hall, Distinguished Scientist at Measurement Standards Laboratory of New Zealand (MSL)
- Robert Hanisch, Director of the Office of Data and Informatics (NIST), USA
- Brett Hyland, Project lead at UN/CEFACT and NATA Australia
- Martin Koval, Chair of the FORUM-MD ad hoc Task Group on Harmonizing DCC and DRMC, Czech Metrology Institute (CMI), Czech Republic
- Thomas Krah, CEO of German Calibration Service DKD, Germany
- Mark Kuster, Chair of "Metrology Practices" and "MII & Automation Committees" at NCSLI, USA
- · Hector Laiz, Director of Metrology and Quality, INTI, Argentina
- Kim Nguyen, Assessor of the Board of D-Trust GmbH, Germany
- Anjali Sharma, Senior Principal Scientist and Head, NPLI, India
- Shanna Schönhals, Group Lead and Head of DCC Development Team at PTB, Germany
- Alexis Valqui, Digital Transformation Consultant, Peru

Back to "Table of Contents" above

Agenda and Chairpersons

2025-02-25 (Tuesday)

UTC	2025-02-25 (Tuesday) Plenary Session (Chair: Martin Koval)
12:00	Welcome and Key Notes (Martin Koval, Cornelia Denz, Cui Shan)
12:30	#P01 Brett Hyland - Interoperable conformity data exchange in a digital quality infrastructure - update from UNECE
13:00	#P02 Ciro Alberto Sánchez - Advancing Digital Calibration Certificates: The INM's Role in Modernizing Metrological Services in Latin America



13:30	Break		
	Key Aspects of DCC I	Applications NMI I	
	(Chair: Shanna Schönhals)	(Chair: Brett Hyland)	
14:00	#001 Thomas Krah	#006 Rodolfo Manuel Gomez Rodriguez Concepts and frameworks fostering interoperability within a digital quality	
1 14:00		infrastructure: DCC development in Colombia and QI infrastructure interoperability under an Enterprise Architecture Framework	
14:20	#002 Julian Haller Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability	#007 Ryu Dong-Hun Calibration support system development of KTL (Korea Testing Laboratory) using certificate database established by OCR	
14:40	#003 Shanna Schönhals Implementation of Digital Calibration Certificates for Force Calibration According to ISO 376	#008 Praiya Thongluang The Web-Based Digital Calibration Certificate Generation Platform developed by NIMT	
15:00	#004 Mark Kuster A Taxonomy of Radiometric-Measurand Metadata	#009 Itzel Dominguez Generating Mass DCC based on DKD expert report using a spreadsheet	
15:20	#005 Andrew Freund A Matter of Trust.	#010 Yoshitha Weerasinghe Real-World Applications of Digital Calibration Certificates (DCC): Use Cases and Implementation Stories	

Back to "Table of Contents" above

2025-02-26 (Wednesday)

UTC	2025-02-26 (Wednesday) Plenary Session (Chair: Hugo Gasca)		
12:00	#P03 Tuukka Mu	stapää - Advances in the DCC enablen	nent at Beamex
12:30	#P04 Gean Marcos Gerony	mo - Python-based DCC Tools Develop	oment and Implementation
13:00	#P05	Thomas Engel - DCC in production mo	ode
13:30		Break	
	Tools & Software (Chair: Anjali Sharma)	Industry Applications (Chair: Carlos Galvan)	Applications NMI II (Chair: Jariya Buajarern)
14:00	#011 David Balslev-Harder Getting Started with Digital Calibration Exchange (DCX): Software Tools, Interfaces, and Workflow Integration	#016 Christian Sander Instructions for using the DCC scheme to create a digital calibration certificate for gauge blocks	#021 Henrike Fleischhack Digital Calibration Certificates for Radioactivity and FAIR Data Exchange
14:20	#012 Alexander Reissert GEMIMEG 2.0: Evolution of the concept and implementation of the GEMIMEG Tool	#017 Aldo Adrián García González Automation in calibration process using DCC, primary to secondary labs.	#022 Diego Coppa Transport of DCC with cryptographic signature in PDF/A
14:40	#013 Justin Jagieniak Formula Evaluation and MathML Processing in Python: A High-Precision Framework for XML documents like the DCC	#018 Tomasz Sołtysiński Semantic registration of DCC and DCoC onto Information Model of International Data Spaces	#023 Blair Hall A service to authenticate documents issued by MSL



15:00	#014 Benedikt Seeger Consistent Representation and Usage of Tabulated Data in the DCC. Enable Easy-To-Use Computer-Aided Processing and Visualization of Your Data	#019 Lisa Busser DCC in a Mass-Calibration Software	#024 Beste Korutlu Progressive Shift towards Digitalization in TUBITAK UME Mass Laboratory
15:20	#015 Arif Sanjid A framework for digitising dimensional metrology	#020 Jens Leinhos Progress Towards Producing DCC with Standard Structured Measurand Information	#025 Søren Kynde Practical Applications of Digital Calibration Exchange (DCX) covering Certificates, Requests, and Templates

Back to "Table of Contents" above

2025-02-27 (Thursday)

UTC	2025-02-27 (Thursday) Plenary Session (Chair: Hektor Laiz)		
12:00		P06 Martin Koval - Motivation for usin	
12:30		ng Digital Calibration Certificates Usir ess Heterogeneity and Harness FAIR	
13:00	#P08 Sajiv	/ Perera - DCC Sharing within a Gove	ernment Agency
13:30		Break	
	Schema & Semantics (Chair: Robert Hanisch)	Key Aspects of DCC II (Chair: Mark Kuster)	DCR & DRMC (Chair: Alexis Valqui)
14:00	#026 Moritz Jordan Augmenting the DCC Framework with the DSI Ontology: Enabling Semantic Web Integration in Calibration Data Management	#030 Blair Hall Calibration and metrological traceability	#035 Heidi Foldal Implementation of Digital Calibration Request, DCR
14:20	#027 Muhammed-Ali Demir The latest version of the DCC	#031 Christian Müller-Schöll Traceability chains using the DCC	#036 Martin Cradock Østerlund Benefits of developing DCR with a focus on machine readability and automation.
14:40	#028 David Balslev-Harder An In-Depth Look at the DCX Schema	#032 Ulrich Harmes-Liedtke Applying FAIR Principles to Quality Infrastructure Data for Economic Analysis - Approaching Challenges in Data Availability, Accessibility, and Interoperability	#037Michael Melzer The Digital Reference Material Document Derived from the DCC
15:00	#029 Muhammed-Ali Demir Compatible sub schemata of the DCC Schema	#033 Jon Gabirondo-López A new data structure for improved scientific data reporting a FAIR compliant manner	#038 Dinis Camara The Digital NIST: Update on the Digital Transformation of Reference Materials
15:20	701 : 9		Open Exchange on DCR and DRMC
15:40	Closing Session		

Back to "Table of Contents" above



Day 1 2025-02-25 (Tuesday) Plenary Session (Chair: Martin Koval)

Welcome address of the president of PTB, Prof. Dr. Cornelia Denz, Germany

Dear representatives of the National Metrology Institutions, dear speakers of the 5th DCC conference, dear attendees, it is a great pleasure welcoming you to the 5th International Conference on the Digital Calibration Certificates - DCC number 5. I am happy that you have followed the call to the digital DCC conference in such a large number.

In previous years, I have referred to the corresponding number of our DCC conference series in my welcome speech. On the occasion the 3rd DCC conference, I referred to the German saying *Aller guten Dinge sind Drei* or *All great things come in three*. On the occasion of the 4th DCC conference, I referred to a compass with its four celestial directions.

And I would like to continue this today. With the 5th edition of the DCC conference, we celebrate a special birthday.

For most children, reaching the age of five means starting school and increasing their independence. This is why the 5th birthday is often considered a milestone.

This is without any doubt also true for the 5th DCC conference, especially with respect to the number of participants and their world-wide representation.

The number five is often accompanied – saying it in metrological terms – by the prefix Big. In Africa, the Big Five represent royal animals as the lion, the leopard, the rhinoceros, the elephant, and the buffalo. The term was originally coined by hunters to refer to the five most difficult animals in Africa to hunt. Nowadays it is widely used by safari tourists for spotting these animals. The Big Five animals of Africa are also famous since they appear on South African Rand banknotes.

Trait theory is another field referring to the role of number five. The Big Five model describes human personality in five core dimensions:

openness, conscientiousness, extraversion, agreeableness and neuroticism.

The model is often referred to as the OCEAN model due to the initial letters of the English names of the five trait dimensions.

Extraversion is sociability, being bold and energetic; agreeableness is kindness, being sympathetic or cooperative; openness to experience, intellect, or imagination includes creativity and intrigue; conscientiousness is thoughtfulness in the sense of efficiency and being organized; neuroticism refers to emotional stress.

The rational of the model is the assumption that a personality is made up of varying degrees of the five core elements. Specific personality profiles are then created from the individual distributions.

Is it possible to transfer the idea of the Big Five traits to the DCC? To do so, I'd like to take a closer look at the successes, challenges and developments of DCC since the last conference and compare them to the five personality profiles.

First, considering the challenges of the DCC in the near future, one definitely finds the trait of openness. A high degree of openness shows a thirst for knowledge, a willingness to experiment and to tackle new aspects with imagination, interest, and joy of abstract thinking.

The harmonization as well as the variability of DCC schemas require precisely these aspects of openness. Openness to different formats for example that can be exchanged or transferred from one to another in a variable way is important for the future of the DCC.



This includes openness to harmonization of various metrological disciplines as well as harmonization of certificates for different purposes. For example, harmonization will be important between DCC and DCR, the digital calibration request, between DCC and DPP, the digital product pass, or between DCC and DRMC, the digital reference material certificate. The challenging future question is, if these certificates can be directly derived or adapted from the DCC schema, or if it needs to be schematically different.

I am absolutely confident that there is world-wide a very high level of openness and many ideas how to face this challenge. Let me name some examples.

In September 2024, the metrological community was meeting at the premier event in measurement science and applications in Hamburg. The IMEKO World Congress brought together more than 1.000 metrological minds, and digitalization in metrology and the DCC played a prominent role in many sessions and workshops.

A timely workshop was held in conjunction with the Conference International de Poids et Measure, CIPM, and its newly founded horizontal structure Metrology and Digitalization. It was focused on basics of metrological traceability in the digital era, especially traceability in today's digital measurement infrastructure, and subsequently addressed applications such as digital twins, virtual data spaces, sensor networks and the DCC, all with respect to traceability *and also* provenance of data.

Furthermore, due to the fact that digital calibration certificates are developed both by national metrology institutes and calibration laboratories worldwide, a workshop dedicated solely to the DCC took also place at the conference.

It provided a deep insight into the international status of the DCC development. It combined short pitches with news from different Regional Metrology Organizations, and project results of the EURAMET Technical Committee TC-IM 1448. The newly established CIPM Task Group on Harmonization of DCC and DRMC, one of seven task and working groups of the Forum Metrology and Digitalization of CIPM, was also introduced and contributed user success stories.

These examples show that the DCC is also a social animal, that loves to gain energy by cooperation - a strong characteristic of extraversion. I assume that the DCC is highly scoring in extraversion as it is likely to start more and more effective interactions, allowing the DCC to attract a wider circle of interested stakeholders.

Furthermore, the DCC is highly scoring in agreeableness because it tends to be cooperative and includes attributes such as FAIRness. Being findable, accessible, interoperable and reusable can be interpreted as features for common understanding, contributing to the harmonization – and with it the satisfaction of different stakeholders with the DCC.

This can also be associated with the dissemination of the DCC. In the digital transformation era, new opportunities emerge almost every day in trade and industry through digital data traffic. This upheaval is challenging for the quality infrastructure, consisting of metrology, standardisation, accreditation, conformity assessment and market surveillance.

Here, the DCC can contribute digital traceability, provenance as well as FAIR data quality as a highly relevant profile for quality infrastructure and industry: it enables the seamless integration of calibration data into industrial processes and ensures a greater transparency - for more quality, productivity and lower costs.

One of the companies that collaborates with PTB on this avenue, is Testo Industrial Services, which relies on the digital calibration certificate and encourages its customers to digitize. For this purpose, a best practice example for a caliper gauge was established together with PTB. Also, the implementation of the DCC by large pharma industry companies as Boehringer Ingelheim shows, how important the DCC is for different industrial sectors.



In order to establish secure and robust calibrated measurement systems for the digital transformation, it is also important to develop internationally agreed basic standards to ensure reliable communication of digital information and certificates in the world-wide quality infrastructure processes.

An important standard is the International System of units in the digital version - the SI digital reference point. It has now been fully established by the international Metrology at BIPM. It is a tremendous enabler of the digital transformation in metrology and gives the DCC the right base for traceability of the measurements behind the certificates.

At the same time, digitalization has implications for development cooperation. For example, digital transformation can lead to a mitigation of existing economic or social disparities for developing countries and allows them to connect to the digital services of the quality infrastructure as DCC.

The fourth of the five personality traits I like to connect to the DCC is the way in which negative emotions are dealt with. And that is Neuroticism.

In such a dynamic and at the same time protracted process as digitalization and the necessary interaction of different stakeholders, with different interests, priorities and sensitivities, it would be perfectly understandable that worries about different aspects of digitalization are appearing. However, I experience that neuroticism is nearly not existing in the community around the DCC. This is on the one hand because it actually leads to many advantages, to efficiency, stability, and resilience, but also because the community is discussing together important developments of the DCC in conferences like today and come quickly to common grounds.

This is also true within today's conference program: Advances in machine-actionable format, data tailored to the quality infrastructure, industrial implementations and convincing use cases and success stories of digital certificates, and solutions enabling uptake of FAIR data across systems and processes, fostering interoperability within a digital quality infrastructure are all hot topics that will be will be presented and discussed in different sessions throughout the three days of the conference.

Last but not least there is conscientiousness as the fifth personality trait. It is defined by high levels of thoughtfulness, and goal-directed behaviors. This is not only fully true for the DCC but also for this conference.

In this spirit, I would like to take the opportunity to express my sincere gratitude to the DCC team at PTB, head by the powerful trio of Wiebke Heeren, Franziska Decker and Shanna Schönhals, who did a tremendous job organizing this large conference, together with the 15 active and engaged members of the International Program Committee. Please give them a very warm digital applause.

I want to close by wishing you a great conference with the big five: inspiring talks, intensive discussions, many interactions, expanding networks, and innovative, forward-looking ideas – all for versatile digital certificates of tomorrow.

Thank you.



Back to "Table of Contents" above

Presentation of Martin Koval

5th International DCC Conference DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

Martin Koval CIPM - FORUM-MD-TG-H-DCC/DRMC (chair)

14 Members of the International Programme Committee (IPC)

David Balslev-Harder, Team Leader at Dansk Fundamental Metrologi (DFM), Denmark

Jariya Buajarern, Head of Digital Center and Materials Metrology, Quality System and International Cooperation Office, NIMT, Thailand

Carlos Galván, Dirección Metrología Dimensional, CENAM, Mexico

Hugo Gasca Aragon, Co-Chair of IMEKO TC6 Digitalization

Blair Hall, Distinguished Scientist at Measurement Standards Laboratory of New Zealand (MSL)

Anjali Sharma, Senior Principal Scientist and Head, NPLI, India

Robert Hanisch, Director of the Office of Data and Informatics (NIST), USA

Brett Hyland, Project lead at UN/CEFACT and NATA Australia

Martin Koval, Chair of the FORUM-MD *ad hoc* Task Group on Harmonizing DCC and DRMC, Czech Metrology Institute (CMI), Czech Republic

Mark Kuster, Chair of "Metrology Practices" and "MII & Automation Committees" at NCSLI, USA

Hector Laiz, Director of Metrology and Quality, INTI, Argentina

Kim Nguyen, Assessor of the Board of D-Trust GmbH, Germany

Shanna Schönhals, Group Lead and Head of DCC Development Team at PTB, Germany

Alexis Valqui, Digital Transformation Consultant, Peru

5TH INTERNATIONAL DCC CONFERENCE



Chair and Co-Chair of the International Programme Committee (IPC)

· Chair: Martin Koval, CMI

· Co-Chair: Shanna Schönhals, PTB

5TH INTERNATIONAL DCC CONFERENCE

Key Notes



Cornelia Denz, President of PTB and Chair of FORUM-MD



Cui Shan, NMC A*Star, Chairof APMP Focus Group on Digital Transformation

5711 INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE



Plenary sessions

Plenary Session	DAY 1	DAY 2	DAY 3
12:00	Welcome and Key Notes (Martin Koval, Cornelia Denz, Cui Shan)	#P03 Tuukka Mustapää Advances in the DCC enablement at Beamex	#P06 Martin Koval Motivation for using DCC
12:30	#P01 Brett Hyland Interoperable conformity data exchange in a digital quality infrastructure - update from UNECE	#P04 Gean Marcos Geronymo Python-based DCC Tools Development and Implementation	#P07 Anjali Sharma Modeling Digital Calibration Certificates Using Object-Oriented Technologies to Address Heterogeneity and Harness FAIR Principles
13:00	#P02 Ciro Alberto Sánchez Advancing Digital Calibration Certificates: The INM's Role in Modernizing Metrological Services in Latin America	#P05 Thomas Engel DCC in production mode	#P08 Sajiv Perera DCC Sharing within a Government Agency
13:30 - 14:00	Break		

5th INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

DAY 1

Parallel Sessions	Key Aspects of DCC I	Applications NMI I	
14:00	#001 Thomas Krah Harmonisation of DCCs in DKD – an update	#006 Rodolfo Manuel Gomez Rodriguez Concepts and frameworks fostering interoperability within a digital quality infrastructure: DCC development in Colombia and QI infrastructure interoperability under an Enterprise Architecture Framework	
14:20	#002 Julian Haller Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability	#007 Ryu Dong -Hun Calibration support system development of KTL(Korea Testing Laboratory) using certificate database established by OCR	
14:40	#003 Shanna Schönhals Implementation of Digital Calibration Certificates for Force Calibration According to ISO 376	#008 Praiya Thongluang The Web-Based Digital Calibration Certificate Generation Platform developed by NIMT	
15:00	#004 Mark Kuster A Taxonomy of Radiometric -Measurand Metadata	#009 Itzel Dominguez Generating Mass DCC based on DKD expert report using a spreadshee	
15:20	#005 Andrew Freund A Matter of Trust.	#010 Yoshitha Weerasinghe Real-World Applications of Digital Calibration Certificates (DCC): Use Cases and Implementation Stories	

511 INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE



DAY 2

Parallel Sessions	Tools & Sofware	Industry Applications	Applications NMI II
14:00	#011 David Balslev-Harder Getting Started with Digital Calibration Exchange (DCX): Software Tools, Interfaces, and Workflow Integration	#016 Christian Sander Instructions for using the DCC scheme to create a digital calibration certificate for gauge blocks	#021 Henrike Fleischhack Digital Calibration Certificates for Radioactivity and FAIR Data Exchange
14:20	#012 Alexander Reissert GEMIMEG 2.0: Evolution of the concept and implementation of the GEMIMEG Tool	#017 Aldo Adrián García González Automation in calibration process using DCC , primary to secondary labs.	#022 Diego Coppa Transport of DCC with cryptographic signature in PDF/A
14:40	#013 Justin Jagieniak Formula Evaluation and MathML Processing in Python: A High-Precision Framework for XML documents like the DCC	#018 Tomasz Sottysiński Semantic registration of DCC and DCoC onto Information Model of International Data Spaces	#023 Blair Hall A service to authenticate documents issued by MSL
15:00	#014 Benedikt Seeger Consistent Representation and Usage of Tabulated Data in the DCC. Enable Easy-To-Use Computer-Aided Processing and Visualization of Your Data	#019 Lisa Busser DCC in a Mass-Calibration Software	#024 Beste Korutlu Progressive Shift towards Digitalization in TUBITAK UME Mass Laboratory
15:20	#015 Arif Sanjid A framework for digitising dimensional metrology	#020 Jens Leinhos Progress Towards Producing DCC with Standard Structured Measurand Information	#025 Soeren Kynde Practical Applications of Digital Calibration Exchange (DCX) covering Certificates, Requests, and Templates

5TH INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

DAY 3

Parallel Sessions	Schema & Semantics	Key Aspects of DCC II	DCR & DRMC
14:00	#026 Moritz Jordan Augmenting the DCC Framework with the DSI Ontology: Enabling Semantic Web Integration in Calibration Data Management	#030 Blair Hall Calibration and metrological traceability	#035 Heidi Foldal Implementation of Digital Calibration Request, DCR
14:20	#027 Muhammed-Ali Demir The latest version of the DCC schema V3.3.0	#031 Christian MüllerSchöll Traceability chains using the DCC	#036 Martin Cradock Østerlund Benefits of developing DCR with a focus on machine readability and automation.
14:40	#028 David Balslev-Harder An In-Depth Look at the DCX Schema	#032 Ulrich Harmes-Liedtke Applying FAIR Principles to Quality Infrastructure Data for Economic Analysis - Approaching Challenges in Data Availability, Accessibility, and Interoperability	#037Michael Melzer The Digital Reference Material Document Derived from the DCC
15:00	#029 Muhammed-Ali Demir Compatible sub schemata of the DCC Schema	#033 Jon Gabirondo-López A new data structure for improved scientific data reporting in a FAIR compliant manner	#038 Dinis Camara The Digital NIST: Update on the Digital Transformation of Reference Materials
15:20	Open Exchange on Schema and Semantics		Open Exchange on DCR and DRMC

5th INTERNATIONAL DCC CONFERENCE
DCG FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE



Conference Chairs

Session	Chairs
DAY 1	
Plenary Session	Martin Koval
Key Aspects of DCC I	Shanna Schönhals
Applications NMII	Brett Hyland
DAY 2	
Plenary Session	Hugo Gasca
Tools & Sofware	Anjali Sharma
Industry Applications	Carlos Galvan
Applications NMIII	Jariya Buajarern
DAY 3	
Plenary Session	Hector Laiz
Schema & Semantics	Robert Hanisch
Key Aspects of DCC II	Mark Kuster
DCR & DRMC	Alexis Valqui

5th INTERNATIONAL DCC CONFERENCE C FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

Conference Team

Oksana Baer Franziska Decker Lutz Doering Muhammed Ali Demir Wiebke Heeren Justin Jagieniak Moritz Jordan Shan Lin

Jan Loewe



In case of technical questions or issues please contact the team at $\underline{dcc\text{-}conference\text{-}2025@ptb.de}$

5TH INTERNATIONAL DCC CONFERENCE



Tready Wednesday Wed

Conference Agenda

5TH INTERNATIONAL DCC CONFERENCE DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

Back to "Table of Contents" above

P01 Interoperable conformity data exchange in a digital quality infrastructure - update from UNECE

Presenting author: Brett Hyland, Project Lead at UN/CEFACT, NATA Australia

E-mail address: Brett.hyland@nata.com.au

Additional authors: Nil

Abstract

Conformity data related to traded goods differs in quite fundamental ways compared with calibration and other equipment assurance data, yet there are potential payoffs if a degree of interoperability could be achieved. The United Nations Economic Commission for Europe (UNECE) has presented for public review a draft policy recommendation for governments, UNECE Recommendation 49 - Transparency at Scale, which outlines the need for interoperable and platform-independent streams of information for substantiating claims about product sustainability attributes. A central element of this recommendation is a conformity credential, designed to digitally capture basic product conformity information within a hosted file that also links to the underlying conformity certificate (which may be in analogue, digital or hybrid form and may also be encrypted). Numerous implementations of this general approach are proceeding under a specification known as the United Nations Transparency Protocol, in which the credential file is cryptographically verifiable back to its issuing body in a tamperevident manner and includes provision for referencing relevant endorsement credentials (such as those issued by accreditation bodies) in either digital or analogue form. It is hoped that increased awareness of these developments may facilitate discussion regarding digitalisation matters between the scientific and trade measurement community and the wider conformity community.



Back to "Table of Contents" above

Presentation of Brett Hyland



Update from UNECE

25 February 2025

Brett Hyland
NATA Australia
& UN/CEFACT Project Lead






UNECE Recommendation 49 Public Review

A new **United Nations Economic Commission for Europe (UNECE)**Recommendation is now **OPEN** for Public Review:

Recommendation No. 49 Transparency at Scale



- This a draft recommendation to member states, promoting the value of transparent supply chain data exchange using international and interoperable standards and based on an open architecture
- To be considered for adoption at the July 2025 UN/CEFACT Plenary








Reducing costs, increasing participation

To encourage maximum inclusion, digital trade systems cannot favour rich countries or create unjustified barriers for small businesses.

This means:

- adopting internationally accepted approaches, emphasising open development and free resources
- adopting an open architecture, so that stand-alone technology platforms do not add friction to data exchange or be leveraged to disadvantage certain participants
- providing a backbone of trust to supply chain data and making the trust anchors verifiable by any participant - but leaving decisions regarding the acceptability of specific trust anchors up to the buyers.



Perhaps we should start at the beginning

- why is the UN interested in supply chain transparency?





The problem of greenwashing

United Nations Sustainable Development Goals depend on the reliability of claims being made about product sustainability.

A race to the top: A race to the bottom: Even well-intentioned It's easy It's hard Consumer businesses must fake confidence to fake to fake claims to survive. improves claims! claims! Consumer Low confidence Higher Business is confidence drops means no price motivated to prices are differential justified make provable claims

Digitalisation can help shift the balance, favouring a race to the top



Transparency at Scale

UNECE Recommendation 49 'Transparency at Scale' is about finding equitable and scalable ways for discovering and verifying information that supports a claim made about a product.

But does it have anything to do with Metrology? *Potentially 'yes'* if product claims are based on measurement and where access to a testing laboratory's certificates (DCC/DRMC/...) may become relevant. Or if a certified cargo quantity is included in the trade data.









A completely different question (but possibly still interesting) is whether entire chains of measurement traceability might be constructed using a similar architecture, noting some common ground with trade data:

- A common need to link data with physical objects and a potential role for product passports (for instrument/artefact/material as well as for traded goods)
- A common need for verifiable access to certificates/reports
- Both involve certificates/reports that are subject to revision/withdrawal
- Common assessment/accreditation frameworks exist
- Data confidentiality may not suit publicly accessible registers





UN/CEFACT

The United Nations Centre for Trade Facilitation and eBusiness (UN/CEFACT) is an intergovernmental body, established in 1966, which aims to deliver simple, transparent and effective processes for global commerce.

In particular, UN/CEFACT facilitates national and international transactions, through the simplification/harmonization of processes and information flows.

One major area of focus:

Digital exchange of trade data that supports product sustainability claims.







Some data challenges to be addressed

Making data trustworthy!

Some suppliers exploit non-verifiable data points to give non-compliant products the appearance of compliance.

Interoperablility!

Closed (non-interoperable) data exchange systems mean that many participants adopt a 'wait and see' approach to digitalisation, due to the dependency on other supply chain actors for enabling data flow.

Accessibility vs confidentiality!

There is tension between requirements for disclosure and protection of commercially sensitive data.






Some milestones

<u>June 2022</u> – a workstream was created for the development of a standard for conformity data exchange/access


<u>July 2024</u> - a Business Requirements Specification was published titled: Digital Product Conformity Certificate Exchange – High Level Process

October 2024 – The United Nations Transparency Protocol (UNTP) released as a specification for testing

<u>February 2025</u> - Public review version of UNECE Rec#49 lists UNTP as a representation of relevant principles <u>uncefact.unece.org/display/uncefactpublicreview</u>








Let's take a quick dive into UNTP

A framework for certificate discovery and verification




Support for issued certificates within UNTP


Digitalisation of certificates is not within the scope of this work

The certificate - whether analogue, digital or hybrid - remains the key reference for conformity data and so the focus is providing a surrounding framework offering:


- 1. A means of discovering relevant certificates from products of interest in a tamper-evident & machine-readable manner.
- 2. Verifiable connections between data points, including any trust anchors like testing/certification bodies, accreditation bodies and regulators.
- 3. Compatibility with confidentiality requirements of supply chain actors (including third party selective redaction).













Interoperability protocol - not a platform

Leave data where it is - but link it together when needed Use any software you like - so long as it conforms to UNTP



Thousands of platforms, millions of value-chains, billions of transactions







UN / CEFACT


Resources


All UNTP resources are available without charge and are royalty free https://uncefact.github.io

 UNTP embraces relevant international protocols, including the open source and free World Wide Web Consortium (W3C) protocols for Decentralised Identifiers (DID), Verifiable Credentials and compatible alternatives


Industry adoptions of UNTP can be registered, either as an 'implementation' (for selected UNTP elements) or as an 'extension' (with extensions owned and managed by the party undertaking the extension).

https://uncefact.github.io/spec-untp/docs/implementations https://uncefact.github.io/spec-untp/docs/extensions











Exchange of ideas

It would be very nice to leverage existing communication channels (formal and informal) to exchange ideas between the trade data world and the world of metrology, noting that:

- Some laboratories are involved in providing test results for traded goods while also undertaking metrology activities and, similarly, most accreditation bodies work across both testing and calibration (often certification as well)
- · Exchanging ideas across different domains might enhance compatibility and assist in planning for future directions

On a personal note, I would like to express my deep gratitude to the International DCC Conference for doing exactly this!!







Thank you!

For more information
Brett Hyland
Brett.Hyland@nata.com.au

Back to "Table of Contents" above

P02 Advancing Digital Calibration Certificates: The INM's Role in Modernizing Metrological Services in Latin America

Presenting author: Giro Alberto Sanchez, Instituto Nacional de Metrologia de Colombia (INM)

Email: csanchez@inm.gov.co

Abstract

The digital transformation of quality infrastructure has become a key driver for enhancing efficiency, traceability, and interoperability in metrological processes. Digital Calibration Certificates (DCC) are central to this transformation, offering machine-actionable information that supports seamless data exchange among laboratories, accreditation bodies, and stakeholders. This abstract focuses on the pivotal role played by the National Metrology Institute (INM) in promoting the adoption and development of DCCs, highlighting its efforts in preparation, standardization, and socialization.

The INM has been actively involved in initiatives under the Inter-American Metrology System (SIM) to promote the digitalization of metrological services. Its general experience in this field includes the production of reference materials via a timestamped environment and participation in various seminars organized by the PTB of Germany, BIPM, and SIM, which provided key guidelines on the development of digital infrastructures. These efforts have laid the groundwork for future DCC implementation.

A real-world use case from the INM's Temperature and Humidity Laboratory illustrates the early stages of a project aimed at integrating DCCs into routine calibration processes. Supported by standards for digital languages provided by PTB, this initiative anticipates national industry needs and enhances compliance with ISO/IEC 17025 requirements regarding traceability and reporting. The adoption of DCCs not only enhances data integrity and reduces errors but also supports the modernization of metrological services.



Key challenges encountered during the implementation include ensuring interoperability with existing quality management systems, meeting stakeholder expectations for usability, and aligning with international frameworks for digital quality infrastructure. By addressing these challenges, this project contributes to the development of a harmonized approach to DCC adoption across the region.

This work concludes by outlining recommendations for future developments, including the creation of regional guidelines for DCC implementation, further stakeholder engagement, and the exploration of software solutions that facilitate integration. These efforts aim to foster greater consistency in the application of DCCs, ultimately supporting the modernization of metrological services in Latin America.

Keywords: Digital Calibration Certificates, ISO/IEC 17025, quality infrastructure, metrology, interoperability, Latin America.

Back to "Table of Contents" above

Presentation of Giro Alberto Sanchez





5TH INTERNATIONAL DCC CONFERENCE

DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

Advancing Digital Calibration Certificates: The INM's Role in Modernizing Metrological Services in Latin America

Ciro A. Sánchez, Jeyson A. Monroy and Adrian L. Blandón Instituto Nacional de Metrología de Colombia – INM

February 25th, 2025

Introduction and Context





Introduction




- Overview of Digital Calibration Certificates (DCC) and their impact on metrological traceability.
- Strategic importance of digital transformation in metrology.

www.inm.gov.co



Instituto Nacional de Metrología de Colombia – INM









INM's Role in Digital Transformation



- · Leading the implementation of DCC in Latin America.
- Enhancing interoperability and data integrity in metrological processes.

www.inm.gov.co



Inter-American Metrology System - SIM







Introduction to SIM

- Inter-American Metrology System (SIM)
 - A regional organization promoting metrology cooperation across the Americas.
 - Enhances the quality infrastructure by fostering collaboration among National Metrology Institutes (NMIs).
 - Aims to support trade, health, safety, and environmental protection through reliable measurements.

www.inm.gov.co



Digital Transformation in Metrology at SIM

- **Driving Digital Quality Infrastructure**: Enhancing reliability and efficiency in calibration and testing.
- **Promoting Digital Metrological Traceability**: Ensuring accurate and consistent measurements across member countries.
- Collaboration and Innovation: Facilitating cooperation between NMIs to develop digital solutions for metrology.
- **Strategic Vision**: Leading the region in adopting international standards for digitalization in metrology.





SIM's Working Group on Digital Transformation (SIM-MWG14-M4DT)



- Objective and Focus Areas
 - Leads digital transformation initiatives within the SIM community.
 - Focuses on:
 - Digital Quality Infrastructure
 - Digital Metrological Traceability
 - Metrology for AI

www.inm.gov.co



Achievements and Plans

- Barothermohygrometer Prototype (2023 2024): A prototype was developed to measure environmental conditions, enhancing digital metrological infrastructure.
- **2025 Goal:** To share the measurements obtained with the prototype across different countries, promoting digital transformation in the region.
- Calibration and DCC Issuance: Plans include calibrating the barothermohygrometer and issuing DCCs, ensuring interoperability and traceability in the digital environment.



Implementation in the Temperature and Humidity Laboratory










Progress Report



www.inm.gov.co



• Project 2404076, undertaken by the Measurement Automation Group, titled 'Digital Calibration Certificate in the INM Temperature and Humidity Laboratory', was an initiative within the framework of the digital transformation of the services offered by the National Institute of Metrology (INM).



- The main objective of this project was to adapt the Digital Calibration Certificate (DCC) prototype, which was previously developed for the Time and Frequency Laboratory in 2023, specifically for the Premium Legal Time service. This adaptation aims to enable the issuance of digital calibration certificates for the thermometer calibration service in the Temperature and Humidity Laboratory.
- By implementing this digitalization process, the project seeks to enhance calibration and certification procedures, improving the efficiency and traceability of the services provided by INM.





System Architecture and Integration



www.inm.gov.co



• **Framework:** Django with Python, leveraging flexibility and scalability for digital certificates.

Database Adaptation:

- Modified the existing database to support thermometer calibration data.
- Ensured data integrity and traceability through a structured storage system.

Script Development and Automation:

- Customized scripts to automate certificate generation, reducing manual intervention.
- Enhanced efficiency and minimized errors in the calibration documentation process.





Digitalization Workflow



www.inm.gov.co



· Certificate Generation Process:

- o Input: Calibration data collected from thermometers.
- Processing: Data validation, calculation of measurement uncertainty, and certificate formatting.
- Output: Digital Calibration Certificate (DCC) in XML format, ensuring compatibility and interoperability.

Integration in Existing Infrastructure:

- 。 Seamless connection with laboratory management systems.
- Ensured compliance with ISO/IEC 17025 requirements for calibration traceability.

www.inm.gov.co





Technical Challenges and Solutions

Interoperability Issues:

- Challenge: Integrating DCC into the existing technological infrastructure and ensuring data exchange.
- Solution: Utilized XML schema standards to enable seamless communication between systems.

User Adoption and Training:

- Challenge: Transition from traditional to digital calibration certificates.
- Solution: Conducted training sessions and developed user guides for laboratory personnel: digital certificates.

www.inm.gov.co




Data Security and Integrity:

- Challenge: Ensuring the authenticity and protection of digital certificates
- Solution: Implemented encryption protocols and digital signatures.



Results and Impact





www.inm.gov.co





• Efficiency Gains:

- Certificate issuance times are expected to decrease, minimizing manual errors.
- Calibration workflows will be streamlined, enhancing overall productivity.

Accuracy and Traceability:

- Accuracy in data processing and measurement uncertainty calculations will improve.
- 。 Digital record-keeping and data integrity will strengthen traceability.

www.inm.gov.co





Operational Improvements:

- Resource allocation will be optimized, enabling staff to focus on core metrological tasks.
- Client satisfaction is anticipated to rise with faster certificate delivery times.

www.inm.gov.c



Next Steps and Strategic Vision





www.inm.gov.co





Expansion of DCC Implementation:

- Extend the digital certificate system to other calibration services in the laboratory.
- Collaborate with international metrology institutes for standardization and best practices.

. 2025 Objectives:

- 。 Full integration of DCC in the Temperature and Humidity Laboratory.
- 。 Participation in international intercomparisons and validation of DCCs.
- Contribution to the Digital Transformation Group at SIM, promoting regional standardization.

www.inm.gov.co



Impact on Metrological Services

- Enhanced efficiency and accuracy in calibration processes.
- Improved traceability and data integrity.





Strategic Vision

- Proactive approach to digital transformation
- Anticipating future industry demands
- Aligning with international best practices


www.inm.gov.co



Challenges and Solutions

- Resistance to change and adoption barriers
- Solutions: Training, collaboration with stakeholders, and technical support








- DCC as a key enabler of digital transformation in metrology
- INM's leadership in modernizing metrological services in Latin America
- Strategic anticipation of industry needs

www.inm.gov.co







- Thank you!
- Questions?





Contact information

csanchez@inm.gov.co Physical Metrology INM

www.inm.gov.co

Back to "Table of Contents" above

Day 1 (2025-02-25 Tuesday)

Parallel Session "Key Aspects of DCC I"

(Chair: Shanna Schönhals)

001 Harmonisation of DCCs in DKD - an update

Presenting author: Thomas Krah, PTB Germany

Email: thomas.krah@ptb.de

Additional authors: Martin Czaske, Muhammed-Ali Demir, Gisa Foyer, Julian Haller,

Wolfgang Schmid

Abstract

Harmonisation is one of the key elements for the DCC to become the common exchange format for calibration data. One aspect of the harmonisation is that the calibration laboratories creating DCCs need to know which information should be written in each element of the DCC. Another aspect is that the users who process the calibration data, especially when dealing with multiple measurands, need to be able to read all the different DCCs in a simple way, without having to adapt their software to each measurand or calibration supplier. One way of harmonising the DCC could be to create a specification within a standard. This would probably be the most simple and effective way. However, it usually takes many years for a standard to become established.



In Germany, we chose the German Calibration Service (Deutscher Kalibrierdienst, DKD) with its well-established structures and the extensive knowledge of its members to speed up the process of harmonising the DCC. In all 13 technical committees, the members of the DKD are working on standardising the DCC.

Several committees have already published some of their work. There are community specific DCC examples available, expert reports which describe the DCC examples and give additional explanations, and a machine interpretable dictionary for the used refTypes is accessible for everybody worldwide. Currently is an interoperability challenge from the subcommittee mass in preparation to validate if their DCC example together with their expert report covers all relevant aspects and thus leads to an unambiguous content of the DCCs the participants of this challenge created.

Furthermore, DKD experts developed an easy-to-use tool which shows if all mandatory and if applicable optional requirements of the ISO 17025 are covered in any given DCC.

Back to "Table of Contents" above

Presentation of Thomas Krah







Harmonisation of DCCs in DKD - an Update

5th International DCC Conference 2025

T. Krah, M. Czaske, W. Schmid, Section 9.1
Gisa Foyer, WG 1.14
M.-A. Demir, WG 1.24
J. Haller, Sartorius Lab Instruments





Introduction



What is the DKD?

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



Nationales Metrologieinstitut

What is the DKD?



Abbreviation for **D**eutscher **K**alibrier**d**ienst (German Calibration Service)



DKD's mission is to achieve high quality in calibration laboratories

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin 25.02.2024




Nationales Metrologieinstitut International DCC Conference 2025



What is the DKD?





Harmonisation of DCCs in DKD



Harmonisation of DCCs in DKD



Harmonisation of DCCs in DKD



Technical committees

- 1. Direct Current and Low Frequency
- 2. High Frequency and Optics
- 3. Force, Acceleration and Acoustics
- 4. Length
- 5. Temperature and Humidity
- 6. Pressure and Vacuum
- 7. Mass and Weighing Instruments
- 8. Metrology in Chemical and Material Properties
- 9. Materials Testing Machines
- 10. Torque
- 11. Flow Quantities
- 12. Metrology in Laboratory Medicine
- 13. Measurement Uncertainty



SC_{general}

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin 25.02.2024



Harmonisation of DCCs in DKD



Composition of the members in SC gene

- All TCs represented.
- From each TC:
 - the TC chair.
 - If applicable the sub-committee chair.
 - Further representatives
- Area of operations of members:
 - Laboratories
 - Software manufacturers
 - Device manufacturers
 - NMIs





Nationales Metrologieinstitut International DCC Conference 2025

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



Output of the TCs



Output of the TCs



Physikalisch-Technische Bundesanstalt Braunschweig und Berlin 25 02 2024



Nationales Metrologieinstitut

Output of the TCs: expected





- Development of DCCs for selected important measurands and measurement equipment types (Good practices / templates).
- Harmonised wording.
- Expert reports.



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin 25.02.2024



Nationales Metrologieinstitut International DCC Conference 2025



Output of the TCs: delivered



- Development of DCCs for selected important measurands and measurement equipment types (Good practices / templates).
 - DCC mass available.
 - DCC length.
 - Code extracts available.
- Harmonised wording.
 - Thesaurus created. RefTypes machine-interpretable and with persistent IDs https://digilab.ptb.de/dkd/refType/vocab/
- Expert reports.
 - Expert report mass available https://www.ptb.de/cms/fileadmin/intermet/dienstleistungen/dkd/archiv/Publikationen/Expertenberichte/DKD-E_7-2_rev1_en.pdf
 - Report length available. https://www.ptb.de/cms/fileadmin/internet/dienstleistungen/dkd/archiv/Publikationen/Expertenberichte/DKD -E_4-3_2024-12.pdf
 - Report admin. block in draft version for members downloadable.



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin 25.02.2024

10



Output of the TCs: delivered





тс	Example DCC	Expert report	
Mass	Mass	Available https://www.ptb.de/cms/fileadmin/internet/dienstleistungen/dkd/ archiv/Publikationen/Expertenberichte/DKD -E 7-2 rev1_en.pdf	
Length	Gauge Block	Available https://www.ptb.de/cms/fileadmin/internet/dienstleistungu archiv/Publikationen/Expertenberichte/DKD -E_4-3_2024-1	
Scales/weighing instruments	Scales	In progress	
Metrology in Laboratory Medicine	Amount of substance concentration/fraction	In progress	
Temperature and Humidity	Resistance thermometer	In progress	
Force, Acceleration and Acoustics	Force transducer Acceleration-sinusoidal and acceleration-shock	In progress	
Material Testing Machines	Tensile testing machine	-	
Pressure & Vacuum	Digital pressure measuring device, acc. DKDR 6-1 (de, en, es) Vacuum measuring device, acc. DKDR 6-2 (only in de)	In progress	
Torque	Calibration of static torque measuring devices, acc. to DIN 51309	-	Status of work:
Metrology in Chemical and Material Properties	Pipettes	-	Ready & available Ready, to be released soon
DC & LF	Multifunction calibrator	-	First results
Flow Quantities	Fluid flow	-	In progress
Measurement Uncertainty	Various code snippets	In progress	

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin





Output of the TCs: delivered







DKD ISO 17025 Checker

Idea behind it:

- Provide an easy tool to check if a given DCC fullfills all requirements of the ISO/IEC 17025 towards calibration reports.
- Simplify the work of laboratory stuff, customers and auditors.

Realisation:

- Usage of a XSLT file to check and process the content of a given ${\bf D}{\rm CC}.$
- Provide a web application for checking the DCCs. The web application is running locally on a computer without connection to any server.

Publication:

- The Expert report general conventions and recommendations for a harmonised issueing of digital calibration certificates (DCCs)
- The XSLT file for checking DCCs against the ISO/IEC 17025
- The corresponding web app

Will be released in a bundle in April.





Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin









DKD ISO 17025 Checker


Functionality:

- The XSLT file transforms a given DCC (XML) into a HTML file which displays the result of the check in human readable form.
- The XSLT file contains xpath expressions to find the right data and routines for checking against the ISO/IEC 17025 requirements.

Output:

- If everything is ok, green entries are shown.
- In column Data is the content of the DCC elements for the requirements shown.
- If a requirement is violated, red entries are
- If possible, it is stated what is the cause of the violation.





Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Nationales Metrologieinstitut



Interoperability challenge DCC mass




Idea:

- Can we create a DCC which everyone can read just with the help of the DKD expert report mass and the DKD good practice example mass.
- Assessing interoperability of mass DCCs through comparison of test DCCs issued by different parties.

Output:

- For manually created DCCs: Prove that you correctly understood the PTB DCC schema and DKD-E 7-2.
- For DCCs created by dedicated software: Prove that you correctly understood and implemented the PTB DCC schema and DKD-E 7-2.
- Update the expert report and the good practice example if necessary.







Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin 25.02.2024

Nationales Metrologieinstitut
5th International DCC Conference 2025

Interoperability challenge DCC mass



- Information for interested participants:
 - Contacts:
 - Gisa.foyer@ptb.de
 - Julian.haller@sartorius.com
- For participation, it will be necessary that you :
 - Can create a DCC for a weight set either with a dedicated software (plus possible manual modifications) or manually.
 - Follow the PTB DCC Schema in version 3.2.1 (or 3.3.0).
 - Follow the DKD expert report DKD -E 7-2, Revision 1 (01/2024) - DOI: 10.7795/550.20240119B.

Task	When?	
Provide examples	02/25	
Create DCCs and submit them	03-05/25	
Develop queries for the scenarios		
Apply queries and provide first	06-07/25	
analysis	00 07720	
Discuss and analyse results	08-10/25	



DCC development is about to start. You can still join!

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



Nationales Metrologieinstitut International DCC Conference 2025







Physikalisch-Technische Bundesanstalt Braunschweig und Berlin

Bundesallee 100 38116 Braunschweig

Dr. Thomas Krah

phone: +49 531 592-9451 E-mail: thomas.krah@ptb.de

Dr. Wolfgang Schmid phone: +49 531 592-9100 E-mail: wolfgang.schmid@ptb.de

Dr. Gisa Foyer

phone: +49 531 592-1145 E-Mail: <u>gisa.foyer@ptb.de</u>

www.ptb.de Stand: 02/25 Dr. Julian Haller

+49 551 308-4523 phone: E-Mail: julian.haller@sartorius.com

Dr. Martin Czaske

+49 531 592-8020 phone: E-mail: martin.czaske@ptb.de

Muhammed-Ali Demir

+49 531 592-1241 phone: E-Mail: muhammed.demir@ptb.de

Back to "Table of Contents" above



Presenting author: Dr. Julian Haller, Sartorius Lab Instruments GmbH & Co. KG,

Email: julian.haller@sartorius.com

Additional authors: Nil

Abstract:

The benefits of digital calibration certificates (DCC) for efficient and modern test equipment management and the sensible use of calibration data are well known. The PTB has created a very flexible DCC scheme that makes it possible to provide content in different ways. This carries the risk that implementations from different labs may differ from each other and that there will be non- interoperable realizations.

For this reason, so-called expert reports are currently being prepared in expert groups of the German Calibration Service (DKD) to ensure the harmonized use of the DCC scheme for the respective subject areas.

While such an expert report for the harmonized creation of DCCs for weights was already published in 2022 and revised in 2024, the publication of such an expert report for the harmonized creation of DCCs for non-automatic weighing instruments (NAWIs) in accordance with the calibration guideline EURAMET cg-18 is imminent.

The aim of this presentation is to present the current status of the expert report and to give an overview of its contents.

Back to "Table of Contents" above

Presentation of Julian Haller



Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

Dr. Julian Haller, Sartorius Lab Instruments GmbH & Co. KG, 25.02.2025



DKD Working Groups

- DKD: Deutscher Kalibrierdienst (German Calibration Service)
- A panel to promote the uniformity in metrology and with the aim to offer complementary technical assistance, in which PTB and accredited calibration laboratories work closely together.

Working group "DCC Mass"

Working group "DCC Weighing Instruments"

- The panel's objective is to develop metrological foundations for calibrations.
- 13 Technical committees:
 - 1. Direct Current and Low Frequency
 - High Frequency and Optics
 Force and Acceleration

 - 4. Length
 5. Temperature and Humidity
 6. Pressure and Vacuum
 - 7. Mass and Weighing Instruments
 8. Chemical Measurands and Material Properties

 - 9. Materials Testing Machines

 - 10. Torque 11. Flow Measurands
 - 12. Measurands in Laboratory Medicine 13. Measurement Uncertainty

For cross-sectional harmonisation among the technical committees, see



2

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

DKD Working Group "DCC NAWIS"

D. Knopf, G. Foyer, A. Scheibner, S. Hackel (PTB) K. Fritsch (Mettler-Toledo) M. Häfner (Häfner) H. Lauer (Bizerba) S. Osang (Minebea, DKD-FA) M. Turino (Kern) Z. Zelenka (BEV, EURAMET TGM) J. Haller (Sartorius)

- Define conventions concerning structure, identification and use cases
- Final target: A guide that, if followed, ensures that any import algorithm can read in a DCC issued by any lab.
- Everything based on PTB's DCC schema!



It's all about being FAIR....

Measurement results | Measurement uncertainties

```
Repeatability
Test load (nominal): 100 g
                                                                                                                                                                                                                                                                                     Eccentricity
                                                                                                                                                                                                                                                                                       Test load (nominal):
                                                                                                                                                                                                                                                                                                                                                                                                                                                                 100 g
                                            100.0001 g
                                                                                                                                                                                                                                                                                     Front left
                                                                                                                                                                                                                                                                                                                                                                                                                                                  100.0002 g
                                                                                                                                                                                                                                                                                                                                                                                                                                                     99.9997 g
                                            100 0002
                                                                                                                                                                                                                                                                                     Back left
                                             99.9998 g
                                                                                                                                                                                                                                                                                    Front right
                                                                                                                                                                                                                                                                                                                                                                                                                                                   99.9999 g
   <dcc:result>
                     closed to the content of the co
Not Findable: Which values are for repeatability and which are for eccentricity?
  </dcc:quantity>
</dcc:data>
</dcc:result>
```

4

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

It's all about being FAIR....

Measurement results | Measurement uncertainties

Eccentricity	
Test load (nominal):	100 g
Center	99.9999 g
Front left	100.0002 g
Back left	99.9997 g
Back right	100.0000 g
Front right	99.9999 g
	Test load (nominal): Center Front left Back left Back right

```
Not Interoperable: What if other languages/other terms are used?
```



Eccentricity
Test load (nominal):

It's all about being FAIR....

Measurement results | Measurement uncertainties
Repeatability
Test load (nominal): 100 g

2	99.9999 g	Front left	100.0002 g	
3	100.0002 g	Back left	99.9997 g	
4	99.9999 g	Back right	100.0000 g	
5	99.9998 g	Front right	99.9999 g	
<dc< td=""><td>c:result refType="NAWI repeatabilityMeasurement"></td><td></td><td></td><td></td></dc<>	c:result refType="NAWI repeatabilityMeasurement">			
	<dcc:data></dcc:data>			
	<dcc:quantity></dcc:quantity>			Fin
	<si:reallistxmllist></si:reallistxmllist>			10
	<si:valuexmllist>0.1000001 0.0999999</si:valuexmllist>		0.0999998	(Sea
	<si:unitxmllist>\kilogram<td>List></td><td></td><td></td></si:unitxmllist>	List>		
				Acc
	cc:result>			
<dc< td=""><td>c:result refType="NAWI_eccentricityMeasurement"></td><td></td><td></td><td></td></dc<>	c:result refType="NAWI_eccentricityMeasurement">			
				Inte
	<dcc:quantity></dcc:quantity>			(Sea
	<si:reallistxmllist></si:reallistxmllist>		90 AG	
	<si:valuexmllist>0.0999999 0.1000002</si:valuexmllist>	0.0999997 0.1000000	0.0999999	this
	<si:unitxmllist>\kilogram<td>List></td><td></td><td></td></si:unitxmllist>	List>		
				Rei
				INC
<td>cc:result></td> <td></td> <td>'</td> <td></td>	cc:result>		'	

indable Gearch for refType="NAWI_repeatability")

Scarcifior refrype="ivAvvi_repeatabl

Accessible

100 g

Interoperable

(Search will work for every issuer that follows this convention)

Reusable

4

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

FAIRness of DCCs via DKD expert reports

Overarching expert report with general rules and conventions:
In progress, to be published soon

Conventional Allowards Announced and Conventional Conventional Conventional Conventional Conventional Conventional Convention









Other communities:
In progress
Temperature
Humidity
Lab Medicine
Flow
Pressure
Electrical Quantities
.....



Current status of the "DCC NAWI" expert report

- · Final draft has been agreed
- Report will be submitted to DKD for publication in March
- Expected publication in Q2/2025




6

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

Content of the expert report

Table of contents



Basic information about DCCs and D-SI

- Chapters 1 and 2 have been taken over almost unchanged from the expert report "DCC Mass"
- Editorial note that chapters 1 and 2 will be included in an expert report "DCC cross-sectional topics" in the medium term and can then disappear from the subject-specific expert reports



Content of the expert report

Chapter 3: describes NAWI-specific aspects for each DCC element

- e.g. in dcc:items:
 - How do I denote/identify MR and MI instruments?
 - How do I denote/identify modular instruments?

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

Content of the expert report

Chapter 4: Specific Application Rules for NAWIs:

- Multiple calibrations (before/after adjustment)
- Indication of adjustment and repair status
- Results for MR and MI instruments



Content of the expert report

Chapter 5: NAWI-specific refType attributes:

- NAWI repeatabilityMeasurement
- NAWI_eccentricityMeasurement
- NAWI_errorOfIndicationMeasurement
- NAWI_auxiliaryMeasurement
- NAWI_range1, NAWI_range2, NAWI_range3, NAWI_range4
- NAWI_resolutionOfDisplayingDevice
- NAWI_verificationScaleInterval
- NAWI_initialMeasurement
- NAWI_finalMeasurement

Proper definition/reference for each refType:

5.3.21 repeatabilityMeasurement

Definition note - Deutsch: Messung der Wiederholpräzision

Definition note - English: Measurement of the measurement repeatability

Element	Element in TemaTres	Beschreibung deutsch	Beschreibung englisch
PRÄFIX:	Broader Term	NAWI_	NAWI_
Quelle:	Bibliographic note	EURAMET Calibration Guide No. 18 Version 4.0 (11/2015) 5.1	EURAMET Calibration Guide No. 18 Version 4.0 (11/2015) 5.1
HINWEIS:	Scope note	für mehr Informationen zur Verwendung siehe Beispiele	for more information on use see Example note
ELEMENT.	Related Term	docresult	dccresuit
BEISPIEL:	Example note	<pre>deccreasit reffype="Wwill_repeatabilityMeasurement"> deccrease</pre>	

10

Table of contents

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

Content of the expert report

Chapter 6: Examples how to find specific information in a NAWI DCC

• e.g. Standard deviations of all repeatability measurements:

```
cxsl:for-each select="//dcc:result[@refType='NAMI_repeatabilityMeasurement']" >
cxsl:for-each select="//dcc:quantity[@refType='math_standardDevlationSample']" >
cyp
cxsl:value-of select="dcc:data/dcc:list/dcc:quantity[@refType='math_standardDevlationSample']"/>
c/pp
c/ssl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each></xsl:for-each</xsl:for-each</xsl:for-each</xsl:for-each</xsl:for-each</xsl:for-each</xsl:for-each</xsl:for-each</xsl:for-ea
```



Content of the expert report

Appendices with example DCCs

- single range balance
- multiple range balance with several measurements
- modular multi-interval balance

12

Julian Haller: Harmonisation of DCCs for non-automatic weighing instruments (NAWIs) through a DKD expert report to ensure full interoperability

Next steps

- · Publish expert report
- · Apply expert report to issue interoperable DCCs for NAWIs
- Collect experiences, feedback,.....
- Revise expert report to implement
 - Corrections
 - Overarching changes/updates
 - · Missing parts (uncertainty in use, minimum net weight)



Thank you.

Connect with me on LinkedIn to stay updated:



Dr. Julian Haller Sartorius Lab Instruments GmbH & Co. KG Scientist | Engineer Metrology Phone +49 551 308 4523 Mobile +49 175 2109435 julian haller@aattriiss.com

Back to "Table of Contents" above

1003 Implementation of Digital Calibration Certificates for Force According to ISO 376

Presenting author: Shanna Schönhals, PTB, Germany

E-mail address: shanna.schoenhals@ptb.de

Additional authors: Dirk Röske dirk.roeske@ptb.de, Christian Schlegel

christian.schlegel@ptb.de, Moritz Jordan moritz.jordan@ptb.de, Shan Lin Shan.Lin@ptb.de (PTB), Alex Cocev, Daniel Reissiger (SCS Concept), Axel Grönboldt (Atlas Copco), Torsten Wallner (HBK GmbH), Siegfried Herbold (STAHLWILLE), Michael Melzer (BAM), Peter Weyrich (CEH Messtechnik), Roland Klinke (MG Sensor), Siegfried Gerber (MPA Uni Stuttgart), Stefan Baus (Gedore), Saher Hassan (Kessler-QMP), Torsten Hahn (GTM), Robert Bausch, Stephan Posch (Kistler Instrumente GmbH), Alfred Botscher (Lorenz Messtechnik)

Abstract

Machine-readable and, ideally, interpretable digital transfer documents present a promising approach to automating processes that require smooth digital data transfer. To ensure interoperability between the issuing and receiving parties, as well as between different stakeholders in the calibration chain, it is crucial to provide standardized solutions when transforming from paper-based to fully digital calibration certificates. This requires the metrological community to reach a consensus on good practices and guidelines for their use. The German Calibration Service (Deutscher Kalibrierdienst, DKD) has established a coordinated process to develop the required solutions, and the work presented here is the result of the work done by the expert group formed by the force and torque community that is contributing to this effort.

In the field of force and torque calibration, specific issues relevant to the community have been identified, and proposed solutions will be introduced and explored in this presentation.



Most force calibrations, particularly in the industrial sector, follow the ISO 376 standard, so the expert group focused on good practices for this calibration process. ISO 376 not only outlines the measurement process and data handling for calibration results but also specifies requirements for calibration reports. As a result, ISO 376 calibration certificates are generally consistent in format, even across different laboratories, including descriptions of processes, raw data, processed results, and final classifications for different measurement ranges based on clear technical criteria. These standardized formats facilitate the transition to machine-readable data structures by limiting the number of variations.

However, machine-interpretable solutions had to be developed for various aspects and issues specific to the community. These included the structuring and reorganisation of tables, while allowing individual series of measurements made in different mounting positions and loading sequences to be addressed in a machine-interpretable way. The group developed solutions for expressing classifications according to the standard and the machine-interpretable interpolation function.

The presentation reflects the current state of the work. The final results of the expert group's activities will be published by the DKD in a special open-access expert report both in German and in English languages.

Back to "Table of Contents" above

Presentation of Shanna Schönhals





Task: develop a good practice certificate for force calibrations according to ISO 376



- ISO 376 standard defines:
 - Measurement process
 - Data handling of calibration results
 - And also requirements for calibration reports
 - CC already quite consistent
 - ► Facilitates transition to DCC
- Machine-actionable solutions sought for:
 - Structuring and re-organizing tables
 - Addressing individual series in different mounting positions
 - Expressing classifications
 - Interpolation function

DCC Working Group on Force and Torque by DKD Technical Committees 3 and 10 outside DKD Founding Selection Documentation Roll-out DCC Task Force Agree on simple and Creating pre-normative Send DCC to customer comprising DKD-TC significant documents 3 and 10 measurement task (ISO (Expert Reports) 376 cal. Cert.) Publish open-access with examples Harmonisation Members İssuing Voluntary participants Creating harmonised Issuing DCCs based on representing calibration DCC examples agreed examples labs and PTB members 19 persons, 13 labs



Starting Point: Analogue Calibration Certificate



- PTB sample calibration certificate for force transducers according to ISO 376
- First draft DCC created with the Excel Tool (Dirk Röske/PTB):
 - Mapping of calibration data → DCC
 - Output: XML DCC
- Discussion resulted in extending the XML example to cover use cases relevant to accredited laboratories, e.g.
 - Additional components of calibration item
 - Additional measuring equipment
 - Influence condition (Transducer zero signal)



Shanna Schönhals | Implementation of Digital 4 Calibration Certificate for Force Calibrations


Some words about refTypes

- refTypes are special attributes in DCCs ensuring association of meaning to data and providing context → machine-interpretability
- Essential: harmonised and FAIR
- DKD involved in definition and filling the database
- Combination of
 - basic refTypes (e.g. basic_temperature)
 - community-specific refTypes (e.g. force_compressive)

All used refTypes entered to Tematres DB: https://digilab.ptb.de/dkd/refType



Currently, 56 refTypes for force identified and described





Achieving Machine Interpretability



Addressing measurement series using a combination of refTvpes and metadata

Tabelle 4: Nullkorrigierte Anzeigewerte der Messreihen

Kraftrichtung: Druckkraft

Table 4: Zero-reduced values indicated for all series

able 4: Zero-re	duced values indicate	d for all series			Force applic	cation: compressive
F	<i>X</i> ₁	X_2	X_3	X ₄ '	X 5	X ₆ '
Kraft	Messreihe 1	Messreihe 2	Messreihe 3	Messreihe 4'	Messreihe 5	Messreihe 6'
Force	Meas. series 1 0 °	Meas. series 2 0 °	Meas. series 3 120 ° auf / up	Meas. series 4' 120 ° ab / down	Meas. series 5 240 ° auf / up	Meas. series 6' 240 ° ab / down
in kN	in mv/v	in mv/v	in mv/v	in mv/v	in mv/v	in mv/v
0	0,000000	0,000000	0,000000	0,000166	0,000000	0,000188
100	0,199962	0,199972	0,199982	0,200146	0,199992	0,200201
200	0,399929	0,399949	0,399969	0,400183	0,399989	0,400243
300	0,599901	0,599931	0,599961	0,600211	0,599991	0,600276
400	0,799878	0,799918	0,799958	0,800230	0,799998	0,800300
500	0,999862	0,999913	0,999963	1,000238	1,000013	1,000312
600	1,199854	1,199914	1,199974	1,200234	1,200034	1,200314
800	1,599859	1,599939	1,600019	1,600193	1,600099	1,600283
1000	1,999900	2,000000	2,000100	3100-00000-0000000000000000000000000000	2,000200	1 (7 0.0 0.0 0.0 0.0 0.0 0.0 0.0 0.0 0.0 0.
0	0,000122	0,000133	0.0400 me.700070707		Carlotte (Sept. 1997)	

Source: Calibration Certificate Sample PTB Dep. 1.2


Shanna Schönhals | Implementation of Digital Calibration Certificate for Force Calibrations according to ISO 376 25.02.20


Achieving Machine Interpretability



Addressing measurement series using a combination of refTypes and metadata

Tabelle 4: Nullkorrigierte Anzeigewerte der Messreihen






Source: Calibration Certificate Sample PTB Dep. 1.2

Shanna Schönhals | Implementation of Digital Calibration Certificate for Force Calibrations according to ISO 376 25.02.2025



Achieving Machine Interpretability






Source: Calibration Certificate Sample PTB Dep. 1.2

Combining tables for loading directions

PIB

- Analogue: separate tables on separate pages for human readability
- DCC: combined tables into one single table for each loading direction (tension and compression for the force, clockwise and anti-clockwise for the torque)
- Improve machine interpretability and simplify the data structure





Empty cells in tables: numerical values



Empty cells indicate values that were not measured

	ullkorrigierte Anz duced values indicate		Messreihen			htung: Druckkrat ation: compressive
F	<i>X</i> ₁	X 2	X ₃	X ₄ ′	X ₅	X ₆ '
Kraft	Messreihe 1	Messreihe 2	Messreihe 3	Messreihe 4'	Messreihe 5	Messreihe 6'
Force	Meas. series 1 0 °	Meas. series 2 0 °	Meas. series 3 120 ° auf / up	Meas. series 4' 120 ° ab / down	Meas. series 5 240 ° auf / up	Meas. series 6' 240 ° ab / down
in kN	in mV/V	in mV/V	in mV/V	in mV/V	in mV/V	in mV/V
0	0,000000	0,000000	0,000000	0,000166	0,000000	0,000188
100	0,199962	0,199972	0,199982	0,200146	0,199992	0,200201
200	0,399929	0,399949	0,399969	0,400183	0,399989	0,400243
300	0,599901	0,599931	0,599961	0,600211	0,599991	0,600276
400	0,799878	0,799918	0,799958	0,800230	0,799998	0,800300
500	0,999862	0,999913	0,999963	1,000238	1,000013	1,000312
600	1,199854	1,199914	1,199974	1,200234	1,200034	1,200314
800	1,599859	1,599939	1,600019	1,600193	1,600099	1,600283
1000	1,999900	2,000000	2,000100	2 cm	2,000200	
0	0,000122	0,000133	code convenience			

Solution: representation as NaN, available with D-SI schema v2.2.0, DCC schema 3.3.0

Shanna Schönhals | Implementation of Digital Calibration Certificate for Force Calibrations according to ISO 376 25.02.2025

Empty cells in tables: string-based classification



able 6: Classificat	ions				
Kraftbereich	/ Force range		Geräteklasse / /	Instrument class	
Von / From	Bis / To	Fall A / Case A	Fall B / Case B	Fall C / Case C	Fall D / Case D
100 kN	1000 kN	00	0,5	00	0,5
200 kN	1000 kN	00	00	00	00
300 kN	1000 kN	00	00	00	00
400 kN	1000 kN	00	00	00	00
500 kN	1000 kN	00	00	00	00

Source: Calibration Certificate Sample PTB Dep. 1.2

- Table restructured to match measured values table
- Use of NaN not appropriate and could lead to misinterpretation
- Undefined if a class cannot be determined
- noClass if class requirements are not met

Kraft Force \kilo\newton	Fall A Case A	Fall B Case B	Fall C	Fall D Case D
0	undefined	undefined	undefined	undefined
100	noClass	0,5	00	0,5
200	00	00	00	00
300	00	00	00	00
400	00	00	00	00
500	00	00	00	00
600	00	00	00	00
800	00	00	00	00
1000	00	00	00	00
0	undefined	undefined	undefined	undefined


Source: Human Readable generated by the GEMIMEG Tool from XML

Shanna Schönhals | Implementation of Digital Calibration Certificate for Force Calibrations according to ISO 376 25.02.2025 11



Empty cells in tables: string-based classification





Example for fitting function of the balanced mean




Ausgleichsfunktionen / Fitting functions X = X(F)force_fittingFunction Die in der Tabelle oben aufgeführten ausgeglichenen Mittelwerte X_a sind mit folgender Funktion berechnet: The balanced mean X_a values stated in the table above have been calculated with the following function: force_balancedMean $X_a = +0.001999767 \cdot F / kN + 1.99 \cdot 10^{-10} \cdot (E / kN)^2 + 1 \cdot 10^{-13} \cdot (E / kN)^3 mV/V$ Unter Berücksichtigung der Umkehrspanne v ergibt sich folgende Funktion für abnehmende force_functionParameters Considering the reversibility v the following function results for decremental forces. force_increasing $X_f = [+ 0.002000139] F / kN + 1.91 \cdot 10^{-10} \cdot (F / kN)^2 - 4.33 \cdot 10^{-18}$ $(F/kN)^3 + 1,67 \cdot 10^4 | mV/V$ force_functionParameters Reported in dcc:result force_decreasing

- refType force_fittingFunction
- Equation represented by dcc:formula using dcc:mathml (Content MathML)
- Coefficients of each variable is given using the element dcc:list with the child element dcc:quantity
- Each quantity receive a refType force_functionParameters.



Report on Good Practice for Force (ISO 376)





Conclusions and next steps



- Producing an additional, minimal example for simplicity
- Presenting and discussing results at DKD TC 3 and TC 10 meeting in May 2025
- Finalising and releasing the expert report
- Transferring solutions to torque very similar structures
 - E.g. DIN 51309
- Using force good practice as template for comparison



Acknowledgement to members of the DCC Task Force of DKD TCs 3 and 10 PTB



- Moritz Jordan, Shan Lin, Dirk Röske, Christian Schlegel, Shanna Schönhals (PTB)
- Alex Cocev, Daniel Reissiger (SCS Concept)
- Axel Grönboldt (Atlas Copco)
- Torsten Wallner (HBK GmbH)
- Siegfried Herbold (STAHLWILLE)
- Michael Melzer (BAM)
- Peter Weyrich (CEH Messtechnik)

- Roland Klinke (MG Sensor)
- Siegfried Gerber (MPA Uni Stuttgart)
- Stefan Baus (Gedore)
- Saher Hassan (Kessler-QMP)
- Torsten Hahn (GTM)
- Robert Bausch, Stephan Posch (Kistler Instrumente GmbH)



Back to "Table of Contents" above

A Taxonomy of Radiometric-Measurand Metadata 004

Author: Neil Swift, Measurement Standards Laboratory, New Zealand

Email: neil.swift@measurement.govt.nz

Speaker-Author: Mark Kuster, Independent Researcher, Consultant, USA

Email: mjk@ieee.org

NCSL International 141 Mil and Automation Committee



Abstract:

The NCSL International 141 MII and Automation Committee and collaborators continue to develop a taxonomy to uniquely identify and distinguish measurands in machine- actionable digital metrology documents such as calibration certificates, accreditation scopes, and instrument specifications. The committee intends to link one or more measurand taxa to each calibration and measurement capability (CMC) contained in the key-comparison database (KCDB) managed at the International Bureau of Weights and Measures. Further extensions would cover national metrology institutes and accredited laboratories CMCs not directly mapped to KCDB service codes. As a FAIR reference data set, such metadata will contribute to the interoperability critical for metrology's digital transformation and the DCC in particular.

This presentation outlines the taxonomy structure and presents progress toward covering the KCDB CMCs in general and efforts at New Zealand's Measurement Standards Laboratory to extend coverage to radiometry in particular. We report current KCDB coverage, new radiometric taxa, and describe the relations between the taxa and the KCDB's services, sub services, individual services and quantity designations.

the authentication of digital assets by end-users. This demonstrator shows that data spaces, as federated network, are inherently suited to these complex tasks. Implementing self-sovereign identities, digital wallets, and using the Eclipse Data Space Components controlled data sharing via beforehand agreed policies, Quality-X facilitates the secure creation, collection, distribution, and verification of credentials, digital assets, and meta data.

Furthermore, we discuss the potential of semantic models and semantic hubs in achieving interoperability across diverse digital assets. This interoperability is paramount for the seamless and automated interaction of various actors within the QI ecosystem, such as standardization institutes, conformity assessment bodies, and market surveillance authorities. We outline other QI processes that can be simulated in our Testbed: digital issuance and signing of certificates, tamper-proof metadata publication to neutral registries, and dynamic digital asset versioning from inception to lifecycle conclusion.



Back to "Table of Contents" above

Presentation of Neil Swift

A Taxonomy of Radiometric-Measurand Metadata

Neil Swift, Measurement Standards Laboratory, New Zealand, neil.swift@measurement.govt.nz Mark Kuster, mjk@ieee.org

NCSL International 141 MII and Automation Committee

Measurement-Information Infrastructure

NCSLI 141 (MII&AC)

DCC 2025

Tuesday, February 25, 2025

Introduction

Section 1

Introduction



Introduction

Today's Topics

- Introduction
- Motivation and Strategy
- Tools and Process
- 4 Conclusion

		4 D > 4 B > 4 E > 4 E > E	2000
NCSLI 141 (MII&AC)	DCC 2025	Tuesday, February 25, 2025	3/23

Introduction

Acronyms

- BIPM—International Bureau of Weights and Measures
- CMC—calibration and measurement capability
- DCC—digital calibration certificate
- DCR—digital calibration request
- DX—digital transformation
- FAIR—findable, accessible, interoperable, reusable
- KCDB—key comparison database

- M-layer—metrology information layer to support measurement systems
- NCSLI—NCSL International
- NMI—national metrology institute
- PID—persistent identifier
- PR—photometry-radiometry
- SIRP—SI reference point
- SoA—scope of accreditation

MII (measurement information infrastructure)

set of normative standards that unambiguously define data structures, taxonomies, service protocols and security for locating, communicating and sharing measurement information

Tuesday Fahrusay 25 2025

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 4 / 23



Section 2

Motivation and Strategy

NCSLI 141 (MII&AC)

DCC 2025

Tuesday, February 25, 2025

Motivation and Strategy

Motivation

- We want interoperability and reusability for machine-actionable measurement data.
 - Time and labor savings
 - Automatic data validation and verification
 - New productivity, services, data value and insight
 - Digital documents such as DCCs won't realize their potential otherwise.
- This requires standardized metrology metadata.
- One of least standardized data elements: measurand descriptions
 - Various DCC and other DX projects currently seek a measurand identifier.
 - Wanted: unique measurand PIDs

4日ト4日ト4日ト4日ト ま めので

NCSLI 141 (MII&AC) DCC 2

Tuesday, February 25, 2025




MII Proposal for Measurand Metadata

- Standardize a measurand taxonomy.
- Link taxa to KCDB CMC IDs.
- Tag digital document data and service offerings with the taxa as metadata.

Spec: Source.Voltage.DC

SoA: Measure.Voltage.DC

Cert: Source. Voltage. DC Cert: Measure.Voltage.DC



NCSLI 141 (MII&AC)

Tuesday, February 25, 2025 7 / 23

Motivation and Strategy

Solution: abstract taxon definitions customized for any context

Measure. Coefficient. Current. Spectral Irradiance

Definition: https://cie.co.at/eilvterm/17-25-063, https://cie.co.at/eilvterm/17-21-053

Required Parameters (instantiated with ranges)

- SpectralIrradiance
- DetectorType {SiliconSingleElement, SiliconTrap, InGaAs, GaN, Ge, Photomultiplier, Pyroelectric, Thermopile, CryogenicRadiometer, Photocell, ...}
- Wavelength
- SpectralBandwidth
- Optional Parameters (instantiated with ranges)
 - Irradiance
 - OpticalPower
 - Current[.Output]

NCSLI 141 (MII&AC)



Uptake and Interest

Known interest in the MII Measurand Taxonomy

- Metrology companies: Cal Lab Solutions, IndySoft, Keysight
- "Enhancing the interoperability of Calibration Certificates", Diego Nahuel Coppa, Marcos Bierzychudek, Andrés Torn, Héctor Laiz, FORUM-MD Workshop 2025
- "Proposal to improve the interoperability of the DCC", Diego Coppa, DCCC 2024
- "Toward standardized language and structure for machine interpretable DCCs", David Balslev-Harder, DCCC 2024
- "Digital workflows and e-services in metrology for small and medium NMIs", Søren Kynde, DTI's DCC Theme Day 2023
- "Machine Readability—Automating the Extraction of Data from DCCs", David Balslev-Harder, DCCC 2023

Uptake requires a wide-ranging taxonomy.



We would like to link the taxonomy to KCDB CMCs.

- Started with one service category (photometry) to test the process
- Extend coverage across the KCDB: electrical-RF, radiometry, ...
- Extend coverage for additional measurand demand down the traceability chain.
 - Derived NMI measurement services
 - Lower laboratories' accredited CMCs
 - Unaccredited CMCs

NCSLI 141 (MII&AC)

The MII taxonomy will point to the KCDB since it already exists but future digital documents and enterprise systems may use the taxonomy as a FAIR standardized metadata resource.





Differences in Approach

BIPM proceeds with its own digital transformation, which includes PIDs and digitalized CMCs.

So what does the MII Measurand Taxonomy add?

- Coverage for all CMCs: NMI services, accreditation scopes, unaccredited services
- Interoperability for all measurement data: CMCs, DCCs, instrument specs, DCRs
- Comprehensive M-layer coverage
 - All quantity and unit systems (e.g., Dobson units)
 - All scale types (e.g, ordinal scales for hardness)
 - Aspects that disambiguate traditional quantity systems
 - Machine-targeted unit conversions and scale transformations

All MII metadata systems will point back to BIPM CMCs, SIRP, etc., whereever relevant.

		4 D P 4 B P 4 B P B P 9 Q 0
NCSLI 141 (MII&AC)	DCC 2025	Tuesday, February 25, 2025 11 / 23
	Tools and Process	

Section 3

Tools and Process





Tools and Process

Tools

- Open-source taxonomy editor at https://github.com/CalLabSolutions/Metrology.NET_Public
- Project hosted on GitHub at https://github.com/NCSLI-MII/measurand-taxonomy
- Provides the taxonomy as a combined XML file and as separate XML taxa
- In process: a repository to integrate quantities, units, measurands, CMCs
 - Pull data from the M-layer and KCDB APIs
 - Link taxa to M-layer quantities (aspects) and units (scales)
 - Link taxa to KCDB CMCs and service categories
 - Monitor coverage, find orphans, expose gaps, facilitate development



Third effort: radiometry branch of the PR area. Steps:

- Review documentation: "Classification of Services In Photometry And Radiometry".
- Oraft initial taxa to match the services, sub services or individual services.
- Focus on taxon names first, though parameter sets influence the names
- Pull all the relevant KCDB CMC-entry data.
- Sort and filter on service, subservice, individual service, quantity and instrument.
- Analyze the method, units, uncertainty, comments and parameter data.
- Supplement the data from other sources—experts, literature, standards, etc.—as required.
- Oraft a taxon set to represent the unique measurands and their parameters.

The information assembled at BIPM through CIPM and its Consultative Committees greatly facilitates taxon creation. The BIPM's KCDB digitalization should speed the process considerably.

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 14 / 23



Tools and Process

Example KCDB CMC Data

KCDB Field(s)	Data	
metrologyAreaLabel	PR	
branchValue, serviceValue	Properties of detectors and sources, (2) Properties of detectors	
subServiceValue	(2.2) Responsivity, spectral, irradiance	
individualServiceValue	(2.2.2) Broad band detector	
quantityValue	Responsivity, spectral, irradiance	
instrument	Broad band detector	
instrumentMethod	Substitution with spectral responsivity standards	
cmc	unit: $A/(W/m^2)$, $V/(W/m^2)$, reading/ (W/m^2)	
traceabilitySource	_	
comments	Other type of detector could be measured	
internationalStandard	_	
parameters	Wavelength: 400 nm to 500 nm; Bandwidth: 4 nm to 50 nm;	
	Irradiance: 1 uW to 50 uW	

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 15 / 23

Tools and Process

Branch: Properties of detectors and sources, Service: Detectors

Sub Service	Individual Service	Quantity	Taxon(s)—PRELIMINARY
Responsivity, spectral, power	Broad band detector	Responsivity, spectral, power	Measure.Ratio*.OpticalPower
Responsivity, spectral, irradiance	Broad band detector	Responsivity, spectral, irradiance	Measure.Ratio*.SpectralIrradiance*
Î	Spectroradiometer	ſ	Measure.Ratio*.SpectralIrradiance*
Responsivity, spectral, radiance	Broad band detector	Responsivity, spectral, radiance	Measure.Ratio*.SpectralRadiance*
Responsivity, laser, power	General detector	Responsivity, laser, power	Measure.Ratio*.OpticalPower.CW.Laser
Responsivity, laser, energy	General detector	Responsivity, laser, energy	Measure.Ratio*.Energy.Pulse.Laser
Responsivity, solar, irradiance	General detector	Responsivity, solar, irradiance	Measure.Ratio*.Irradiance.Solar
I	1	Į.	Measure.Current.DC.DSR
1	Broadband detector	[Measure.Ratio*.Irradiance.Solar
Resp., UV, BB irradiance	UV radiometer	Resp., UV, BB irradiance	Measure.Ratio*.Irradiance.UV
Resp., UV, BB rad. exposure	UV radiometer	Resp., UV, BB radiant exposure	Measure.Ratio*.RadiantExposure.UV
Resp, solar, spectral, irradiance	Solar spectroradiometer	Resp., solar, spectral, irradiance	Measure.Ratio*.SpectralIrradiance.Solar*

^{*}These sub services also include ... Coefficient.Voltage...and ... Coefficient.Current...taxa (omitted for space).

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 16 / 23



Tools and Process

Branch: Properties of detectors and sources, Service: Sources

Sub Service	Individual Service	Quantity	Taxon(s)—PRELIMINARY
Irradiance, spectral	Tungsten lamp	Irradiance, spectral	Measure.SpectralIrradiance*
1	Deuterium lamp	J	1
Radiance, spectral	Tungsten lamp	Radiance, spectral	Measure.SpectralRadiance*
1	Deuterium lamp]	1
I	General source	1	1
Power, spectral total radiant	Laser	Power, spectral total radiant	Measure. Optical Power. CW. Laser
Ī	1	1	Measure. Optical Power. Pulse. Laser
Radiant intensity, spectral	Tungsten lamp	Radiant intensity, spectral	Measure.SpectralRadiantIntensity*
I	Deuterium lamp	1	I
Radiant flux, spectral	General source	Radiant flux, spectral	Measure.SpectralRadiantFlux*
Ì	Tungsten lamp		Î

^{*} Tokens including "Spectral" indicate spectral-density quantities, e.g., per nm.

Some KCDB CMC entries inevitably contradict themselves. Digitalization and input validation will help.

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 17 / 23

Tools and Process

Example KCDB-Linkage Results

Taxon	kcdbCode	
Measure.Coefficient.Voltage.SpectralIrradiance	AFRIMETS-PR-ZA-000006G1-1	
Measure.Coefficient.Current.SpectralIrradiance		
Measure.Ratio.SpectralIrradiance		
Measure.Coefficient.Current.OpticalPower	APMP-PR-JP-000005T7-2	
Measure.Coefficient.Voltage.Irradiance.Solar	EURAMET-PR-CH-00000DR2-1	
Measure.Energy.Pulse.Laser	SIM-PR-US-00000A2B-2	
Measure.SpectralRadiance	SIM-PR-CA-000009MC-1	

◆ロト ◆団 ト ◆ 豆 ト ◆ 豆 ◆ りへの

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 18 / 23



Conclusion

Section 4

Conclusion



Conclusion

Photometry and Radiometry Service Coverage as of Feb 22nd, 2025

PR Services

- Photometry
 ✓
- Properties of detectors

NCSLI 141 (MII&AC)

- Spectral emission properties of sources √
- Spectral properties of materials—taxa drafted
- Spectrally-integrated measurements for sources and detectors—taxa drafted
- Ocolour and other spectrally-integrated measurements of materials—taxa drafted
- Fibre optics

Some PR sub services have no CMC entries in the KCDB.





Conclusion

KCDB CMC Coverage as of Feb 22nd, 2025

Area Code	CMC Count	Metrology Area	Taxon Coverage
AUV	Acoustics, Ultrasound, Vibration	1260	scattered
EM	Electricity and Magnetism	4640	≈ 79 % *
L	Length	1696	scattered
М	Mass and related quantities	2973	scattered
PR	Photometry and Radiometry	1551	≈ 56 % *
QM	Chemistry and Biology	6455	none
RI	Ionizing Radiation	3708	none
Т	Thermometry	3001	scattered
TF	Time and Frequency	825	scattered
	Total	26109	17 % to 29 %

^{*} Reported at NCSLI 2024 Conference

NCSLI 141 (MII&AC) DCC 2025 Tuesday, February 25, 2025 21 / 23

Conclusion

Recommendations

Establish and use a measurand taxonomy

- Unique measurand PIDs with unlimited human-readable aliases
- Fully qualifies measurands for interoperable digital documents
- (DCCs, instrument specs-DCRs, SoAs, ...)
- Provenance to BIPM: KCDB, SIRP, etc.

Collaboration

- Available for metrologists to augment, review and critique
- Current information
 - GitHub for configuration management (submission, review, approval): https://github.com/NCSLI-MII
 - Open-source taxonomy and SoA editors: https://github.com/CalLabSolutions/Metrology.NET_Public
 - Further info: http://miiknowledge.wikidot.com/

Tuesday February 25 2025 22 /22

NCSLI 141 (MII&AC)

DCC 202

Tuesday, February 25, 2025

22 / 23

^{*} Reported at DCC 2024, 2025 Conferences



Conclusion

Acknowledgments

Many thanks go to

- Neil Swift, MSL, New Zealand for subject-matter expertise
- PTB and the DCC conference organizers for the kind invitation
- NCSL International for its MII support
- NCSL International 141 MII and Automation Committee members for their MII development work
- Cherine-Marie Kuster

And Thank You for your time!

Collaboration opportunities? Please bring your expertise!

 Questions?
 Image: Control of the control

Back to "Table of Contents" above

005 A Matter of Trust

Presenting author: Andrew Freund

Senior Product Manager DTR TC TDS - Trusted Digitalisation Solutions M +49 (0) 171-532

96 42,

Email: Andrew.Freund@bdr.de

Additional authors: Technical contact:

Peter Schrameyer, Product Management Fiskalisierung, DTR M PB

D-Trust GmbH | Ein Unternehmen der Bundesdruckerei

Kommandantenstr. 15, 10969 Berlin

Tel + 49 (0) 170 – 301 85 77, Peter.Schrameyer@d-trust.net, www.d-trust.net

Abstract

It is not hard to argue that in the digital world we trust too much and take a lot of things for granted. For example, we assume that our paying for things with our smart watches and phones is completely secure despite the presence of NFC Scanners that can rob us and defraud us in seconds. We like the convenience of solutions like our Banking Apps and Google Pay and the speed and effectiveness of these solutions and quickly jump over possible security problems with "oh, it can't happen to me. I'm careful". In the corporate world, nobody looks at digital security with a critical eye unless it's necessary. When it's necessary, it's usually due to pain. That pain usually has large monetary consequences.

By passing the eIDAS Regulations, the EU have created means to anchor and support the trust the world puts into digital products. In the regulation are prescribed digital certificates to confirm website origins or secure connections, and digital signatures and seals that have the same legal standing as ink and stamps are but a few examples. All of which firmly hold up our quickly given trust in digital solutions.

In our presentation we will show that Qualified Electronic Seals confirm the origin of digital documents and protect against unnoticed changes. We will explain how a digital calibration



certificate wrapped in a qualified electronic seal anchors and supports the baked in quality of a digital calibration certificate. A Qualified Digital Seal is a foundation for the trust the consumers of Digital Calibration Certificates have in those certificates.

Back to "Table of Contents" above

Presentation of Andrew Freund

d-trust.



How do we trust?



06.03.2025 Teil der Bundesdruckerei-Gruppe bat-



Convienience or easy to steal?


d-trust.



06.03.2025 Teil der Bundesdruckerei-Gruppe bd.

The Pain





06.03.2025 Teil der Bundesdruckerei-Gruppe bd. 4

Teil der Bundesdruckerei-Gruppe





eIDAS Trust Services d-trust.

The trust services of the eIDAS Regulation ensure security and transparency in the digital world



Electronic Signatures
The individual's digital signature



Delivery Services Secure digital messaging as with the post office or bailiff



Electronic Seals
The trustworthy digital mark of Government and industry



Verification Service
The automatic independent verifier



Time Stamp
The official Time like on an electronic photo



Archive Service The forever digital archive



Website Certificates (QWAC)
The secure identification of a website owner






seal-me: D-Trust Qualified Remote Sealing Service

d-trust.



seal-me: Qualified Remote Sealing Service



Remote Sealing Service


d-trust.

Qseal-as-a-service



With the qualified remote sealing service, you can handle your sealing processes electronically from start to finish.

Simple and uncomplicated.





Securing and Anchoring our Trust in DCCs with QESeals

d-trust.



06.03.2025

· Sealing a DCC XML



Challenges

- · DCC is an stable XML with instable content with many types and attributes being optional
 - This can make the XADES Signature Type inpractible for the QTSP and can lead to errors in the signature process
- Guaranteeing that the DCC XML has not in any way been changed since issuance
 - All elements of an XML can be changed in a simple text editor by an intercepting party
- Guaranteeing the issuer of the DCC XML, the source of the document
 - In an age where Hackers route through multiple IP Adresses, determining the source of an XML document can be tricky

There is a solution already being used multiple D-Trust Customers in Health Care and Government for non-PDF Documents



06.03.2025 - Copyright 2019 Bundesdruckerei GmbH -

Teil der Bundesdruckerei-Gruppe



• The Solution: Remote Sealing using Enveloping Specified in RFC 5652 - Interworking

d-trust.



Customer application uses seal-me API and addresses an enveloping proxy with it

Data protection: the request from the customer application to the enveloping proxy contains <u>plain text documents</u>. This does not create a data protection problem, as the proxy lies within the customer's IT!

Customer application receives a seal with the embedded, unchanged plain text document back from the enveloping proxy

Enveloping-Proxy calls D-TRUST

The plain text document is hashed in the Proxy on the way to D-TRUST

D-TRUST only recieves the document hash

The document is enveloped in the CADES Signature on the return path to the Customer Application

Enveloping is performed as specified in RFC 5652 Section 5.2 "EncapsulatedContentInfo"

06.03.2025

- Copyright 2025 Bundesdruckerei GmbH -

Teil der Bundesdruckerei-Gruppe bd.

11

· Enveloping according to RFC 5652 - Principles and Advantages



Enveloping according to RFC 5652 Section-5.2

The eContent field is empty in the seal-me signature as seal-me only recives hashes

The eContent field will be addressed in the Enveloping Proxy and filled with the original unchanged plain text document on return path to the customer application.

Can be implented with standard libraries such as Bouncy-Castle

ASN.1 Struktur aus RFC 5652

```
EncapsulatedContentInfo ::= SEQUENCE {
     eContentType ContentType,
eContent [0] EXPLICIT OCTET STRING OPTIONAL
```

Advantages

- · Data protection of highly sensitive documents is easy to ensure despite remote signatures
- Signature and document can be processed and archived together in a compact form (as with PDF), which is an advantage over remote CADES signatures
- The process is independent of the data format of the plain text document to be signed
- · Plain text documents can be binary content of any size



06.03.2025

- Copyright 2025 Bundesdruckerei GmbH -

Teil der Bundesdruckerei-Gruppe 💆.



A Secured DCC





06.03.2025

Teil der Bundesdruckerei-Gruppe



Andrew Freund Senior Product Manager

Tel.: +49 (171) 5329642 E-mail: andrew.freund@bdr.de

Hinweis: Diese Präsentation ist Eigentum der D-Trust GmbH. Sämtliche Inhalte – auch auszugsweise – dürfen nicht ohne die Genehmigung der D-Trust GmbH vervielfältigt, weitergegeben oder veröffentlicht werden. ©2022 by D-Trust GmbH.



14

Back to "Table of Contents" above



Day 1 (2025-02-25 Tuesday)

Parallel Session "Applications NMI I"

(Chair: Brett Hyland)

O06 Concepts and frameworks fostering interoperability within a digital quality infrastructure: DCC development in Colombia and QI infrastructure interoperability under an Enterprise Architecture Framework

Presenting author: Rodolfo Manuel Gomez Rodriguez

Email: jefeoidt@inm.gov.co

Additional authors: Jose Eduin Culma Caviedes jeculma@inm.gov.co

National Metrology Institute of Colombia

Abstract

For the DCC adoption, development, and implementation in the INM Colombia, we are working decisively in identifying, structuring, and generating a data governance model for all the processes we have, including data generated during measurements. So, it has been recognized that it's necessary to employ a reference framework that allows us to define our main objectives, scopes, capabilities and maturity levels.

In this regard, some activities have been carried out to recognize the different data and variables used in the laboratories, with the objective of achieving a data collection and visualization that in the future be in accordance with standards that are useful for presenting measurement data in a format that aligns with the consensus in the metrological world, thus allowing its reading and interoperability.

To govern the data, we must identify the data we have. In this context, activities have been undertaken to review all management processes at the INM, allowing the identification of data flows for each process and enabling the initial characterization of these data.

Furthermore, according to the criteria adopted in the country to leverage data, the INM is designing and implementing actions and guidelines aimed at ensuring data quality, security, and privacy. So, while we are working on identifying tasks to do in the future, we are also advancing in structuring guidelines and procedures that will guide the future creation of data and consequently, their utilization. This allows us to perform in data quality, becoming Findable, Accessible, Interoperable and Reusable (FAIR).

Now we have an opportunity to achieve progress on another fundamental aspect of Enterprise Architecture: the identification of artifacts that are reusable across many tasks, referred to solutions building blocks (SBB). The reusability of SBB must be seen as an opportunity for being more effective, because some tasks that were solved by someone (area, laboratory or researcher), can be analysed by another regarding its viability of use and appropriation in another process and in the future by another institution, specially from the Quality Infrastructure.

This is a key element for moving forward under a structured framework to the digitalization of data, and therefore the construction of the DCC in the different NMI Colombia laboratories, as well as in the mechanisms for data interoperability with other stakeholders, especially with our partners.



The integration of Quality Infrastructure will allow us to advance collectively towards solutions that are useful and transversal for industries, where our new building has walls made by SBB as bricks and data as glue.

Back to "Table of Contents" above

Presentation of Rodolfo Manuel Gomez Rodriguez



Theme 4 –Concepts and frameworks fostering interoperability within a digital quality infrastructure

DCC development in Colombia and QI infrastructure interoperability under an Enterprise Architecture Framework

National Metrology Institute Colombia

February 2025

5TH INTERNATIONAL DCC CONFERENCE

DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE





Agenda

- 1. Policies and regulatory framework related to Digital Transformation in Metrology.
- 2. Enterprise Architecture Framework

www.inm.gov.co







1. Policies and regulatory framework related to Digital Transformation in Metrology

www.inm.gov.co

National Digital strategy 2023 - 2026

Main objective: unleash the potential of the digital transformation to overcome the economic, social and environmental challenges through the strengthening of its enablers and a strong and a decisive boost to **the use and appropriation of data and digital technologies**.



This strategy focuses on eight axes are:


- 1. Digital connectivity to change lives:
- 2. Access, use and exploitation of data to drive social transformation;
- 3. Digital security and trust to guarantee people's freedoms and integral development;
- 4. Digital skills and talent as a driver of opportunities;
- Artificial Intelligence and other emerging technologies to generate economic and social value;
- Public digital transformation to strengthen the State -Citizenship link;
- 7. Digital economy for productive transformation;
- 8. Digital society for inclusive, equitable, and sustainable development.

Departamento Nacional de Planeación

www.dnp.gov.c



Colombian Digital Government Policy (Decree 767 – 2022) and Digital Transformation Guidelines



Guidelines and standards applicable to public digital transformation (Decree 1263 of 2022)

- 1. Use and exploitation of data
- 2. Interoperability
- 3. digitization and automation of procedures, services and processes
- 4. Use of demand aggregation mechanisms
- 5. Cloud services
- 6. Strategic Plan for TI
- 7. Exploration for new regulations
- 8. Artificial Intelligence



www.doo.gov.c


2. Entreprise architecture framework

www.dnp.gov.co




Stakeholders in Digital Transformation - Colombian QI





Quality Infrastructure Sub-System


www.inm.gov.co






Departamento Nacional de Planeación Digital Transformation in metrology through **Enterprise Architecture**

















Thank you!

Rodolfo Gómez – jefeoidt@inm.gov.co

Back to "Table of Contents" above

O07 Calibration support system development of KTL (Korea Testing Laboratory) using certificate database established by OCR

Presenting author: Ryu Dong-Hun

Email: dhryu@ktl.re.kr
Additional authors: None

Abstract

KTL (Korea Testing Laboratory) has developed a data-driven calibration certificate management system, KArgos, for the digital transformation of calibration certificates. This system utilizes AI-based OCR (Optical Character Recognition) technology to convert calibration certificates, which exist in various document formats such as PDF, Excel, Word, and images, into a database (DB). KArgos provides statistical data based on the accumulated calibration results, such as general performance level of test equipment and fault possibility in calibration data. This was achieved by extracting data from various unstructured documents and standardizing it for storage. Additionally, KArgos oU'ers a function to correct the equipment's identification information (e.g., manufacturer, model name, serial number). As part of its future plans, KArgos will implement an online service for technical managers that automatically verify data, identify any anomalies, and notifies the technical manager during the digital sign stage of the calibration process. KArgos is expected to significantly enhance the digitization and data utilization of calibration certificates, contributing to the improvement of the eU'iciency and reliability of the calibration process.

Back to "Table of Contents" above



Presentation of Ryu Dong-Hun







Summary

- Certificates Database accumulation and analysis example
- Future plan

Example

- 1. DB accumulation using OCR
- 2. Grouping similar results I : manufacturer correction
- 3. Grouping similar results II : nominal value estimation
- 4. Analysis using the past results

Future Plan


- 1. Making a certificate check agent to enhance the quality.
- 2. Making a client service to digitized certificates.

1. Introduction: KTL






1. Introduction: KTL




1. Introduction: Certificates and database






2. Research field: Calibration Certificate OCR




2. Research field: Al-Based Similarity Search





2. Research field : Al-Based Similarity Search




2. Research field: AI-Based Similarity Search





2. Research field: AI-Based Similarity Search



2. Research field: Digitalization of Certificate

Grouping the results II

- Reference value may vary slightly and making a group could be difficult
- In some cases, nominal value estimation is required. Could be a source of error.

Cert. #1	Cert. #2		
100.0	100.1 SAME?		
100.0	101.1 SAME??		
100.0	104.5 ??		


 Having nominal value in DB helps a lot in grouping the measurement results








2. Research field: Digitalization of Certificate



2. Research field: Digitalization of Certificate

Example 1 torque wrench

- Minimal results included in DB to maximize application fields (Ref., Error, U)
- Results are checked compared to
- 1) Guideline range of the same model
- 2) Previous results of the same instrument






2. Research field: Digitalization of Certificate

It is possible to check how many certificates are present on the same manufacturer & model.


This guideline was made based on the 125 results of same model



2. Research field: Digitalization of Certificate

Example 2 Micro pipette


- Example of outlying data
- Current guideline is 25 % and 75 % range of all previous results
- Still needs more elaborated tool to set up the permissible range





2. Research field : Digitalization of Certificate

- Current guideline is 25 % and 75 % range of 484 results
- Manual input by each field expert is possible.



3. Future Plan

Certificate pre-check service:

The Certificate Pre-Check Service employs Al-powered OCR to extract and analyze data from certificates, identifying critical errors before approval

- 1. A calibration technician performs calibration
- 2. To prepare a calibration certificate.
- 3. To submit a calibration certificate for approval.




<u>+ Certificate Pre-Check Service :</u> detects errors in calibration data (incorrect results, equipment information, certificate forms, etc.) and notifies key review points

- 4. The technical manager reviews the certificate and makes the final approval.
- 5. The approved certificate is provided to the customer.



3. Future Plan



Summary

- Certificates Database accumulation and analysis example
- Future plan

Example

- 1. We issue 120k+ calibration certificates every year and want to make use of results
- 2. Trained AI to extract data from the past pdf certificates and accumulated DB
- 3. To make a statistical analysis, more work were needed.
 - unique identifier : manufacturer & model & SN works. But correction is needed.
 - nominal value needs : reference value may vary very slightly. Nominal value is useful.

Future Plan

- 1. Making a certificate check agent to enhance the quality.
- 2. Making a client service to digitized certificates.

20





Back to "Table of Contents" above

The Web-Based Digital Calibration Certificate Generation Platform developed by NIMT

Presenting author: Praiya Thongluang

Email: praiyat@nimt.or.th Additional authors:None

Abstract

The National Institute of Metrology Thailand (NIMT) has continuously studied the Digital Calibration Certificates (DCCs) in order to find the most suitable type of DCC for our staffs and customers. We have investigated and have experimented with both PDF/A3 format and XML format. Both formats possess distinct advantages. The PDF/A3 allows attachment of additional files possible and facilitates the inclusion of a digital signature. This format ensures both versatility and ease of use. On the other hand, the XML is suitable for machine-readable processes and offers significant benefits for automation and integration with other systems.

Upon these findings, NIMT has adopted a hybrid approach to DCC generation in order to complement strengths of both formats. Currently, NIMT's DCC is generated in PDF/A3 format with an XML file as an attachment. To facilitate this process, NIMT has developed a dedicated web-based platform accessible from dcc.nimt.or.th. This platform is made available for the general XML generation. Hence, the platform consists of two sections: one tailored for NIMT staffs with security and linkage to other NIMT's platforms and another part for general users.

For NIMT's part, the platform seamlessly integrates user information, administrative data, and calibration work orders with NIMT's existing systems. This integration is to ensure accuracy and efficiency when generate DCCs. NIMT staffs may create DCCs in PDF/A3 format with XML files attached, and both components are digitally signed for security and authenticity. For general users' part, the platform enables the generation of DCCs in XML format only, fulfilling the machine-readable requirements.



Additionally, the platform includes functionality verify schema and Schematron of DCCs in order to ensure that the DCC adheres to predefined standards. In order to validate digital signatures, the platform leverages the Electronic Transactions Development Agency (ETDA) verification service, guaranteeing compliance with recognized digital signature protocols. This comprehensive approach ensures that NIMT's DCC format meets the needs of users, both human and machine, while maintains the highest standards of accuracy, security, and interoperability.

Back to "Table of Contents" above

Presentation of Praiya Thongluang



TABLE OF CONTENT

- INTRODUCTION

 THE HYBRID APPROACH (PDF/A3 AND XML)
- IMPREMENTATION PROCESS SECURITY AND DIGITAL VERTIFICATION

/////

FEATURES AND FUNCTIONALITIES CONCLUSION OF WEB-BASED PLATFORM

2025 National Institute of Metrology (Thailand)





INTRODUCTION

In 2022, NIMT surveys customers and NIMT's staff about their opinion with Digital Calibration Certificate (DCC). The summary of the surveys is described below.

- 1. Most participants agree with changing the calibration format to digital.
- 2. Their **apprehension** depends on the type of organization and job position.

For the type of organization

- The private sector has biggest concerned about changing organizational culture
- The government sector has most concerned about the staff's lack of the new knowledge to support the DCC system.

For the job position

- The laboratory head and staff are concerned about the operation process.
- The department head focuses on budget and costs.







© 2025 National Institute of Metrology (Thailand)



IMPLEMENTATION PROCESS











- DCC in PDF/A3 and using datalink as a middleware
- Using Developer function in MS Excel to generate DCC in XML
- Using Python to generate DCC in XML
- The Web-based DCC generate platform








/////

NIMT's schema version 1.1

- Announcement NIMT DCC Schema version 1.1 based on PTB XML DCC v.3.2.1 with some modification.
- · Publish Schema Repository in ETDA website for reference in Thailand.





© 2025 National Institute of Metrology (Thailand)

URES AND FUNCT

1.Automated Accreditation Scope Verification

The system automatically verifies accreditation scope and selects the correct certificate template.

2.Hybrid Approach (PDF/A3 & XML)


- The system generates DCCs in both PDF/A3 and XML formats, ensuring data integrity and accessibility.
 Both formats are digitally signed by NIMT for authenticity.

3. Verification Platform

- XML Schema & Schematron validation
- Digital signatures for both PDF and XML files
- Compatibility with ETDA's platform for cross-validation



NIMT Web Portal



© 2025 National Institute of Metrology (Thailand)

DCC Automating accreditation scope verification concept

Automated accreditation scope verification has been implemented in the NIMT DCC web portal for certificate template type selection




© 2025 National Institute of Metrology (Thailand)






Hybrid approach

PDF/A3



XML



Both document was signed by NIMT

© 2025 National Institute of Metrology (Thailand)

q

Verification Platform

(FOT NIMT's schema version 1.1)

NIMT's Platform



🕢 🐝 XML - Schema

Schematron

🗙 🐝 XML- Digital Signature

× •>> PDF - Digital Signature

ETDA's Platform



X ->> XML - Schema

X •>> XML - Schematron

/////

PDF - Digital Signature

© 2025 National Institute of Metrology (Thailand)







DIGITAL CALIBRATION CERTIFICATE

Can create DCC via NIMT's platform in PDF/A3 format (Attach an XML file)

TYPE CLASSIFICATION

Automatically classify the type of certificate (NIMT, TISI, or CIPM)

•>>>

DIGITAL SIGNATURE

Can sign digital signature and timestamp both PDF and XML file

VERIFICATION

NIMT's platform supports DCC validation

© 2025 National Institute of Metrology (Thailand)



General User

- Can generate platform in PDF/A3
 - · Add feature to help customer extract the xml file

NIMT's staff User

Improve platform to support the complex structure data

Validation

- Support digital signature and time-stamp verify
 - · Accreditation scope validation

© 2025 National Institute of Metrology (Thailand)

12





Back to "Table of Contents" above

O09 Generating Mass DCC based on DKD expert report using a spreadsheet

Presenting author: Itzel Dominguez Mendoza (CENAM),

E-mail address: idomingu@cenam.mx

Additional authors: Carlos Galvan (CENAM), Siegfried Hackel (PTB), J. Armando Lopez-

Celis (CENAM), Gregorio Alvarez-Calva (CENAM), Luis Peña-Perez (CENAM)

Keywords: DCC, DCC-Schema, Excel, Subschema, Machine-readable, Digitalization,

DKD-Expert reports

Abstract

The DCC has been reported on in the context of many international conferences. The DCC format, which has been developed internationally with all parties involved (industry, laboratories, NMI, accreditors) under the coordination of PTB, is used and further developed by many. A comprehensive collection of quotations can be found in [1].

Meanwhile, there are several activities under way to integrate the DCC into production operations. In particular, it should be emphasised that intensive work is being done to create Good Practice (GP) DCCs. In addition to the initial activities, which can be found on the DCC homepage [2], the DKD [3] is particularly involved in the creation of Good Practice DCCs. These are published in the DKD expert reports. A first expert report on the determination of mass is already available [4]. Further expert reports are shortly to be finalised.

There is also a need in smaller calibration laboratories without larger IT departments, where Microsoft Excel is very often used to create calibration reports. To make it easier for these laboratories to get started with the DCC, the authors have developed a procedure [5] for automatically transferring the calibration measurement results to the corresponding area of the DCC. Based on a DCC subschema, which is tailored to the DKD expert report for determining the mass, an algorithm was further developed that carries out the transfer.



References:

- [1] Hackel, Siegfried, et al., Digitization versus Digitalization & Digital Transformation in the field of calibrations and their subsequent use (2024). Physikalisch-Technische Bundesanstalt (PTB). Available at: https://doi.org/10.7795/120.20240126
 - [2] https://www.ptb.de/dcc, see tile "Good Practice"
 - [3] https://www.ptb.de/cms/metrologische-dienstleistungen/dkd.html
- [4] Foyer, Gisa, et al., Expert Report DKD-E 7-2: Instructions on how to use the DCC schema to create a digital calibration certificate for weights and weight sets,

https://www.ptb.de/cms/fileadmin/internet/dienstleistungen/dkd/archiv/Publikationen/Expertenberichte/DKD-E 7-2 rev1 en.pdf

[5] Lopez-Celis J. Armando, Galvan-Hernandez Carlos, DCC measurement results XML code through a spreadsheet, available https://doi.org/10.1016/j.measen.2024.101483

Back to "Table of Contents" above

Presentation of Itzel Dominguez





Generating Mass DCC based on DKD Expert Report using a Spreadsheet

Itzel Dominguez (CENAM), Carlos Galvan (CENAM), Siegfried Hackel (PTB), José Armando Lopez-Celis (CENAM), Gregorio Alvarez-Calva (CENAM), Luis Peña-Perez (CENAM)







Initially

- Joint lecture by CENAM and PTB
- > Aim of presentation:
 - Demonstrating a way of DCC automation
- > Background:
 - Some colleagues feel that DCC is too complex





Outline

- ➤ Siegfried Hackel (PTB)
 - Introduction
 - DKD Expert Report: DCC for weights and weight sets
 - DCC-Subschema for DKD Expert Report





Introduction



- DCC has been well received by the community
- Many publications
 - Listing: https://doi.org/10.7795/120.20240126

In particular, it is used in the broad area of the pyramid

Plenary talk tomorrow: "DCC in production mode"







DKD Expert Report : DCC for weights and weight sets


- ▶ 1st DKD Expert Report: DCC Good Practice
 - https://www.ptb.de/cms/fileadmin/internet/dienstleistungen/dkd/archi v/Publikationen/Expertenberichte/DKD-E_7-2_rev1_en.pdf
- Many presentations about DCC Good Practice
 - See e. g. parallel sessions
 - o Key Aspects of DCC I
 - Industry Applications







DCC-Subschema







DCC-Subschema for DKD Expert Report

- DCC is based on a generic approach
 - Almost all calibration reports can be displayed with it
 - Specific informations on each value possible
 - o e.g.: Mobile calibration of particle filters
 - Very often not necessary at this level of detail
 - e. g.: Ambient conditions in an air-conditioned laboratory








DCC-Subschema for DKD Expert Report

- > The path to the subschema
 - What is not necessary?
 - o Specific informations on each value
 - E. g. ambient conditions are global
 - Are list-in-list elements necessary?
 - o No
 - ...



An Example

> DCC Schema










An Example






An Example

> DCC Schema











Data Analysis Process

The data analysis process for the time being and to favor natural change, it was decided to find a way to export the results directly from the work environment to the XML file.







DCC Subschema for DKD Expert Report











Generic spreadsheet template

The MS Excel template was subdivided into the following 4 areas:




Adapted subschema for use in MS Excel





Types of elements to bind



3.3.0





Nonrepeating XML elements








Types of elements to bind





Repeating XML elements









An XML mapping cannot be exported if ...

</Solve list of lists>



```
<xs:complexType name="usedMethodListType">

<xs:annotation>
| List of the methods used in the calibration.
| </sxs:documentation>
| </sxs:documentation>
| </sxs:documentation>
| </sxs:documentation>
| </scidental form of the calibration of the calibratio
```

 ${\sf Extract\ from\ DCC\ subschema: type\ definition \textbf{``usedMethodListype''.}}$







An XML mapping cannot be exported if ...

</Solve list of lists>



```
<!-- mandatory list of real elements -->

<sxs:element name="real" type="sizealInListType" maxOccurs="unbounded">

<sxs:anontation>

| <sxs:documentation xmlslang="en">
| Meta data element definition for a real measurement quantity in list.

This implementation differs from the pure real quantity in the way that the unit component is optional in order to allow a combination with a global unit in the list of real quantities.

</sxs:documentation>

</sxs:documentation>

<sxs:documentation xmlslang="en">
| Sxs:dement name="realOnce" type="sizealInListType"><!--Added-->

</scs:documentation xmlslang="en">
| Meta data element definition for a real measurement quantity in list.

This implementation differs from the pure real quantity in the way that the unit component is optional in order to allow a combination with a global unit in the list of real quantities.

</scs:documentation>

</scs:documentation</scs:documentation>

</scs:documentation</sc>

</scs:documentation</sc>
```

Extract from D-SI schema: new type definition "realOnce"





An XML mapping cannot be exported if ...

</Solve choice schema construct>



Extract from DCC subschema: type definition"dataType", choice construct.












"Remove all tags added to item names with the editor of your choice."







Revert the XML generated to the original valid schema







One example of this subschema can be found at:



Currently, this method has been successfully tested in several calibration services, such as gauge blocks by interferometry, step gauges, RTDs and weights in mass.











More questions? Please contact:

CENAM Metrology Institute;

km 4.5 Carretera to Los Cues, Municipality El Marqués 76246 Querétaro Mexico

Itzel Dominguez

E-Mail: idomingu@cenam.mx

Physikalisch-Technische Bundesanstalt Braunschweig und Berlin

Bundesallee 100 38116 Braunschweig Germany

Dir. u. Prof. Dr. Siegfried Hackel Phone: +49 531 592-1017 E-Mail: siegfried.hackel@ptb.de

www.ptb.de/dcc

Back to "Table of Contents" above

O10 Real-World Applications of Digital Calibration Certificates (DCC): Use Cases and Implementation Stories

Presenting author: Yoshitha Weerasinghe, Sri Lanka Standards Institution, Metrology

Division, Sri Lanka

E-mail address: yoshitha@slsi.lk

Abstract

Digital Calibration Certificates (DCC) are revolutionizing quality assurance by providing efficient, traceable, and automated solutions for calibration processes. This paper explores practical use cases and user stories that highlight the real-world applications and benefits of implementing DCC in diverse industries. In quality management systems, DCCs enable seamless integration, significantly reducing manual interventions and errors. They support industries in achieving regulatory compliance, such as with ISO/IEC 17025, by providing consistent, traceable documentation for calibration processes. Additionally, DCCs facilitate global collaboration by allowing stakeholders across borders to exchange harmonized and interoperable data smoothly. User stories from various sectors illustrate the transformative impact of DCCs. Metrology laboratories have reported improved efficiency and reduced administrative overhead by automating data handling and report generation. In industrial applications, companies in fields such as automotive manufacturing have adopted DCCs to digitize their calibration workflows, resulting in cost savings and enhanced audit readiness. Similarly, in the healthcare sector, the use of DCCs has improved the calibration of medical devices, ensuring reliability in diagnostics and better patient outcomes. The paper also outlines key implementation strategies for adopting DCCs, including collaboration with software developers to create machine-readable formats, training end-users for effective adoption, and building stakeholder confidence through pilot projects. These examples demonstrate the transformative potential of DCCs in advancing digital quality infrastructure, providing inspiration and a roadmap for broader adoption.



Keywords: Digital Calibration Certificates, quality assurance, interoperability, regulatory compliance, healthcare diagnostics, calibration workflows.

Back to "Table of Contents" above

Presentation of Yoshitha Weerasinghe



REAL-WORLD APPLICATIONS OF DIGITAL CALIBRATION CERTIFICATES (DCC): USE CASES AND IMPLEMENTATION STORIES

Presented by: Yoshitha Weerasinghe

Institution: Sri Lanka Standards

Institution, Metrology Division

Email: yoshitha@slsi.lk



INTRODUCTION

- Revolutionizing Quality Assurance with Digital Calibration Certificates (DCC)
- Their role in modern quality assurance
- Overview of today's discussion

BENEFITS OF DCCS

Efficient, traceable processes

Fewer manual errors


Seamless integration with quality systems





GLOBAL COLLABORATION • Harmonized calibration data • Facilitates international data exchange • Enhances global cooperation









USE CASE: HEALTHCARE SECTOR



Calibration of medical devices



More accurate diagnostics



Improved patient outcomes

IMPLEMENTATION STRATEGIES



COLLABORATE WITH SOFTWARE DEVELOPERS



USE MACHINE-READABLE FORMATS



TRAIN END-USERS EFFECTIVELY



BUILDING STAKEHOLDER CONFIDENCE



Pilot projects for initial testing



Collect feedback for improvements



Share success stories




CHALLENGES AND SOLUTIONS

- Technical integration issues →
 Collaborate with IT experts
- Resistance to change → Conduct user training
- Data security concerns →
 Implement secure data protocols



CONCLUSION Recap of benefits and real-world applications Future potential for digital quality infrastructure Encouragement for broader adoption





Back to "Table of Contents" above

Day 2 (2025-02-26 Wednesday) Plenary Session (Chair: Hugo Gasca)

P03 Advances in the DCC enablement at Beamex

Presenting author: Tuukka Mustapää, Beamex Oy Ab, Finland

E-mail address: tuukka.mustapaa@beamex.com

Additional authors:

Abstract

During the recent years, Beamex has been involved in multiple projects and codevelopment Initiatives related to the adoption of the DCC in the process industry. As the development of the DCC has progressed, the focus of the co-development initiatives has been shifting from prototypes and proof-of-concepts towards the readiness to support early adaptors in taking the DCC into use in their production environments.

As there are still many topics around the DCC that require further development, the enablement of the use of the DCC will be take place in multiple stages. This presentation discusses the ongoing projects and development activities at Beamex, the expected outcomes of those projects, and the outlines of a roadmap for the eventual support for the DCC in Beamex systems.


Initially, the support for the DCC will cover the possibility for a customer to import calibration results directly from a DCC into their calibration management system. The capability to import data from DCCs will be dependent on the establishment of good practices for the application in question and likely requiring co-development. In practice, this means that early on the scalability for different applications will have some limitations.

More extensive automation of the data exchange processes will require additional harmonization such as a standardized format for digital calibration requests (DCR) for enabling the two-way communication between the customer and calibration service provider. Additionally, the machine-to- machine communication can be further enhanced by use of centralized interfaces that can streamline the communication within calibration ecosystems consisting of numerous customer-service provider relationships.



Back to "Table of Contents" above

Presentation of Tuukka Mustapää



5th International DCC Conference

OUTLINE

- · Beamex in brief
- · Beamex calibration ecosystem & DCC
- · DCC related development
- · Learnings from PoCs
- DCC enablement
 - · Initial phase
 - · Extending DCC support
 - Maximizing value
- Summary

26 FEBRUARY 2025

ADVANCES IN DOC EVABLEMENT AT BEAMEX, TJUKKA MUSTAPAA



5th International DCC Conference

BEAMEX IN BRIEF

Who are we?

- A trusted partner for calibration solutions and services
 - · Field and workshop calibration equipment
 - · Calibration management systems
 - · Accredited calibration laboratory
 - · Expert services
- Technology and expertise available in over 90 countries
- · Over 15 000 customers worldwide

beamex

Our customer industries:









Pharmaceutical

Chemical

nical Power Generation

Oil & Gas



Beverage







Food &

Service Companies

Water & Wastewater

Other industries

26 FEBRUARY 2025 ADVANCES IN DCC ENABLEMENT AT BEAMEX, TUUKKA MUSTAPÄÄ

5th International DCC Conference


BEAMEX CALIBRATION ECOSYSTEM

Combination of calibration equipment, software tools and services

- On-prem and cloud-based calibration management solutions
- · Digital workflows

26 FEBRUARY 2025

· Platform for collaboration






5th International DCC Conference

BEAMEX CALIBRATION ECOSYSTEM & DCC

Aims of the DCC have a lot in common with ideas behind the Beamex calibration ecosystem

- · Optimization of calibration processes
- · Secure digital data flow
- · Data-based decision making



5th International DCC Conference


26 FEBRUARY 2025

26 FEBRUARY 2025

BEAMEX CALIBRATION ECOSYSTEM & DCC

However, many things will also need to be reconsidered or redefined

- Transition from a closed to an open ecosystem
- Differences in accredited and industrial processes
- Dependencies and compatibility with other systems



heamex



5th International DCC Conference

DCC RELATED DEVELOPMENT

- · Beamex has been involved in multiple co-development projects
- · Proof-of-concepts (PoCs):
 - · Importing calibration results from DCCs
 - · Exporting calibration results as DCCs
 - · Digital calibration requests (DCR)
 - · DCC exchange platform
- · Participation in DKD's DCC working groups

28 FEBRUARY 2025 ADVANCES IN DOC EVABLEMENT AT DEAMEX, TUUKKA MUSTAPÄÄ

5th International DCC Conference

26 FEBRUARY 2025

LEARNINGS FROM POCS

- Overall, there is still a lot of work to be done regarding harmonization and good practices
 - · The DCC needs to be flexible to support all use cases
 - The good practices must be precisely defined for them to be usable
- Current state of the DCC development allows us to get started with adding support for DCCs in our systems
- Ongoing development and PoCs will provide the basis for supporting DCC in production environments
- Even limited capabilities can provide value for the early adopters and help them prepare for wider use of DCCs



ADVANCES IN DOG ENABLEMENT AT DEAMEX, TUUKKA MUSTAPÄÄ



5th International DCC Conference

DCC ENABLEMENT - INITIAL PHASE

beamex

Basic features to support DCC for early adopters that we can already get started with

- Conceptualization and definition of requirements for the DCC services
- Definition of DCC related features for the first releases
- Primary focus: importing calibration results from DCCs
- · Support for limited number of use cases
- Implementing enhancements helping in enabling the use of DCCs



26 FEBRUARY 2025

ADVANCES IN DOC ENABLEMENT AT BEAMEX, TJUKKA MUSTAPÄÅ


5th International DCC Conference

beamex

DCC ENABLEMENT - EXTENDING SUPPORT

Features that we can get started with when certain conditions are met

- Adding support for more use cases as the good practices are defined
- Extending DCC support to more products and services
- Introduction of DCR to allow end-to-end communication in both ways in a similar manner
- Developing an interface that provides the basis for the data exchange platform



26 FEBRUARY 2025

ADVANCES IN DCC ENABLEMENT AT BEAMEX, TUUKKA MUSTAPÄÄ

beamex




5th International DCC Conference

DCR

DCR is needed as an assurance that the whole calibration process is completed according to plan

- Calibration procdures are defined in the calibration managment system
- Exchange of the DCR and DCC will be included in the audit trail
- Differences between accredited and industrial processes
- Agreements on the used procedures will need to be in place first



26 FEBRUARY 2025

ADVANCES IN DCC ENABLEMENT AT BEAMEX, TUUKKA MUSTAPÄÄ


5th International DCC Conference

COMMON INTERFACE

Exchange of DCCs and DCRs can be done similarly to exchange conventional calibration certificates, but it is not ideal

- Establishing an interface between the parties involved in the data exchange enables further automation of the process
- Simplified exchange of DCCs and DCRs between organizations that have several contractors or customers





26 FEBRUARY 2025

ADVANCES IN DOC ENABLEMENT AT BEAMEX, TUUKKA MUSTAPÄÄ



5th International DCC Conference

DCC ENABLEMENT - MAXIMIZING VALUE

beamex

Features that will maximize the benefits from digitalization, but require most of the aforementioned features and frameworks to be established first

- · Setting up the platform-based calibration ecosystem
- Introducing additional services to enhance the platform
- · Big data and analytics based decision making
- Investigating new opportunities to leverage the machine-actionable data




26 FEBRUARY 2025

ADVANCES IN DOC ENABLEMENT AT BEAMEX, TJUKKA MUSTAPÄÅ

, 2

5th International DCC Conference

SUMMARY



- Support and features will be incrementally added to Beamex systems
- A lot of collaborative development is still needed to optimize the processes concerning DCCs

For more information

Publications


26 FEBRUARY 2025

- · White papers, Beamex blog and more: resources.beamex.com
- · Social media: LinkedIn, YouTube, ...

ADVANCES IN DOC ENABLEMENT AT BEAMEX, TJUKKA MUSTAPÄÄ

beamex





Back to "Table of Contents" above

P04 Python-based DCC Tools Development and Implementation

Presenting author: Gean Marcos Geronymo, Inmetro (Brazil)

E-mail address: gmgeronymo@inmetro.gov.br

Additional authors: Nil

Abstract

This paper introduces DCC Tools, a suite of open-source Python-based software tools developed to streamline the handling of Digital Calibration Certificates (DCCs). These tools are designed to generate DCCs compliant with the XML schema proposed by the Physikalisch-Technische Bundesanstalt (PTB) [1] [2] and facilitate the embedding and extraction of human-readable PDF versions of calibration certificates. The transition from traditional paper-based or PDF certificates to machine-readable DCCs addresses the limitations of current practices, which often involve manual transcription of calibration data, leading to inefficiencies and errors. The proposed tools aim to enhance the digitalization of calibration processes, enabling seamless integration with modern digital workflows and improving accessibility, interoperability, and efficiency.

The DCC Tools package includes several key functionalities: dccGenerator, which generates PTB-compliant DCCs from JSON inputs; dccAddPdf, which embeds human-readable PDFs into DCCs; and dccGetPdf, which extracts PDFs from XML DCCs. Future plans include the development of dccReader, a tool for extracting measurement data from DCCs and converting it into JSON or spreadsheet formats. For this task we plan to use the pyDCC library developed by Siemens, which is available on GitHub [3]. The system architecture is serverless, using AWS Lambda for scalability, cost-efficiency, and ease of integration with existing workflows. The tools are designed to operate within the AWS Lambda free tier, making them accessible to calibration laboratories of various sizes and budgets. The tools are implemented



as Python scripts, so they can be easily adapted for running on a local server or private cloud if needed.

The adoption of DCCs represents a significant step toward true digitalization in calibration processes. Unlike traditional methods, which rely on paper or closed digital formats like PDF, DCCs store data in a machine- readable XML format, enabling automated processing and integration with digital systems. This is particularly important in the context of Industry 4.0, where real-time data exchange and automation are critical. The proposed tools also support the inclusion of digital cryptographic signatures, enhancing trust and security in the calibration process. DCC Tools software is freely available on GitHub [4] under the GPLv3 license, providing a practical and cost-effective solution for calibration laboratories seeking to modernize their workflows.

In conclusion, the DCC Tools represent a significant advancement in the digitalization of calibration processes. By enabling the generation, embedding, and extraction of machine-readable DCCs, these tools reduce the reliance on manual transcription, minimize errors, and improve productivity. The serverless architecture ensures scalability and cost-efficiency, making the tools accessible to a wide range of laboratories. Future work will focus on expanding the tools' capabilities and integrating them with emerging standards and technologies, such as the Digital SI (D-SI) format and metrology clouds, to further enhance their utility and impact.

References:

- [1] Hackel, Siegfried, Shanna Schönhals, Lutz Doering, Thomas Engel, und Reinhard Baumfalk. "The Digital Calibration Certificate (DCC) for an End-to-End Digital Quality Infrastructure for the Industry 4.0". *SCI* 5, Nr. 1 (2023). https://doi.org/10.3390/sci5010011.
- [2] PTB DCC Project. Available online: https://www.ptb.de/dcc (accessed on 20 th January 2025).
- [3] PyDCC. Available online: https://github.com/siemens/pydcc (accessed on 27 th January 2025).
- [4] DCC Tools GitHub Repository. Available online: https://github.com/gmgeronymo/dcc tools accessed on 27th January 2025).



Back to "Table of Contents" above

Presentation of Gean Marcos Geronymo

Python-based DCC Tools Development and Implementation 5th International DCC Conference

Gean Marcos Geronymo

Inmetro - Brazil

February 26th, 2025

Gean Marcos Geronymo (Inmetro - Brazil

Python-based DCC Tools Development and Implementat

February 26th, 202

1/38

Digital Calibration Certificates: True Digitalization

Digital Transformation

- Beyond simple digitization:
 - Content transfer not enough
 - Need for machine interpretation
 - Structured data format

DCC Advantages

- Machine-readable format
- Interpretable data structure
- Efficient storage system

Traditional Approach

- Paper to PDF conversion
- Focus on human readability
- Limited machine processing

Implementation Benefits

- Clearly defined format
- Precise data storage
- Automated processing
- System integration ready

◆□▶ ◆□▶ ◆□▶ ◆□▶ ◆□◆

Gean Marcos Geronymo (Inmetro - Brazil


Python-based DCC Tools Development and Implementat

ebruary 26th, 2025

2 / 38



Digital Calibration Certificate Structure - PTB proposal



Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025 3 / 38

Digital System of Units (D-SI)

Overview of D-SI

- XML format for seamless integration of measurement data
- Includes:
 - Numerical values
 - Associated units
 - Measurement uncertainties (value and distribution)

Visual Representation

- D-SI data model: See Slide 5
- Application example: Resistance value in ohms (Slide 6)

Gean Marcos Geronymo (Inmetro - Brazil)



D-SI Structure

SmartCom Uniform data format Minimum requirement Maximum information Y = y [SI] $U_{95}(k)$ distribution UTC Coordinated Universal type of distribution coverage factor

Figure: DCC Structure. Source: https://gitlab1.ptb.de/d-ptb/d-si/xsd-d-si/-/ blob/master/wiki/doc/RealQuantity.md

measurement value

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025 5/38

measurement uncertainty + confidence interval

D-SI XML Example

```
<si:quantity>
   <si:value>0.9999397</si:value>
   <si:unit>\metre\tothe{2}\kilogram\second\tothe{-3}\ampere\tothe{2}</si:unit>
   <si:expandedUnc>
     <si:uncertainty>2.10E-07</si:uncertainty>
     <si:coverageFactor>2</si:coverageFactor>
     <si:coverageProbability> 0.9545 </si:coverageProbability>
   </si:expandedUnc>
   <si:dateTime>2023-02-09T13:37:37-03:00</si:dateTime>
 </si:real>
</si:quantity>
```

Figure: Example of a fully compliant D-SI quantity statement. The electrical resistance unit (ohm) is stated in term of the SI base units ($m^2 \text{ kg s}^{-3} \text{ A}^2$) for a "platinum" quality class in terms of machine-readability. Alternatively, the unit can be represented as \ohn for a "gold" quality class.

February 26th, 2025



Digitalization of Calibration Process

Metrology Cloud (MC)

- Most desirable path for digitalization
- Enables seamless integration of DCCs
- Supports real-time data exchange

DCC Tools Support

- Initial DCC adoption
- Future integration with MC
- Flexible for various calibration workflows

DCC Tools Benefits

- Digitalization even with simple methods (e.g., email)
- Real-time action on instrumentation
- Key for smart factories and Industry 4.0

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025

Existing Solutions

PTB Tools

- GEMIMEG Tool:
 - Web-based, open-source
 - Manual data creation
- Excel Tool:
 - Macro-enabled
 - Generates DCC XML from blocks

SIEMENS: PyDCC

- Python library for DCC access and processing
- Simplifies extraction and manipulation of DCC data
- Enables seamless integration into digital manufacturing
- Available on : https: //github.com/siemens/pydcc



Our Proposal: DCC Tools

Overview

- Initial prototype available on https:// github.com/gmgeronymo/dcc_tools
- Implemented in Python
- Designed for cloud-based technologies (AWS Lambda)
- API for system integration

Future Plans

- Implement dccReader using PyDCC
- Extract measurement data as JSON or spreadsheets


Main Tools


- dccGenerator:
 - Generates PTB DCC compliant XML from JSON
- dccAddPdf:
 - Adds human-readable PDF to XML
- dccGetPdf:
 - Extracts human-readable PDF from **XML**


Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025 9 / 38

DCC Tools Diagram










Figure: DCC Tools diagram.



DCC Tools: Input Data Format Selection

Traditional Lab Workflows

- Common approaches:
 - Excel spreadsheets
 - Manual data entry
 - Standalone systems

Study Case Environment

- Modern lab infrastructure:
 - Web-based systems
 - SQL databases
 - LabVIEW automation
 - MS Access integration

JSON Format Benefits

- Already structured data
- Easy API integration
- Language-agnostic
- Lightweight format
- Better human readability than XML

Future Extensions

- Excel support planned
- Multiple input formats
- Custom data mappings

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025

DCC Tools: Simplified Data Structure Approach

PTB DCC Structure

- Comprehensive XML schema:
 - Complex structure
 - Covers all calibration types
 - Extensive metadata options
 - Generic requirements

Our Approach

- Lab-specific JSON schema:
 - Simplified structure
 - Custom field mapping
 - Minimal required data

Real-world Usage

- Labs typically use:
 - Specific certificate types
 - Limited data fields
 - Consistent structure
 - Repetitive patterns

Benefits

- Reduced complexity
- Easier implementation
- Better maintainability
- Lower error risk



JSON to DCC: Transformation Strategy

Implementation Strategy

- dccGenerator function:
 - Accepts simple JSON
 - Maps to DCC structure
 - Validates output
 - Ensures compliance

Customization Features

- Lab-specific templates
- Configurable mappings
- Flexible field definitions
- Extensible structure

Key Advantages

- Labs work with familiar data
- No need to understand full DCC
- Automated transformation
- Validated output

Future-proofing

- Easy schema updates
- New certificate types
- Additional field mappings
- Version compatibility

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025

Implementation Strategy

Initial Prototype Focus

- Target advanced labs:
 - Automated systems
 - Structured databases
 - API-ready infrastructure
- Quick implementation:
 - Minimal data transformation
 - Direct integration

Development Roadmap

- Phase 1: JSON API
- Phase 2: dccReader (JSON)
- Phase 3: Excel support (read and write)

Integration Benefits

- Seamless data flow
- Reduced processing time
- Lower error rates
- Simplified deployment



System Architecture

Serverless Design

- Core compute service: AWS Lambda
- Advantages:
 - Automatic scaling
 - Pay-as-you-go pricing
 - Reduced operational overhead

Suitable Use Cases

- Limited trusted clients
- Internal systems
- Efficient data submission and retrieval

Cost Efficiency

- Operates within AWS Lambda free tier limits:
 - 1M requests/month
- Cost-efficient for calibration labs
- No additional costs for typical usage

Lambda Function URLs

- Direct and efficient data handling
- Simplifies communication process
- Reduces API Gateway overhead

February 26th, 2025 15 / 38

AWS Lambda Serverless Architecture

What is Serverless?

- No server management needed
- Auto-scaling capabilities
- Pay-per-use model

AWS Lambda Features

- Function-as-a-Service (FaaS)
- Multiple language support
- Event-driven execution
- Native AWS integration

DCC Tools Implementation

- Direct function URLs
- Stateless architecture
- HTTPS encryption
- Free tier compatible



Alternative Deployment: Traditional Server Architecture

Adapting to Traditional Server Approach

The certificate generation service can be easily adapted from serverless to a traditional server architecture using Python FastAPI:

- Core business logic remains unchanged
- More predictable costs for consistent workloads
- Full control over runtime environment
- Simplified local development and debugging
- Multiple deployment options (servers, containers, platforms)
- Privacy requirements

February 26th, 2025 17 / 38

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

Example Calibration Certificate



Gean Marcos Geronymo (Inmetro - Brazil)

Python-based DCC Tools Development and Implementat




Constants



Variables







ser<mark>findiatria Bravileira de Componentes Eletrônicos (IRCE)</mark> erego:<mark>(Rua das Indústrias, 1234, Bairro Industrial, Cidade Tocnológica, RJ, CEP 12345-678</mark> Modelo/Tipo: 742A-1 Código de Identificação: RES-100-001 Laboratório Executor aborutório Responsável: Liberatório de N esponsável Técnico: Ream Murcos Gerony écnico Executor: Márcio Cândido da Silve uta da Calibração: [03122024]

Edson Afonso Chefe da Divisão de Metrologia Elétrica

THEFTIO

Características do Item Resistência Nominal: 1 Ω

Certificado de Calibração

DIMCI 2502/2024 Número do Certificado

Ds resultados da cultbração cudo contidos au Tubeia 2, onde V.m. é o valor medido. A incerteza de medições aprapridida (1) releadad foi declimada como a incentera pudido combinada multiplicada pois finir de abrungleria que, pur sum and instrutejões 6, com um interno reletivo de prusa de Brendade v., correspondo a um internolabilidade de abrungleria de VSA5 %. A incenteza de medições exposidada for relatada de accedo com a albalocação valunção. De Dados de Medições—Costa para Exposições de Investe au foranções—Costa Qual-

Gean Marcos Geronymo (Inmetro - Brazil)

Python-based DCC Tools Development and Implementat

February 26th, 2025

200



JSON Version of the Calibration Certificate (Part 1)

```
"nome_lab": "Laboratório de Metrologia em Padronização Elétrica",
"sigla_lab": "Lampe",
"nome_div": "Divisão de Metrologia Elétrica",
"sigla_div": "Diele",
"num_certif": "2502/2024",
"num_certif": "2502/2024",
"num_processo": "0052600.005271/2024-53",
"tipo_item": "Resistor Padrão",
"fabricante": "Fluke",
"modelo": "742A-1",
"num_serie": "SN-2023-RES-001",
"cod_identificacao": "RES-100-001",
"valor_nominal": "1",
"valnom_unit": "\\metre\\tothe{2}\\kilogram\\second\\tothe{-3}\\ampere\\tothe{2}",
"data_calibracao": "2024-12-10",
"cod.": true,
"data_calloracao"
"cmc": true,
"chefe_div": "Edson Afonso",
"chefe_lab": "Gean Marcos Geronymo",
"tecnico_executor": "Márcio Cândido da Silva",
"desc_tecnico_executor": "Técnico Executor",
"desc_tecnico_executor" : "lecnico Executor",
"cliente": {
  "nome": "Indústria Brasileira de Componentes Eletrônicos (IBCE)",
  "email": "contato@ibce.com.br",
  "cidade": "Cidade Tecnológica",
```

Part 1 of the JSON representation of the calibration certificate.

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025 21/38

JSON Version of the Calibration Certificate (Part 2)

```
"pais": "BR",
"cep": "12345-678",
"uf": "RJ",
"endereco": "Rua das Indústrias",
"numero": "1234"
              ),
"condicoes_ambientais": [
31
32
                              "name": "Temperatura no banho de ar",
"value": "22.94",
"unc": "0.11",
"unit": "\\degreecelsius"
33
34
35
36
37
38
                               "name": "Umidade relativa",
                               "value": "53.0",
"unc": "1.6",
"unit": "\\percent"
40
41
43
44
45
              ],
"rastreabilidade": [
                               "name": "Resistor Padrão",
"certificado": "DIMCI 0145/2024",
"cod_id": "TR 654"
47
```

Part 2 of the JSON representation of the calibration certificate.

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat



JSON Version of the Calibration Certificate (Part 3)

```
}
            ],
"resultados": [
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
66
67
                           "value": "1.00005512", "unit": "\\metre\\tothe{2}\\kilogram\\second\\tothe{-3}\\ampere\\tothe{2}", "unc": "0.25",
                "unctype" : "ppm",
"k": "2"
            ],
"condicoes_resultados": [
                           "name": "Corrente Aplicada",
"value": "0.003",
"unit": "\\ampere"
68
69
70
```

Part 3 of the JSON representation of the calibration certificate.

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025 23 / 38

XML Version of the Calibration Certificate (Part 1)

```
<dcc:digitalCalibrationCertificate xmlns:xsi="http://www.w3.org/2001/XMLSchema-instance"
    xmlns:dcc="https://ptb.de/dcc" xmlns:si="https://ptb.de/si" xsi:schemaLocation="https://ptb.de/dcc
    https://ptb.de/dcc/v3.2.0/dcc.xsd" schemaVersion="3.2.0">
     <dcc:administrativeData>
        <dcc:dccSoftware>
            <dcc:software>
              <dcc:name>
                  <dcc:content>dccGenerator</dcc:content>
               </dcc:name>
            <dcc:release>0.2</dcc:release>
</dcc:software>
        </dcc:dccSoftware>
        <dcc:coreData>
           cc:corePata>
<dcc:countryCodeIS03166_1>BR</dcc:countryCodeIS03166_1>
<dcc:usedLangCodeIS0639_1>pt</dcc:usedLangCodeIS0639_1>
<dcc:mandatoryLangCodeIS0639_1>pt</dcc:mandatoryLangCodeIS0639_1>
<dcc:uniqueIdentifier>DIMCI 2502/2024</dcc:uniqueIdentifier>
           <dcc:identifications>
  <dcc:identification>
                 <dcc:issuer>calibrationLaboratory</dcc:issuer>
<dcc:value>0052600.005271/2024-53</dcc:value>
                  <dcc:name>
                     <dcc:content>Processo Inmetro</dcc:content>
                  </dcc:name>
               </dcc:identification>
           </dcc:identifications>
            <dcc:beginPerformanceDate>2024-12-10</dcc:beginPerformanceDate>
           <dcc:endPerformanceDate>2024-12-10</dcc:endPerformanceDate>
                                                     Python-based DCC Tools Development and Implementat
Gean Marcos Geronymo (Inmetro - Brazil)
```



XML Version of the Calibration Certificate (Part 2)

```
<dcc:performanceLocation>laboratory</dcc:performanceLocation>
        </dcc:coreData>
<dcc:items>
31
           <dcc:item>
             <dcc:name>
                <dcc:content>Resistor Padrão</dcc:content>
34
35
36
37
             <dcc:manufacturer>
<dcc:name>
                  <dcc:content>Fluke</dcc:content>
38
39
                </dcc:name>
             </dcc:manufacturer>
<dcc:model>742A-1</dcc:model>
40
41
             <dcc:identifications> <dcc:identification>
                  <dcc:issuer>manufacturer</dcc:issuer>
44
45
                  <dcc:value>SN-2023-RES-001</dcc:value>
                  <dcc:name>
                     <dcc:content>Número de Série</dcc:content>
                  </dcc:name>
                </dcc:identification>
                <dcc:identification>
                  <dcc:issuer>customer</dcc:issuer>
<dcc:value>RES-100-001</dcc:value>
50
51
                     <dcc:content>Código de Identificação</dcc:content>
                  </dcc:name>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025

XML Version of the Calibration Certificate (Part 3)

```
</dcc:identification>
                  </dcc:identifications>
56
57
           </dcc:item>
</dcc:items>
59
           <dcc:calibrationLaboratory>
     <dcc:contact>
                 <dcc:name>
  <dcc:content>Laboratório de Metrologia em Padronização Elétrica(Lampe)</dcc:content>
                 </dcc:name>
                 <dcc:eMail>samci@inmetro.gov.br</dcc:eMail>
<dcc:phone>+55 21 2679 9077/9010</dcc:phone>
<dcc:location>
64
65
                    icc:location>
<dcc:city>Xerém - Duque de Caxias</dcc:city>
<dcc:countryCode>BR</dcc:countryCode>
<dcc:postCode>25250-020</dcc:postCode>
<dcc:state>RJ</dcc:state>
68
69
70
71
72
73
74
75
76
77
78
79
                    <dcc:street>Av. Nossa Senhora das Graças</dcc:street><dcc:streetNo>50</dcc:streetNo>
                  </dcc:location>
              </dcc:contact>
           </dcc:calibrationLaboratory>
           <dcc:respPersons>
              <dcc:respPerson>
                 <dcc:person>
                     <dcc:name>
                        <dcc:content>Edson Afonso</dcc:content>
                     </dcc:name>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat



XML Version of the Calibration Certificate (Part 4)

```
</dcc:person>
              </dcc:role>Chefe da Divisão de Metrologia Elétrica</dcc:role>
<dcc:mainSigner>true</dcc:mainSigner>
83
85
           </dcc:respPerson>
           <dcc:respPerson>
             <dcc:person>
87
89
                  <dcc:content>Gean Marcos Geronymo</dcc:content>
             </dcc:person>
<dcc:role>Chefe do Laboratório de Metrologia em Padronização Elétrica</dcc:role>
91
92
93
           </dcc:respPerson>
94
95
           <dcc:respPerson>
             <dcc:person>
96
97
                  <dcc:content>Márcio Cândido da Silva</dcc:content>
98
99
                </dcc:name>
             </dcc:person>
<dcc:role>Técnico Executor</dcc:role>
100
101
           </dcc:respPerson>
102
         </dcc:respPersons>
         <dcc:customer>
104
105
           <dcc:name>
             <dcc:content>Indústria Brasileira de Componentes Eletrônicos (IBCE)</dcc:content>
           </dcc:name>
<dcc:eMail>contato@ibce.com.br</dcc:eMail>
           <dcc:location>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025

XML Version of the Calibration Certificate (Part 5)

```
<dcc:city>Cidade Tecnológica</dcc:city>
                <dcc:countryCode>BR</dcc:countryCode>
111
112
                <dcc:postCode>12345-678</dcc:postCode>
<dcc:state>RJ</dcc:state>
                <dcc:street>Rua das Indústrias</dcc:street>
                <dcc:streetNo>1234</dcc:streetNo>
114
115
116
             </dcc:location>
           </dcc:customer>
           <dcc:statements>
             <dcc:statement>
                <dcc:declaration>
                  </dcc:declaration>
             </dcc:statement>
<dcc:statement>
                <dcc:norm>ABNT NBR ISO/IEC 17025</dcc:norm>
                <dcc:declaration>
                  <dcc:content>O presente certificado de calibração atende aos requisitos da norma ABNT NBR ISO/IEC
17025 e é válido apenas para o item acima caracterizado, não sendo extensivo a quaisquer
outros. Este certificado de calibração somente pode ser reproduzido em sua forma integral.
Reproduções parciais devem ser previamente autorizadas pelo Inmetro.</dcc:content>
                </dcc:declaration>
              </dcc:statement>
             <dcc:statement>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat



XML Version of the Calibration Certificate (Part 6)

```
<dcc:declaration>
131
132
             <dcc:content>Características do Item</dcc:content>
</dcc:declaration>
133
             <dcc:data>
               <dcc:quantity>
                 <dcc:name>
  <dcc:content>Valor Nominal</dcc:content>
135
                  </dcc:name>
                  <si:real>
                   <si:value>1</si:value>
                    <si:unit>\metre\tothe{2}\kilogram\second\tothe{-3}\ampere\tothe{2}</si:unit>
141
                  </si:real>
                </dcc:quantity>
           </dcc:data>
</dcc:statement>
143
         </dcc:statements>
145
      </dcc:administrativeData>
147
      <dcc:measurementResults>
        <dcc:measurementResult>
149
           <dcc:name>
             <dcc:content>Resultados e Declaração da Incerteza de Medição</dcc:content>
151
           </dcc:name>
           <dcc:usedMethods>
153
             <dcc:usedMethod>
154
155
               <dcc:name>
                 <dcc:content>Declaração da Incerteza de Medição</dcc:content>
                <dcc:description>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025

XML Version of the Calibration Certificate (Part 7)

```
158
                       <dcc:content>As incertezas expandidas de medição (U) relatadas são declaradas como a incerteza
                              padrão combinada multiplicada pelo fator de abrangência k, que, para uma distribuição t, com um número efetivo de graus de liberdade \nu eff , o qual corresponde a uma probabilidade de abrangência de 95,45 %. A incerteza de medição expandida foi relatada de acordo com a publicação Avaliação de Dados de Medição - Guia para a Expressão de Incerteza de medição - GUM 2008.</d>
                 </dcc:description>
</dcc:usedMethod>
                 <dcc:usedMethod>
                    <dcc:name>
                       <dcc:content>Método de Medição</dcc:content>
                    </dcc:name>
165
                    <dcc:description>
                       <dcc:content>O resultado fornecido refere-se ao valor médio de seis séries de trinta medições
                              pelo método de comparação de corrente na configuração 04 (quatro) terminais. Utilizou-se uma Ponte Automática de Resistência modelo 6010D.</dc:content>
167
                    </dcc:description>
                 </dcc:usedMethod>
169
              </dcc:usedMethods>
              <dcc:measuringEquipments>
                 <dcc:name>
                    <dcc:content>Os resultados da calibração são rastreados ao Sistema Internacional de Unidades (SI),
                           por intermédio dos padrões metrológicos nacionais. As medições realizadas estão referenciadas aos padrões relacionados.</dc:content>
                 </dcc:name>
                 <dcc:measuringEquipment>
                    <dcc:name>
                        <dcc:content>Resistor Padrão</dcc:content>
                    </dcc:name>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025



XML Version of the Calibration Certificate (Part 8)

```
<dcc:certificate>
                   <dcc:referral>
                     <dcc:content>DIMCI 0145/2024</dcc:content>
                   </dcc:referral>
<dcc:referralID>TR 654</dcc:referralID>
181
183
                   <dcc:procedure>analog</dcc:procedure>
                   <dcc:value>analog</dcc:value>
                </dcc:certificate>
              </dcc:measuringEquipment>
           </dcc:measuringEquipments>
<dcc:influenceConditions>
              <dcc:influenceCondition>
<dcc:name>
                <dcc:content>Temperatura no banho de ar</dcc:content></dcc:name>
                <dcc:data>
                   <dcc:quantity>
                    <si:real>
  <si:value>22.94</si:value>
                       <si:unit>\degreecelsius</si:unit>
<si:expandedUnc>
198
199
                         <si:uncertainty>0.11</si:uncertainty>
                          <si:coverageFactor>2</si:coverageFactor>
                       <si:coverageProbability>0.9545</si:coverageProbability>
</si:expandedUnc></si:</pre>
                     </si:real>
                   </dcc:quantity>
                </dcc:data>
              </dcc:influenceCondition>
 Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025
```

XML Version of the Calibration Certificate (Part 9)

```
<dcc:influenceCondition>
                   <dcc:name>
209
210
                      <dcc:content>Umidade relativa</dcc:content>
                   </dcc:name>
211
212
                   <dcc:data>
                     <dcc:quantity>
                        <si:real>
  <si:value>53.0</si:value>
                          <si:unit>\percent</si:unit>
<si:expandedUnc>
                             <si:uncertainty>1.6</si:uncertainty>
<si:coverageFactor>2</si:coverageFactor>
                           <si:coverageProbability>0.9545</si:coverageProbability>
</si:expandedUnc></si:</pre>
                     </si:real>
</dcc:quantity>
                </dcc:data>
</dcc:influenceCondition>
             </dcc:influenceConditions>
<dcc:results>
                <dcc:result>
                     <dcc:content>Resultados</dcc:content>
                   </dcc:name>
                  <dcc:data>
<dcc:quantity>
                        <si:real>
                           <si:value>1.00005512</si:value>
```

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat



XML Version of the Calibration Certificate (Part 10)

```
\label{linear_second_tothe} $$ \sin : \frac{2} \left( \frac{-3}{ampere \cdot tothe \{2\} \cdot \sin i} \right) $$
236
237
                       <si:expandedUnc:
                          <si:uncertainty>2.50E-07</si:uncertainty>
                         <si:coverageFactor>2</si:coverageFactor>
<si:coverageProbability>0.9545</si:coverageProbability>
                       </si:expandedUnc>
                     </si:real>
                     <dcc:influenceConditions>
                       <dcc:influenceCondition>
                         <dcc:name>
                            <dcc:content>Corrente Aplicada</dcc:content>
                         </dcc:name>
                            <dcc:quantity>
                              <si:real>
                                <si:value>0.003</si:value>
                                <si:unit>\ampere</si:unit>
                              </si:real>
                            </dcc:quantity>
                         </dcc:data>
255
                       </dcc:influenceCondition>
                     </dcc:influenceConditions>
                   </dcc:quantity>
                </dcc:data>
              </dcc:result>
            </dcc:results>
260
         </dcc:measurementResult>
       </dcc:measurementResults>
     </dcc:digitalCalibrationCertificate>
  Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025
```

Example - embedded PDF

```
</dcc:quantity>
                                                           </dcc:data>
                                                    </doc:influenceCondition>
                                               </dcc:influenceConditions>
                                         </dcc:quantity>
                                     </dcc:data>
                              </dcc:result>
259
                         </dcc:results>
260
                    </dcc:measurementResult>
261 </dcc:measurementResults>
262 <dcc:document><dcc:fileName>CC_DIMCI_2502_2024.pdf</dcc:fileName><dcc:mimeType>application/pdf</dc
          {\tt c:mimeType}{\small <} {\tt dcc:dataBase64}{\small >} {\tt JVBERi@xLjcKJcOkw7zDtsOfcjigMCBvYmoKPDwvTGVuZ3RoIDMgMCBSL@ZpbHRlci9GbGF}
          \\ \emptyset ZURIY29kZT4+CnN\emptyset cmVhbQp4nM\emptyset byYr1RvJeXyHwzVByRkZmSoLiwVtnbPDB44Y5mDk\emptyset 3k23PW5s/PsTW6ZSUuq9avBhaKpaSVaLiwVtnbPDB44Y5mDk\emptyset 3k23PW5s/PsTW6ZSUuq9avBhaKpaSVaLiwVtnbPDB44Y5mDk0 3k23PW5s/PsTW6ZSUuq9avBhaKpaSVaLiwVtnbPDB44Y5wDk0 3k24PW5wDk0 3k24PW5
          8aSsUekyvXQ/fX02fHDHz//8PbbP7rTl+en3zvXud75sYtT7P0QuzFAPyboPnz/909Pu1+frrSqAUP/Pvz4dHrz5KPvfeyGYej
          DOHVvvus+u0E3dm9++ObFwSG800/w4F9ccPHwTP8nN7iRH0xye3Qnd5areMAXd5G3z3R1VdibwD+HF3C07j9vvni6vnn6asUEp
          K1HYiKNvV8zAXAw50CVBUAI9AyiI1bCmezN2IMkTH1A+vGHxBsYjFPbCqwHWeEFnRtAMQj0xJfA7B+NnEpBAR2IBGQ9HIU8IeN
          NR112WTL1VWKgLG3kSTwK0RP/AsGTeL9TwQ/7ohtSn0h0cerdRnQTI/R5A8q7i05C/z0bg3R71V0VtgnGmayE8ZvsgJkBFZkJh
          KXPQhWYW0M43ik1V9vIjW4LRpONz7K5kNC3KCv5mGay8nQfqArcldIQeohdYlmhCcl3wZuVs0pYH45sBohJFoubxLQneneiFaM\
          7My/AqiVtuStbxh61AGOPXQrj7FLYQcg+BS9KUC+iehT9ngR/kusFbhDc0C3xfvv+CfrQ/WWk//WPJ9c9e1rwvgM39IPevOu+3
          uMPh9lksJuUvdMhvrCyAu9UfJjVFQH4NwmJNq4WRT8REj+fWX0UbMLUU4gC2kDEla16T7TEUsUoDs9DEBsaxJvUZ+wlG5HHcuk
          uYk/DDDKSH64QsEnjInyY47eoiTMPviBzgPUK8kgxSh9mlzGbrhe04jcLQPUed8uu72PNtl9wMprnL1znVGLpBla9Q1mD+f187
 4S8LMF1ScP16LBEN5aIe1HvcPHicLelMEZesmSDx0JHmDcpBNNYbep5qEpCC2q/GcvqLr1Sqwh0x+liSL1rGhdDy/m16ahqiqiy
Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025 34
```



Cryptographic Signatures

Digital Signatures in DCC

- DCC XML can be digitally signed
- Ensures trust between institute and customer

Common Practice

- Adding digital signatures to XML is common
- Various free tools available

Assinador Serpro

- Free tool by a federal state-owned company in Brazil
- URL: https://www.serpro.gov.br/ links-fixos-superiores/ assinador-digital/ assinador-serpro
- Requires a valid digital certificate from a certification authority

Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat

February 26th, 2025 35 / 38

Conclusion

Pragmatic Solution

- Efficient DCC handling via cloud-based serverless application
- Tools offer potential for enhancement and customization

Digitalization Benefits

- Seamless integration into calibration workflows
- Supports genuine digitalization of processes

Productivity Gains

- Adoption of DCCs eliminates manual transcription tasks
- Improves productivity and reduces errors



Thank You!

Thank You for Your Attention!

Contact: **G** M Geronymo

gmgeronymo@inmetro.gov.br



Gean Marcos Geronymo (Inmetro - Brazil) Python-based DCC Tools Development and Implementat February 26th, 2025 37 / 38

Back to "Table of Contents" above



P05 DCC in production mode

Presenting author: Thomas Engel, Siemens AG, Germany

E-mail address: engelthomas@siemens.com

Additional authors: Klaus Fickinger, Sami Koskinen, André Lemke, Christian Sander, Shanna

Schönhals

Abstract

In Summer 2024 a group of DCC enthusiast have teamed up to finalize, apply and validate the DCC for applications for a real-world application scenario of a temperature sensor in pharmaceutical industry. The team covers all stakeholders along the process chain, from system owner (Boehringer Ingelheim) to industry partners as instrument manufacturers, calibration service providers, calibration management systems and process automation and control solutions manufacturers (Beamex, Endress & Hauser, Siemens, Testo) and the German national metrology institute PTB as DCC development partners. The project initiative was structured in three phases.

Phase 1: DCC generation and DCC harmonization among partners and alignment with DKD good practices

Phase 2: Development and specification of a digital calibration request (DCR) to specify all requirements

Phase 3: DCC exchange and application in the asset administration shell, digital product passport and dataspaces

After 5 months, phase 1 is close to completion. The team will present the requirements and benefits from a sensor system owner perspective and share detailed information and lessons learnt when implementing a full process chain for DCC generation in sensor system production and recalibration. The goal of phase 1 was to have a harmonized content of the DCC irrespective of the issuer of the DCC as system manufacturer or calibration service provider. The DCCs was transferred to the calibration management system (CMS) and all calibration information extracted and automatically entered into the CMS. Finally, the CMS data was evaluated and validated by the sensor system owner with respect to completeness and correctness as endpoint of phase 1.

The presentation will be concluded by an update of recent and up-coming DCC schema developments and a perspective on the integration of DCCs into the concepts and systematic of AAS and the digital product passport as they constitute the futures system of digital asset representation and documentation.



Back to "Table of Contents" above

Presentation of Thomas Engel



Boehringer Ingelheim Situation











beamex PTB SIEMENS





Introduction and Role of Instrument Manufacturer
















Current situation at Instrument Manufacturer





The DCC is powered by: beamex PB SIEMENS Ingelheim Endress+Hauser












DCC creation in Production





Role of Calibration Software Solution Provider



BACKGROUND

- + Beamex Calibration Solution has been in place at Boehringer Ingelheim 10+ years
- + Solution is widely used within BI manufacturing network world wide
- + BI has been managing calibration certificates from instuments and service providers manually. Only one copy of the paper and/or PDF certificate has been available. They both are designed human readably only.
- + Vendors paper/PDFs are all in different formats and speak different "language" → no harmonization nor labeled data available.
- + With DCC we are able to get vendor's calibration data as part of the existing solution and utilize data for analysis and process optimization













Role of Calibration Software Solution Provider



CHALLENGES

- + In this case Calibration solution is part of the enterprise level ecosystem and therefore changes to existing processes need to be planned carefully
- Digital Calibration Request (DCR) is a crucial part of future development to fully automate new digital processes with vendors
- + Current calibration solution is running on premises which requires additional configuration and/or external cloud based services (DCC Hub) to manage DCR/DCC coming outside of existing enterprise ecosystem

NEW SOLUTION

- Receive DCC from vendors and import harmonized and labeled data to existing calibration solution
- New process provides better data quality by avoiding media breakes between BI and their calibration vendors
- + Vendor's calibration data can be used for data analysis and future process improvements














Recalibration of sensors using DCC















Endress+Hauser 🔠

system
















The DCC is powered by: beamex SIEMENS (in Boehringer Ingelheim Endress+Hauser









DCC Production Mode The Game Changer













The DCC is powered by: DEBMEN SIEMENS Boehringer Endress+Hauser



DCC Production Mode The Game Changer





Ingelheim





Meeting 15.01.2025 Gerlingen

DCC Temperature Manufacturer and Service Provider Ready to use in Production Mode











Key Facts about the Digital Calibration Certificate (DCC) XML Schema







XML-based exchange format for calibration certificates



Developed with international partners including industry



Automated Validation via XSD-Schema and Schematron




Complies with many important norms and standards



Endress+Hauser





Evolution of the DCC Schema



- + Version 3.X.X longtime available
- + Parallel development of modular (XSD) Digital Schema X (DX)
- + Basis for DCC v4.X, Digital Calibration Requests (DCR), and further Digital Quality Documents
- + Definition of semantic model (harmonised terms and definitions, ontology)















Digital Quality Documents & Asset Administration Shell





DQD sub-model of the Asset Administration Shell





Digital Quality Documents characterize the performance of a system / product over its entire life span via the following documents by recurring tests over time:

- Calibration (DCC)
- Conformity assessment (DCoC)
- Legal metrology (DCC, DCoC)
- Test reports (DTC)
- > Acceptance tests
- > Field Tests
- Check lists e.g. function / completeness




Boehringer Endress+Hauser



DQD in Dataspaces





The value of a digital quality infrastructure







Presenters and Contacts + Boehringer Ingelheim Klaus Fickinger klaus.fickinger@boehringer-ingelheim.com + Endress & Hauser André Lemke andre.lemke@endress.com + Beamex Sami Koskinen Sami.Koskinen@beamex.com CSander@testotis.de + Testo Christian Sander + Physikalisch Technische Bundesanstalt Dr. Shanna Schönhals shanna.schoenhals@ptb.de + Siemens Dr. Thomas Engel engelthomas@siemens.com Thank you very much for your attention. Questions welcome The DCC is powered by: beamex SIEMENS Boehringer Endress+Hauser

Back to "Table of Contents" above

Day 2 (2025-02-26 Wednesday)
Parallel Session "Tools & Software"
(Chair: Anjali Sharma)

O11 Digital Calibration Exchange (DCX): Software Tools, Interfaces, and Workflow Integration

Presenting author: David Balslev-Harder; DFM – Danish National Metrology Institute

Email: dbh@dfm.dk

Additional authors: Søren Kynde srk@dfm.dk

<u>Abstract</u>

The Digital Calibration Exchange (DCX) framework was developed by DFM with input from members of TC-IM 1448 WG1 and WG2. The aim of DCX is to investigate the impact on software-tools and architectures when the schema emphasizes use of reference pointers within the DCC's, as compared to tree-structure emphasis as applied in the PTB DCC-schema. This presentation will explore these developed software-tools and discuss aspects relevant to user-interface. In particular software tool implemented using Python and Excel, which allows users to read, edit and export data to xml, will be presented.

Additionally, we will introduce a query tool designed for clients to efficiently lookup specific data within Digital Calibration Certificates (DCCs) which the client may want to transfer to internal databases. This tool simplifies the process of preparing the data retrieval, to enable quick and accurate access to essential calibration information.

The presentation will also cover the generation of PDF documents from DCX data, emphasizing the use of human-readable headings for all elements. This feature enhances the readability and usability of the generated documents, making them accessible to a wider audience. Finally, we will discuss the workflow and architecture of DCX in relation to the XSD



schema, providing a comprehensive understanding of how these components interact to ensure data integrity and consistency. In conclusion, the DCX user-interfaces and software tools provide important steps toward offering a powerful and intuitive solution for managing calibration data, streamlining processes, and enhancing the overall user experience.

Outline of presentation

- What is DCX
- Availability of software tools for DCX
- o User interface for DCX for: reading, editing and eporting to xml
- o Querry tool for clients to lookup data in xml
- o PDF generation utilizing Human readable headings for all elements.
- Workflow and architecture relation to DCX schema
- Solution available on github.com/TC-IM-1448/DCC-Tables
- See the other presentations for information on DCX schema and DCX certificate examples.

DFM's work is supported by funds from The Danish Agency for Institutions and Educational Grants.

Back to "Table of Contents" above

Presentation of David Balslev-Harder

Getting Started with Digital Calibration Exchange (DCX): Software Tools, Interfaces, and Workflow Integration

5th international DCC conference,

February 26th 2025

By David Balslev-Harder & Søren R. Kynde






DFM's incentive for DCC:

To support creation of FAIR DCC formats that are globally recognized/standardized, in order to support quality and efficiency in production companies, through automation.

Develop simple user interface software tool that can accept and modify DCC, DCR and Templates regardless of origin.

DFM receives funding by: Danish Agency for Institutions and Educational Grants






Motivation for DCX:


- 1. Explore impact on middleware architecture when schema is designed w. focus on:
 - a) Use of references rather than emphasis on tree structure (PTB-DCC approach):
 - i. Extend use of xs.ID and xs.IDREF and utilizing xs.key and xs.keyref to restrict pointing options
 - ii. Introduce new attribute-names and restricted options to guide machine interpretability
 - iii. avoid multiple re-entry of identical data
 - b) Represent data content in tables for easier integration with Excel and databases and provide better overview in user interface experience.
 - c) Emphasis on descriptive query expressions for finding and interpreting measurement data.
 - d) Same elements/structure for all calibration/metrology areas.
 - e) Define as much as possible in one Schema, to simplify interoperability and facilitate validation data machine interpretability and ability level to support automatic data extraction.
- 2. DCC is more than just a data-file, it may act as a legal document.
 - a) Support human readable fields for all elements in the xml-file.
 - b) Support PDF-generation w. low level of middleware assumptions.
- Simplify software/middleware management
 - a) Use schema in middleware to adapt for received xml's used schema version.
 - b) Provide tools for Excel that support file reading across different metrology area's

dbh@dfm.dk


Motivations were developed through discussions with stakeholders and within TC-IM 1448 WG












Updates since last year

- 1. New default system for Quantities and Units based on ISO 80000 and BIPM SI-Digital-Framework unit representation introduced in schema including:
 - Incorporated in the data result tables.
 - RegEx for validating BIPM-SI unit expression.
 - Section for defining other quantities and units.
- Simplified and improved query expressions for extracting measurement data and meta data.
- 3. Introduced human readable fields for all elements.
- 4. Simplified the layout of administrative data
- 5. Introduced DCX (Digital Calibration Exchange) as name to distinguish from PTB-DCC schema.
- 6. DCR and Templates incorporated in the same schema.
- 7. User interface builds up based on the used schema, enabling dynamic adoption between schema versions.






calibrationResults and measurement table structure new Quantity Unit lines added


17	@scope	reference	indication	bias
18	dataCategory	value	value	value
19	@dataCategoryRef	-	-	-
20	@quantity	5-2 Celsius temperature °C	5-2 Celsius temperature °C	5-2 Celsius temperature °C
21	@unit	°C	°C	°C
22	@quantityUnitDefRef			
23	@settingRef			
24	heading[da]	Referencetemperatur i instru	Instrumentvisning	Visningsfejl
25	heading[en]	Reference Temperature in in	Instrument indication	Instrument Error
26	1	-6.076	-6.06	0.016
27				
	2	5.017	5.05	0.033
28	2	5.017 27.054		10,000,000
	-			


<xs:restriction base="xs:string">

 $\label{eq:constraints} $$\operatorname{den}_{X}^{*}(X) = \operatorname{den}_{X}^{*}(X) +














































Xpath Query for measurement data.

Xpath-Color-code:


- Orange: attributes keywords defined in schema
- Purple: keywords defined in schema which can be selected to identify data.
- Green: keywords specific for the certificate needed to identify data.


Xpath:

- $*//dcx: calibration Result [@tableId="Table1"]/* [@scope="bias"] [@quantity="3-4|volume|m3"] [@unit="\mu L"]/dcx: value/* [@idx="1"]/text() = 10.5$
- *//dcx:calibrationResult [@tableId="Table1"]/*[@scope="bias"][@quantity="3-4|volume|m3"][@unit="µL"]/dcx:conformity/*[@idx="1"]/text() = 'pass'

dbh@dfm.dk









Process for Pilot study of DCX v1.0.0

- 1. Client map their data-base information needs to DCR format.
 - 1. if no DCR we experience low client stakeholder engagement.
- 2. Create mapping for preparing extraction of data from DCC to client DB.
- 3. Tender and PO process will run in parallel
- 4. Can the relevant data be extracted reliably from the different providers.
- 5. Certificates will be generated by several Service Providers to test if they return DCC's with equivalent content, structure and formats.

dbh@dfm.dk

Path forward

- 1. We envision the presented tools as a tool for
 - establishing templates for specific kinds of calibration tasks in different metrology areas, such that both clients and service providers can automate processes based on the templates.
 - 2. Provide simple-tool for small Service Providers so they can receive DCR, and generate DCC's.
- 2. With PTB investigate how the DCX requirements can be merged/harmonized with DCC, TemaTres
- 3. In TC-IM 1448: Harmonize semantics and structures, international perspectiv
- 4. Establish standardized regular expression based on BIPM Digital SI unit expression and a standardized schema for ISO 80000 through TC12.
- Incorporate NCSLI metrology Taxonomy into settingCodes.
- 6. Pilot study and Full process cost analysis (Template, DCR, DCC, extraction)
 - 1. Cost of deployment, maintenance, versatility, for both clients and Service Providers.

lbh@dfm.dk



DFM's incentive for DCC: Members in TC-IM 1448 WG1 and WG2: To support creation of FAIR DCC formats DFM, Denmark: David Balslev-Harder (Chair) that are globally recognized/standardized, Zoltan Zelenka in order to support quality and efficiency in Diego Nahuel Coppa, INTI, Argentina: IPQ, Portugal: Carlos Pires production companies, through NPL, UK: Michael Chrubasik, Ian Smith automation. NPLI. India: Aniali Sharma Shanna Schönhalls, Daniel Hutzschenreuter, PTB, Germany: Gamze Soeylev-oektem, Clifford Brown, Develop simple user interface software tool that can accept and modify DCC, DCR and Quentin Baire. Templates regardless of origin. SMD. Belgium: TI, Denmark: Jonas Vind TUBITAK, Turkev: Erkan Danaci Danish collaboration partners: **Collegues from PTB:** Shanna Schönhalls, DANDIAG: Tove Søgaard, Rene, Charlotte Rugaard Thomas Krah Novo Nordisk: Heidi Foldal, Jan Laursen, Johan Schroll-Flaischer Daniel Hutzschenreuter FORCE Technology: Ahmed Khan Leghari, Andreas Lyndrup Jensen, ... Muhammed-Ali Demir Gamze Soeylev-oektem Teknologisk Institut: Jonas Vind, Peter Friis Østergaard, Jan Nielsen, ... Benedikt Seeger DANAK: Erik Øhlenschlæger Clifford Brown Wiebke Heeren Siegfried Hackel Contact: David Balslev-Harder (dbh@dfm.dk) Frank Härtig ...

Back to "Table of Contents" above

O12 GEMIMEG 2.0: Evolution of the concept and implementation of the GEMIMEG Tool

Presenting author: Alexander Reissert Email: alexander.reissert@ptb.de

Additional Authors: Samuel Eickelberg samuel.eickelberg@ptb.de, Petra Tsesmetzi

petra.tsesmetzi@ptb.de, Abdul Rehman abdul.rehman@ptb.de

Abstract

GEMIMEG II is a successor of the GEMIMEG tool [1] that strives to fulfil the purpose and goals of the predecessor. GEMIMEG II is designed to provide an accessible graphical user interface (GUI) for human users to create a Digital Calibration Certificate (DCC) without requiring prior knowledge of the underlying XML structure nor the DCC schema. It continues the successful work of the previous software tool. Furthermore, it produces a human readable format of the DCC, to facilitate the work of auditors and streamlining accreditation procedures.

GEMIMEG II supports for the first time the DCC schema version 3.3.0 and the support of the Digital System of Units (D-SI) [2] and its corresponding interlinked schema. This improves, e.g., the expanded measurement uncertainty and its correct notation in the digital realm.

GEMIMEG II is provided as a webservice and can be easily accessed via an Internet browser like a web page. A desktop version for offline use will be provided for all major operating systems. By employing container technology, GEMIMEG II is independent of the underlying operating system.

Additionally, GEMIMEG II provides a standardised application programming interface (API). This facilitates the seamless integration of GEMIMEG II into an existing laboratory workflow [3]. Implementing standardised REST interfaces enable a seamless communication and data exchange of upstream process steps and specialised software tools. This provides the opportunity to employ GEMIMEG II without media discontinuity. Furthermore, laboratories can create a fully automatised DCC creation workflow in their customised working environment.



GEMIMEG II is published under an Open-Source license and will be free of charge. This allows to use the tool without restrictions including the right to copy, to change, to publish, to distribute, to lend or rent it, or use it commercially.

- [1] Machine-readable and machine-actionable digital calibration certificate (DCC) and its use in practice S. Hackel, S. Schönhals, T. Krah, L. Doehring – Technisches Messen 2024; 91(1): 51–62, DOI: https://doi.org/10.1515/teme-2022-0117
- [2] SmartCom Digital System of Units (D-SI) Guide for the use of the metadata-format used in metrology for the easy-to-use, safe, harmonised and unambiguous digital transfer of metrological data - Second EditionHutzschenreuter, Daniel, et al., *Zenodo. DOI* 10 (2019).
- [3] <u>Digital transformation in metrology: Building a metrological service ecosystem</u> A
 Oppermann, S Eickelberg, J Exner, T Bock... Procedia Computer Science, 2022

Back to "Table of Contents" above

Presentation of Alexander Reissert





General digital transformation in the lab | Prolog




- Avoid proprietary data formats and platform with export functionalities
- Embrace the FAIR principles
- Make data and algorithms available between labs, groups and disciplines











26. Feburary 2025

Trustworthy machine-readable information | Prolog



Image: PD (https://nzz.ch)


Providing exactly the information that is needed with a high level of confidence

- Interfaces between platforms
- Machine-readable data
- · Protection against manipulation
- · High level of automation


PIB 5th International DCC Conference Dr. Alexander Reissert



Integration Areas | Prolog




Integration: Systems, Data and Processes | Prolog





FAIR principles – not only in science! | Prolog



- Uniquely findable data is the pre-requisite for application of machine learning and Al
- Data must be accessible by other systems in the workflow
- Interoperability is required to use the data in different contexts
- Re-using data may lead to new insights



Core metrology competences underpin the FAIR principles

5th International DCC Conference Dr. Alexander Reissert

```
GEMIMEG Tool 2.0

| Gemimed to the street of ```



### Gemimeg Tool 2.0 | Presentation

#### Backend

- Java / Spring Boot Framework
- REST Interfaces
- OpenAPI Documentation (Swagger)
- Container Technology (Docker)
- Support DCC schema version 3.3.0

Fully functional service unit to create and issue DCC



26. Feburary 2025

## Gemimeg Tool 2.0 | Presentation

#### Frontend

- Angular Framework
- Web Application
- Container Technology (Docker)

Graphical User Interface for the Backend Service





### GEMIMEG Tool 2.0 | Presentation

- Connect GEMIMEG Tool 2.0 now to your service infrastructure through REST APIS
- Or run GEMIMEG Tool 2.0 as a standalone tool
- Create fully digitalised workflows with many different data sources
- Create a DCC via JSON and receive a fully functional DCC XML



10


26. Feburary 2025


### Gemimeg Tool 2.0 | Presentation

- Open-Source License
- Public Beta Test will start at 31. March 2025
- Multilanguage Support more localisation needs your help
  - English,
  - French,
  - German,
  - Spanish and
  - Portuguese
- We need your feedback @ gemimeg@ptb.de







































5th International DCC Conference Dr. Alexander Reissert

26. Feburary 2025



Sth International DCC Conference
Dr. Alexander Reissert

26. Feburary 2025

21



## Gemimeg Tool 2.0 | Presentation

AVAILABLE AT 31. March 2025

VISIT: https://www.gemimeg.ptb.de/gemimeg-tool/2.0



```
GEMIMEG Tool 2.0

Use Case

Use Case

If a previous Elements length; i < ii; ++i) {


Interest = selected Elements[i];

Interest = selected Elements[i];

Interest = selected;

I
```





## Questions? | Discussion



## Thank you for your attention









Operation – Layer Team

AVAILABLE AT 31. March 2025

VISIT: https://www.gemimeg.ptb.de/gemimeg-tool/2.0







Back to "Table of Contents" above

#### O13 Formula Evaluation and MathML: Processing in Python: A High-Precision Framework for XML documents like the DCC

Presenting author: Justin Jagieniak, PTB Germany

Email: Justin.Jagieniak@ptb.de

Additional authors: Siegfried Hackel, Jan Loewe


#### **Abstract**

This presentation introduces a Python library designed for processing and evaluating mathematical expressions with high precision, leveraging sympy library. A key feature of this library is its ability to parse and process XML-based MathML content, enabling the extraction and evaluation of mathematical formulas represented in MathML format. The library supports dynamic variable management, symbolic expression handling, and high-precision numerical computations, making it suitable for applications in scientific computing. By integrating MathML processing capabilities, the library bridges the gap between symbolic mathematics and XML-based mathematical representations, providing a versatile tool for working with complex mathematical data. The presentation will highlight the library's core functionalities, its use cases, and its advantages in scenarios requiring both high computational precision and interoperability with MathML standards for the DCC.



#### Back to "Table of Contents" above

### Presentation of Justin Jagieniak








#### Workflow


The transformation of calibration into an XML document






## The Utility Model Formulas in the utility model by IDIS





From: "IDIS – Initiative Digitale Standards". [Online].
see: https://www.dke.de/resource/blob/2076816/facc9bde1806e2194a3d26a60c79bf77/idiswhitepaper\_1/en\_, download-data.pdf an Loewe | Segfried Hadel | Automating Formula Evaluation in Digital Calibran Certificates






#### Formula in Content MathML

An example how a formula is written in Content MathML









#### Modules of the MathML Evaluator

The three modules of the python program



- The program is divided in 3 modules:
  - XMLFormulaExtractor extracts the formulas of an XML like DCC.
  - ContentMathMLEvaluator evaluate the extracted formula and convert it into a Python algorithm.
  - GUI / Terminal A GUI or Terminal for work with.

https://www.sympy.org | Justin Jagieniak | Jan Loewe | Siegfried Hackel | Automating Formula Evaluation in Digital Calibian Certificates (DCCs) with Python



#### Problem fields for metrology

Identify, analyze and resolve problems



- A formula is not an algorithm. The polish notation of content MathML needs to be interpreted step by step and converted.
  - Resolution 1: The XML needs to be interpreted in a way that is understandable for an algorithm.

    This resolution was developed by us.
  - ▶ Resolution 2: SymPy¹ (open source library) offers many advantages for symbolic mathematics.

    The formula can be converted into an algorithm using a function pointer that references it.
- Values for variables (measurement results) need to be extracted by id and xrefs in the DCC.
  - Resolution: We have achieved this with our own solution in the XMLFormulaExtractor.
- 1. https://www.sympy.org

ustin Jagieniak | Jan Loewe | Siegfried Hackel | Automating Formula Evaluation in Digital Calibration Certificates (DCCs)twPython

#### Problem fields for metrology 2




Identify, analyze and resolve problems

- Normally programming languages use limited data types like float, double, etc.
  - The precision might not be sufficient for specialized fields of metrology
    - ▶ Resolution: SymPy works with a library called mpmath². With mpmath, SymPy can achieve a precision of up to 100 decimals.
- Units of variables needs to be calculated too.
  - ▶ **Resolution**: Units in formulas can also be calculated. This functionality is supported inSymPy. The feature is currently in progress and is planned for future implementation.
- 1. https://mpmath.org











#### Future plans

Summary



#### Distribution

- A REST API hosted by PTB for working with the latest developments from the DCC team at PTB.
- A free-to-use Docker image for running the REST API locally or on your own server.
- A Python library to integrate into your own development projects.

#### Features

- Integrate the SI system of units into the code to handle units in formulas as well.
- Integrate more advanced mathematical functions that can be used in Content MathML.





Back to "Table of Contents" above

## O14 Consistent Representation and Usage of Tabulated Data in the DCC. Enable Easy-To-Use Computer-Aided Processing and Visualization of Your Data

Presenting author: Benedikt Seeger Email: benedikt.seeger@ptb.de

Additional authors: Eva Vanessa Stehr vanessa.stehr@ptb.de, Thomas Bruns

thomas.bruns@ptb.de

#### **Abstract**

In metrological measurements, tabulated data is omnipresent - almost all data consists of different values that stand in relation to each other. The Digital Calibration Certificate (DCC) presents us with a unique opportunity to make this data accessible more efficiently than ever. The accessibility of the data enables direct processing of the data, e.g. as human-readable plots, such that the data becomes easy to understand and interpret correctly. The current state of the DCC schema definition allows for multiple different representation of tabulated data. For efficient use of the DCC as a universal exchange format for metrological data between different applications, it is crucial that the table format is consistent across all DCCs, regardless of calibration task. One such format is proposed in which all kind of tables can be represented, and which can then be efficiently parsed by suitable software.

It is discussed how computers process relational data and how modern design principles enable efficient interaction with and consistency of such data. The different possible structures of the DCC that could represent the relations of the data are discussed. A suitable, efficient format for representing tabulated data is proposed, and its efficiency and reliability is evaluated. We present an open-source reference implementation project in Python to show the underlying de- sign principles that could analogously be used in all object oriented programming languages.

Keywords: Tables; DCC; Indexing



#### Back to "Table of Contents" above


## Presentation of Benedikt Seeger



# **Consistent Representation and Usage of Tabulated Data in the DCC**

5th DCC Conference

Benedikt Seeger, Vanessa Stehr, PTB, 1.73



## **Outline**



- 1 Data in the DCC
  - 1 Tables
  - 2 dcc:quantities
- 2 Representation options of this data
  - 1 Motivation
  - 2 dcc:quantities
  - 3 Tables

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

26.02.2025



## Why we use tables



- Calibration: Establishes a relationship between known quantity values (with uncertainties) and corresponding indications under specified conditions. JCGM 200:2008 International vocabulary of metrology
  - Tables are used in analog and digital domain to store calibration data efficiently.
  - Relational databases and tables share the same basic principles
    - They store "rows" with identical kind and structure of information
    - We look up a specific "row/dataset" by searching for one (or more) specific value

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

26.02.2025

# Physical quantities and tables



|          | Table Columns are |                                                      |                                    |                                 |    |  |  |
|----------|-------------------|------------------------------------------------------|------------------------------------|---------------------------------|----|--|--|
|          | Quantity          | Phyiscal Quantites                                   |                                    |                                 |    |  |  |
| Name /   | Frequenz          | Beschleunigungs- amplitude Betrag Phasenverschiebung |                                    | Index<br>position               |    |  |  |
| rtarrio  | f in Hz           | â in m/s²                                            | $S_{qa}$ in pC/(m/s <sup>2</sup> ) | $\Delta arphi_{	ext{qa}}$ in 1° |    |  |  |
| Label /  | 10                | 5                                                    | 0,1301                             | -0,02                           | 0  |  |  |
| Unit     | 12,5              | 5                                                    | 0,1302                             | -0,03                           | 1  |  |  |
|          | 16                | 5                                                    | 0,1302                             | -0,02                           | 2  |  |  |
| Values — | 20                | 10                                                   | 0,1302                             | -0,02                           | 3  |  |  |
|          | 25                | 10                                                   | 0,1302                             | -0,01                           | 4  |  |  |
|          | 31,5              | 10                                                   | 0,1302                             | -0,01                           | 5  |  |  |
|          | 40                | 10                                                   | 0,1302                             | -0,02                           | 6  |  |  |
|          | 50                | 50                                                   | 0,1302                             | -0,01                           | 7  |  |  |
|          | 63                | 50                                                   | 0,1301                             | 0,00                            | 8  |  |  |
|          | 80                | 50                                                   | 0,1301                             | 0,01                            | 9  |  |  |
|          | 100               | 50                                                   | 0,1301                             | -0,01                           | 10 |  |  |
|          | 125               | 50                                                   | 0,1301                             | -0,01                           | 11 |  |  |
|          | 160               | 50                                                   | 0,1301                             | 0,00                            | 12 |  |  |

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin


Nationales Metrologieinstitut

26.02.202



# **Dcc:quantites**






- Dcc:quantities are structures containing
  - Name (multi language)
  - Description (multi language)
  - · Unit as D-SI string
  - Values (scalar) or lists
  - Uncertanties (scalar) or lists
  - Rich metadata as XML elements
    - · Nested quantities are allowed here
- Software should use **dcc:quantities** as the smallest unit of information.

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitu

# dccQuantities library





Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin


Nationales Metrologieinstitut

02.2025



# dccQuantities library





Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

# dccQuantities library



Read DCC and parse quantities




• Standard Python functions



 Arithmetic operations with uncertainties









Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



## dccTables-Library?



#### A similar library for tables?

- Most data is tabulated
- We need to be able to work with this data digitally!
  - We want to parse the data into a format that is easy to understand for humans, e.g. as plots
- **Problem**: There are a large number of ways to represent tabulated data in the DCC...



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin 26.02.2025 Nationales Metrologieinstitut

**Table formats** 



#### Why is that problematic?

- First, we would need to find every table format that is used anywhere
- Some formats increase the size of file so much that it can not be easily parsed
- Every format needs its own parser



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



#### **Table formats**



- We need a standardized table format in order to be able to productively use data from the DCC!
- Keep in mind: DCC is not intended to be directly read by humans!



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

26.02.2025

. .


Nationales Metrologieinstitut

# **Table format requirements**



#### What are tables?

- Rows of data taken under the same conditions
- Columns of the same kind of data, taken under different conditions
- Entries can be found via the index columns



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



## **Table format requirements**



- Indexing
  - One, two, or *n* dimensions
    - Most cases only have one dimension
  - Numeric or categorical indexes
    - Can be mixed in multi-dimensional tables
- Some metadata (like units) is the same for each element, include option to avoid repetition

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin


Nationales Metrologieinstitu

26.02.2025

13

# Q P

# **Proposed Table Format: 1D**



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



## **Proposed Table Format: 1D**




Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

6.02.2025

# **Proposed Table Format: categoric index**





Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

16



# **Proposed Table Format**



# More than one index vector

- In most cases, it's sufficient to repeat index vector values ("long format")
- Then, format can follow 1D case

| Pos | Index 1 | Index 2 | Index 3 | Result        |
|-----|---------|---------|---------|---------------|
| 0   | 1.3     | а       | 1       | 2.8 ± 0.2     |
| 1   | 1.3     | а       | 2       | 8.2 ± 0.9     |
| 2   | 1.3     | b       | 1       | 9.0 ± 0.6     |
| 3   | 1.3     | b       | 2       | 1.2 ± 0.8     |
| 4   | 3.7     | а       | 2       | $7.7 \pm 0.6$ |
| 5   | 3.7     | b       | 1       | 6.5 ± 0.2     |
| 6   | 3.7     | b       | 2       | 9.2 ± 0.4     |

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

1

Nationales Metrologieinstitu

# **Proposed Table Format: > 1D**



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

25



# Proposed Table Format: > 1D



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

26 02 2025

19

Nationales Metrologieinstitut

## **Proposed Table Format**



- Avoiding index value repetition is only necessary for very large tables (≥ 1000 entries, ≥ 2 dimensions)
- We are working on a proposition for such a format ("flat format")
  - Please contact us if you have use cases, so we can take them into account



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



## **Proposed Table Format: Benefit**



- Proposed format covers all use cases we know of
- · Writing a parser for a strict format is very easy
- · Enables better software, and thus greatly increases usability



Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin



### **Discussion**



Images: Illustrations CC0, Screenshots own work

#### Contact

Benedikt Seeger benedikt.seeger@ptb.de gitlab1.ptb.de/Seeger

Vanessa Stehr vanessa.stehr@ptb.de gitlab1.ptb.de/stehr01

dccQuantities-library: gitlab1.ptb.de/digitaldynamicmeasurement/dccQuantities/-/tree/devel-class-restructuring

Physikalisch-Technische Bundesanstalt ■ Braunschweig und Berlin

Nationales Metrologieinstitut

Back to "Table of Contents" above

#### A framework for digitising dimensional metrology 015

Presenting author: Arif Sanjid Email: sanjid@nplindia.org

Additional authors: Naveen Garg, R Srinivas Rao CSIR- National Physical Laboratory, New Delhi, India

### **Abstract**

The ISO 17450 (2011) entitled Geometrical product specifications (GPS)—General concepts: Model for geometrical specification and verification recommends framework of



dealing developing models for dimensional metrology applications. Moreover, its part 2, describes operators of the coordinate measurement data. Likewise, there are assorted document standards, literature and standard operating procedures to perform dimensional metrology calibrations. The KCDB (the BIPM key comparison database) web portal enlisted the calibration services of length along with their key comparison data. Accordingly, Calibration and measurement capabilities the national measurement institutes (NMIs) across the globe are endorsed after an approved peer reviewed audit.

In the era of industry 5.0, a framework that transforms metrological activities into digital format is inevitable. National Physical Laboratory, India has identified the significance of digitalising the dimensional metrology and initiated the efforts to device digital calibration certificates (DCC) in collaboration with the centre for calibration and testing (CFCT) of NPL, India. These attempts may lead to an evolution of the generic framework to meet the requirement of (Findable, Accessible, Interoperable, and Reusable data and services) FAIR.

This work aims to discuss a framework to digitalise the dimensional metrology practices and application. Present, the proposed abstract model is explained with a dimensional measurement application. The proposed model involves a process flow which emphasizes (1) metrological traceability of the reference measures and materials (2) Geometric product specification of the device under test (3) static and dynamic characteristics of the metrological loop of the measuring system (4) measurement process and raw measurement data. In dimensional metrology, the metrological loop comprises of the influences due to the structural rigidity, geometric imperfection – orientations of the workpiece, and the kinematics - the kinetic interactions. The abstract model is used to compose the measurement data in a format to supplement the requirement of FAIR. The procedure of measurement data is suggested in ISO 17450-2. The digitalization of the dimensional metrology also includes development of digital twins by implementing simulations using different artificial intelligence algorithms and techniques.

Keywords: DCC, FAIR, Artificial intelligence

Back to "Table of Contents" above



# A framework for digitising dimensional metrology

M Arif Sanjid\*, Naveen Garg, Rrinivas Rao CSIR- National Physical Laboratory, New Delhi, India



# Outline

- Introduction
- Dimensional Metrology
- Specific Application
- Future prospects
- Conclusion




# National Physical Laboratory

- History
  - The National Physical Laboratory is one of the earliest national laboratories set up under the Council of Scientific & Industrial Research(CSIR).
  - Late Shri Jawaharlal Nehru laid the foundation stone of NPL on the 4th January 1947.
  - Late Dr. K. S. Krishnan, FRS, was the first Director of the laboratory.
  - The main building of the laboratory was formally opened by Late Deputy Prime Minister. Sardar Vallabhbhai Patel. on the 21st January 1950.



# CSIR - NPL Organization Chart



# NPL Organization for Quality System




Chart # 2. Organization Chart for Quality System



# Dimension Metrology at NPLI





Calibration Certificate: 250/Annual

No. of CMCs - 46 Team members - 6 Revenue: \$60K /Annual







# Dimension Metrology at NPLI











# Digitalization for FAIR

## Requirements for FAIR

- Traceability
- Measurement
- Expression of results
- Digital Twin
  - Machine kinematics
  - GPS of reference artefacts
  - GPS of measuring instruments
  - Simulation using AI algorithms

# **Useful Document Standard**

- ISO 17450: Model for geometrical specification and verification
  - surface model skin model features operators -
- ▶ ISO 1101:2017 : Geometrical product specifications (GPS)
- ISO 14406:2010 GPS Extraction
- ISO 16610-21:2025 (GPS) Filtration Part 21: Linear profile filters: Gaussian filters
- Y14.5 Dimensioning and Tolerancing(R2024)
- Y 14.5.1 M- MATHEMATICAL DEFINITION OF DIMENSIONING AND TOLERANCING PRINCIPLES




# Measurement

- Traceability to SI units
  - Reference Artifact -Length Bar, Hemisphere, Surface specimen, etc
  - Reference instruments- laser wavelength, autocollimator
- Workpiece GPS
- Verification operation
- Measurement techniques
  - Taylor envelope principle
  - Reversal Technique
  - Closure method
  - Multistep error separation method
- Results
  - Measurement Model- Uncertainty estimation
  - **Conformity** assessment


# Specific Application

- DiameterMeasurement
- Digitalization
- the verification of the specification





# Diameter Measurement Application



- Substitution method The circumference of the piston is scanned to detect the peaks (i.e., reversal points) of its curvature.
  - The zenith point searching and the misalignment.

# Operators for diameter Measurement (ISO-17450)

#### Partition-

the ability to locate the centre of a rectangular GB in mechanical comparison calibration is limited to 0.5 mm. Then the uncertainty contribution due to imperfect form is 2 nm

The scanned circumference constitutes the ROI for the given specimen.

$$\Delta\theta_{max} = 2cos^{-1}[1 - 2h_{max}/D]$$

The optimal angle between ROIs that minimizes the uncertainty of the measured quantity—is 120° for the radius and the Euclidean centre position



# Operators for diameter Measurement (ISO-17450)

#### **Extraction-**

The operator decides the reversal point based on the least change (i.e. by one least count) in the reading of probe.

- Instead, the coordinates of the circumference of the specimen under test are sampled.
- The peak of this profile curve is determined by employing the gradient method ( Different Al algorithm can be applied-e.g. Gradient search )
- Such an interpolation improves the uncertainty as we can convincingly assign a triangular distribution to the detection of peak.

# Operators for diameter Measurement (ISO-17450)

#### Filtration-

F meli et al scanned three spheres with a resolution  $10^{\circ}(5^{\circ})$  in longitude (latitude), respectively to map their radius. This method measures the variations of radius as a function of angle

#### Collection-

The parameters associated with the circular cross-section of the specimen are roundness and diameter. Other additional parameters such as straightness, and parallelism, are required for cylindrical surfaces.



# Operators for diameter Measurement (ISO-17450)

#### Association-

- ▶ The surface model : The matrix of the definition of geometric features and their interrelationships are illustrated in ISO 14660-1
- The diameter measurement is associated with (i) at least three measurements of roundness along the length of the cylindrical specimen, (ii) measurement of straightness and parallelism of the cylindrical surface parallel to its axis [Calibration Guide, EURAMET/cg-06/v.01, 2007]



#### Construction-

- The profiles of the diameter and faces of the reference GB is extracted using two LVDT displacement sensors.
- Asymmetric datum is constructed from the coordinates of these profiles. The displacement of translation is read perpendicular to the constructed datum.
- The laser beam of displacement measuring interferometer is set coaxial to the constructed datum line [Meas. Sci. Technol. 8 (1997) 849–856.]. Consequently, the kinematics of the experiment is aligned referring to this datum.



# Operators for diameter Measurement (ISO-17450)

#### **Reconstruction** –

- Usually, the coordinate data is registered by considering their root mean square of residue.
- The consolidated discrepancy has been accounted for the uncertainty estimation.

#### Reduction -

- A spindle axis serves as a datum for the roundness measurements.
- On the contrary, the diameter does not need a datum.
- However, the fusion-based diameter measurement needs a center to construct a reference circle

# Digital Twins

- GPS for automative, manufacturing
- Digital twins for
  - Management of measurement process
  - Optimization of conformity
  - Different applications
- Metrological Loop
  - Datum
  - Kinematics
    - Orientation Abbe
    - Situational Bryan
  - Kinetics
    - Contact force
- Simulation Not unique



# Measuring Equipment

<u>ISO 14253-2:2011</u>, (GPS) — Inspection by measurement of workpieces and measuring equipment —

- Decision rules for verifying conformity or nonconformity with specifications
- Guidance for the estimation of uncertainty in GPS measurement, in calibration of measuring equipment and in product verification

<u>ISO 10012:2003-</u> Measurement management systems — Requirements for measurement processes and measuring equipment

- Statics ISO 3534, ISO 2602, , ISO 2854, ISO 5479, ISO 5725, ISO 7870
- ► ISO 10017:2021Quality management Guidance on statistical techniques for ISO 9001:2015

# Conclusion

- A uniform Matrix for definitions
  - VIM- Based
    - Parameters as per IS 1101
      - Deviation in Diameter
      - Cylindricity Straightness , Roundness
    - Uncertainty of measurement model equation
      - Matrix of coefficients and input data
- Results in the digital certificates
  - Metrological Traceability
  - Measured value , Associated uncertainty
  - Conformity statements -
    - · Qualifiers CMC based, tolerance based



# THANK YOU

# HAVE A GOOD DAY

National Physical Laboratory Dr.K.S Krishnan Marg New Delhi 1100012,INDIA www.nplindia.org



Back to "Table of Contents" above

Day2 (2025-02-26 Wednesday)

Parallel Session "Industry Applications"

(Chair: Carlos Galvan)

O16 Instructions for using the DCC scheme to create a digital calibration certificate for gauge blocks

Presenting author: Christian Sander; Testo Calibration Services GmbH Email:

Additional authors: Katharina Janzen<sup>1</sup>, Guido Bartl<sup>1</sup>, Daniel Heißelmann<sup>1</sup>, Ralf Liske<sup>2</sup>

<sup>1</sup> Physikalisch-Technische Bundesanstalt (PTB)

<sup>2</sup> L & W GmbH

### **Abstract**

#### Introduction

In the course of digitization in metrology, the Physikalisch-Technische Bundesanstalt (PTB) developed the digital calibration certificate (DCC), which enables the reporting of calibration results for all measurands in a machine-interpretable format and meets the requirements of DIN EN ISO/IEC 17025:2018 for the reporting of results. For this purpose, an XML schema definition (XSD) was created, which specifies the data types and structure of an XML document. The developed XSD schema offers flexible and diver se solutions for presenting



the results for each individual measured variable, provided they are reported in a schema-compliant XML document. Within each metrology community, the uniform use of terminology for the communication of measurement results is also essential, which is why the universally defi ned XML elements must be supplemented with human- and machine-interpretable information. This additional information must be individually coordinated in each community to ensure the interpretability of the data exchange. For this reason, a good practice example (GP) for a DCC for gauge blocks was developed within the framework of the DKD Technical Committee on Length in cooperation with PTB, which is presented in this expert report.

The developed and presented DCC is based on schema version 3.3.0. As the schema definition is subject to continuous further development, this presentation represents the status of the discussion at the time of publication.

#### Calibration object

Gauge blocks (e.g. according to EN ISO 3650:1998) made of steel, ceramic or carbide are cuboid blocks with two parallel functional surfaces with very high surface quality and accuracy (distance, parallelism, flatness). Gauge blocks can be assembled to different lengths, which is referred to as 'wringing'. In practice, gauge blocks are used as individual items or in the form of ready-made sets of individual gauge blocks with specific increments of the nominal size (e.g. 125-piece set). A set can contain several gauge blocks of the same nominal size and different materials.

#### Objective

The use of the machine-readable DCC eliminates the error-prone manual transfer of the determined calibration results, such as the centre dimension or the deviation of the centre dimension from the nominal dimension, on the user side when using gauge block sets for metrological purposes. To avoid interpretation errors between the generating and processing systems of the DCC, a uniform representation of the usable data in DCC format is desirable. For the example of gauge block sets, a solution is presented in this report as "good practice".

Back to "Table of Contents" above



### Presentation of Christian Sander





# A good practice example applying the DCC schema in gauge block calibrations

Katharina Janzen <sup>1</sup>, Guido Bartl <sup>1</sup>, Daniel Heißelmann <sup>1</sup>, Ralf Liske <sup>2</sup>, <u>Christian Sander</u> <sup>3</sup>

- <sup>1</sup> Physikalisch-Technische Bundesanstalt (PTB)
- <sup>2</sup> L & W GmbH



#### Introduction



DKD – German Calibration Service – a forum of PTB

- More than 400 calibration laboratories
- Goal: promote uniformity in metrology
- 13 technical committees
  - Implementation of good practice examples of DCCs
  - Standardized DCCs for various device types
  - First DCC of TC Length → calibration of gauge blocks (https://doi.org/10.7795/550.20250131)





2024-04-16

<sup>&</sup>lt;sup>3</sup> Testo Calibration Services GmbH



#### Introduction



Gauge blocks (ISO 3650:1998)

- cuboid blocks with two parallel functional surfaces to transfer the length quantity
- Used as reference standards for dimensional calibrations in industry
- Can be joined/wrung with low uncertainties
- Sets of blocks with various nominal lengths
  - Made of steel, ceramics or carbides
  - Available in different grades/accuracy classes
  - Identification of individual block by marking



DCC ≡Ø


2024-04-16


# Introduction



#### Gauge block calibration

- Functional tests
  - Visual inspection
  - Wringability
- Dimensional calibration
  - a) central length I<sub>c</sub> or
  - b) deviation of central length I<sub>c</sub> from nominal length I<sub>n</sub>
  - Optional: deviation span f<sub>o</sub> and f<sub>u</sub>
- Method: VDI/VDE/DGQ 2618 3.1:2004-01








2024-04-16






#### Structure of DCC with focus elements



# DCC GP gauge blocks




#### dcc:items structure







#### dcc:items structure





DCC schema 3.3.0 enables a structred representation of set of calibration objects




# DCC GP gauge blocks



dcc:identifications of the set of gauge blocks

Information for identifying the complete set of gauge blocks










dcc:itemQuantities of the set of gauge blocks

Properties valid to the complete set of gauge blocks




# DCC GP gauge blocks



dcc:itemQuantities of the set of gauge blocks

Properties valid to the complete set of gauge blocks








dcc:itemQuantities of the set of gauge blocks

Properties valid to the complete set of gauge blocks




# DCC GP gauge blocks



dcc:item (individual gauge block in the set)

Properties valid to idividual gauge blocks








dcc:item (individual gauge block in the set)

Properties valid to idividual gauge blocks




# DCC GP gauge blocks



dcc:item (individual gauge block in the set)


Properties valid to idividual gauge blocks







#### Structure of DCC



# DCC GP gauge blocks



#### dcc:usedMethod

Description of the test or calibration method





#### dcc:measuringEquipments

Applied reference standards (optical flat)

```
<dcc:measuringEquipment>
 <dcc:name>
 <dcc:certificate>
 <dcc:referral>
 <dcc:referralID>{CertificateNumber}</dcc:referralID>
 <dcc:procedure>analogue</dcc:procedure>
 <dcc:value>noValue</dcc:value>
 </dcc:certificate>
 <dcc:identifications>
 <dcc:identification refType="basic_identificationNumber">
 </dcc:identifications>
 </dcc:measuringEquipment>
 dcc:measurementResult
 dcc:measuringEquipments
 refType ="length_wringability"
 dcc:result refld="setOfItems"
 dcc:results
2024-04-10
```



# DCC GP gauge blocks




#### dcc:results (for functional test - wringability)

```
<dcc:result refId="setOfItems">
 <dcc:name>
 <dcc:data>
 <dcc:quantity>
 <dcc:name>
 <si:real>
 <si:value>1</si:value>
<si:unit>\one</si:unit>
 </si:real>
<dcc:measurementMetaData>
 <dcc:metaData>
<dcc:declaration>
 Boolean quantity types
 </dcc:metaData>
 </dcc:measurementMetaData>
</dcc:quantity>
 urgently needed
 </dcc:data>
</dcc:result>
 dcc:measurementResult
 dcc:measuringEquipments
 refType ="length_wringability"
 dcc:result refld="setOfItems"
 dcc:results
 2024-04-10
```





#### Structure of DCC




# DCC GP gauge blocks



#### dcc:results

Dimensional calibration results









#### dcc:results

```
<dcc:name
 <si:value>0.06000014</si:value>
 <si:unit>\metre</si:unit>

 Dimensional calibration

 <si:measurementUncertaintyUnivariate>
 <si:expandedMU>
 results
 <si:valueExpandedMU>0.136e-06</si:valueExpandedMU>
 <si:coverageFactor>2</si:coverageFactor>
<si:coverageProbability>0.95</si:coverageProbability>
 </si:expandedMU>
 </si:measurementUncertaintyUnivariate>
 </si:real>
 </dcc:quantity>
 dcc:quantity refType="length_centralLength"
 dcc: quantity\ refType="length_deviationFromNominalLength"
 dcc:quantity refType="length_upperDeviationFo"
 dcc:result refld="Item_1"
 dcc:quantity refType="length_lowerDeviationFu"
 dcc:quantity refType="lenght_accuracyClass" o dcc:metaData refType="basic_conformity"
 dcc:results
 dcc:result refld="Item_2"
```

<dcc:quantity refType="length\_centralLength">

dcc:quantity refType="length\_accuracyClass" • dcc:metaData refType="basic\_conformity"

# DCC GP gauge blocks

dcc:result refld="setOfItems"



#### dcc:results

 Dimensional calibration results

```
<dcc:quantity refType="length_deviationFromNominalLength">
 <dcc:name
 <si:value>0.14e-06</si:value>
<si:unit>\metre</si:unit>
 <si:measurementUncertaintyUnivariate>
 <si:expandedMU>
 <si:valueExpandedMU>0.136e-06</si:valueExpandedMU>
 <si:coverageFactor>2</si:coverageFactor>
 <si:coverageProbability>0.95</si:coverageProbability>
</si:expandedMU>
 </si:measurementUncertaintyUnivariate>
 </si:real>
</dcc:quantity>
```

```
dcc: quantity\ refType="length_deviationFromNominalLength"
 dcc:quantity refType="length_upperDeviationFo"
 dcc:result refld="Item_1"
 dcc:quantity refType="length_lowerDeviationFu"
 dcc:quantity refType="lenght_accuracyClass" • dcc:metaData refType="basic_conformity"
dcc:results
 dcc:result refld="Item_2"
 dcc:quantity refType="length_accuracyClass" • dcc:metaData refType="basic_conformity"
 dcc:result refld="setOfItems"
```

dcc:quantity refType="length\_centralLength"





## DCC GP gauge blocks



#### dcc:results

dcc:results

 Dimensional calibration results

```
<dcc:quantity refType="length_upperDeviationFo">
 <dcc:name
 <si:real>
 <si:value>0.07e-06</si:value>
 <si:unit>\metre</si:unit>
 <si:measurementUncertaintyUnivariate>
 <si:expandedMU>
 <si:valueExpandedMU>0.07e-06</si:valueExpandedMU>
 <si:coverageFactor>2</si:coverageFactor>
 <si:coverageProbability>0.95</si:coverageProbability>
 </si:expandedMU>
 </si:measurementUncertaintyUnivariate>
 </si:real>
 </dcc:quantity>
 dcc:quantity refType="length_centralLength"
 dcc: quantity\ refType="length_deviationFromNominalLength"
 dcc:quantity refType="length_upperDeviationFo"
dcc:result refld="Item_1"
 dcc:quantity refType="length_lowerDeviationFu"
 dcc:quantity refType="lenght_accuracyClass" • dcc:metaData refType="basic_conformity"
 dcc:result refld="ltem_2"
```

dcc:quantity refType="length\_accuracyClass" • dcc:metaData refType="basic\_conformity"

<dcc:quantity refType="length accuracyClass">



## DCC GP gauge blocks

dcc:result refld="setOfItems"



```
<dcc:name>
dcc:results
 <dcc:content lang="de">Genauigkeitsklasse</dcc:content>
 <dcc:content lang="en">Accuracy class</dcc:content>

 Dimensional calibration

 <dcc:charsXMLList>1</dcc:charsXMLList>
 results
 <dcc:measurementMetaData>
 <dcc:metaData refType="basic_conformity">
 <dcc:name>
 <dcc:content lang="de">Konformität</dcc:content>
<dcc:content lang="en">Conformity</dcc:content>
 </dcc:name>
 <dcc:norm>DIN EN ISO 3650:1999</dcc:norm>
 <dcc:conformity>pass</dcc:conformity>
 dcc:quantit
 </dcc:metaData>
 dcc:quantit
 </dcc:measurementMetaData>
 </dcc:guantity>
 dcc:quantit
 dcc:result refld="Item_1"
 dcc:quantity refType="length_lowerDeviationFu"
 dcc:quantity refType="lenght_accuracyClass" • dcc:metaData refType="basic_conformity"
 dcc:results
 dcc:result refld="ltem_2"
 dcc:quantity refType="length_accuracyClass" • dcc:metaData refType="basic_conformity"
 dcc:result refld="setOfItems"
```



## DCC GP gauge blocks



```
<dcc:result refId="setOfItems">
dcc:results
 <dcc:name>
 <dcc:data>

 Dimensional calibra

 <dcc:quantity refType="length_accuracyClass">
 <dcc:name>
 results
 <dcc:charsXMLList>1</dcc:charsXMLList>
 <dcc:measurementMetaData>
 <dcc:metaData refType="basic_conformity">
 <dcc:name>
 <dcc:content lang="de">Konformität</dcc:content>
 <dcc:content lang="en">Conformity</dcc:content>
 </dcc:name>
 <dcc:norm>DIN EN ISO 3650:1999</dcc:norm>
 <dcc:conformity>pass</dcc:conformity>
 </dcc:metaData>
 </dcc:measurementMetaData>
 dcc:result refld="Item 1"
 </dcc:quantity>
 </dcc:data>
 /dcc:result>
 dcc:results
 dcc:result refld="Item_2"
 dcc:quantity refType="length_accuracyClass" o dcc:metaData refType="basic_conformity"
 dcc:result refld="setOfItems"
```

## DCC GP gauge blocks



#### Summary

- Machine interpretable DCCs for gauge blocks can be implemented using schema 3.3.0 with specific refTypes and ids/reflds
- Sets of objects are well represented by dcc:item and dcc:subitems
- General structure is assumed to be easily adapted to other dimensional artefacts
- The schema 3.3.0 should be extended to boolean quantity types



2024-04-16 26







<u>ww.dkd.eu</u>

Back to "Table of Contents" above

# O17 Automation in calibration process using DCC, primary to secondary labs.

Presenting author: Aldo Adrián García González

Email:

Additional authors:

#### **Abstract**

This work shows how DCC can be implemented at acceleration primary laboratory to calibrate accelerometers that can be implemented directly to acceleration secondary labs. The solution uses excel sheets, python and LabView to generate results subscemes. The integration of all good practices allows to reduce calibration process.

The work also will show.


- o Implemented method.
- Developed software.
- o Results and future work.



#### Back to "Table of Contents" above

## Presentation of Aldo Adrián García González









# ® Economía Steinam Digital Quality Infrastruuture







3 MINTERNATIONAL DCC CORFERENCE




Page 256 of 487





















5 M NOTES NATIONAL DCC CONFERENCE

## ©CC Generator with Sub-Schema from Python






5m international RCD Corference









3 MINTERNATIONAL DCC CONFERENCE






Smintebnational RCC Conference




## ©CC Generator with Sub-Schema from Python






#### 5 IN INTERNATIONAL DCC CORFERENCE































SIN INTERNATIONAL DCC CORFERENCE



# ® Economía Steenam DCC at Transfer laboratory





#### 5 MINTERNATIONAL DCC CORFERENCE

NOT SOME WHICH INVESTIGATION AND THE CONTRACT INTERCORNATIONS AND VIOLE





Different Ways to Generate and Read DCC was created and included at the calibration process.

EXCEL EYTHON LABVIEW



Collaborate with other laboratories to perform an intercomparison using DCC



Implementation of Subschema's



Benefits to reduce error and time in calibration



Method to transfer the knowledge to implement DCC at laboratories.



Transfer the knowledge to other NMI's and users



Integrate a complex sensor network to measure different quantities inside a laboratory



Integrate AI to enable new aplitacions using Machine Learning





Back to "Table of Contents" above

# O18 Semantic registration of DCC and DCoC onto Information Model of International Data Spaces

Presenting author: Tomasz Sołtysiński; Physikalisch-Technische Bundesanstalt (PTB), Abbestraße 2-12, D-10587 Berlin-Charlottenburg, Germany Email:

Additional authors:

## **Abstract**

Many processes in the quality infrastructure require the provision of a document of compliance or conformity with legal, contractual or regulated requirements. For digital quality infrastructure, these proofs need to be provided using digital communication protocols and connected data sources. Data spaces allow to comprise an ecosystem connecting different participants offering their assets, like resources, services and exchanging protocols with metadata and data. Focused on such applications, we discuss and demonstrate different scenarios allowing us to exchange specific protocols, like a digital certificate of conformity (D-CoC) or digital calibration certificate (DCC), as used for instance, in legal metrology.

To enable a semantic-driven flow of content carried by D-CoC and DCC within an ecosystem adhering to the principles of digital quality infrastructure, such as Quality-X, based on International Data Spaces (IDS), a unified and interoperable framework is essential. In this context, a new approach is proposed: registering the vocabulary embedded in certificates onto a semantic information model of the de facto standardized IDS. This ensures secure, traceable, and harmonized metadata exchange. Such a framework lays the foundation for efficient implementation of digital processes, including quality assurance through interlaboratory comparisons and round trials, as well as the validation of AI methods and software.

Combining our approach with specific standardized interfaces being operated by industry, like asset administration shell (AAS), the unified data communication protocol OPC UA and



the industrial digital twin association (IDTA), we discuss solutions paving the way towards effective, interoperable, automated, machine -readable pipelines shaping prospective digital quality infrastructure framework.

Back to "Table of Contents" above

## Presentation of Tomasz Sołtysiński



## **Agenda**



- 1. Information Model vs DCC
- 2. Information Model vs D-CoC
- 3. Semantics for DCC (temperature, pressure or mass)
- 4. Registration of DCC onto Informaton Model of IDS
- 5. Semantics for D-CoC (process monitoring)
- Federated digital ecosystems to ensure trustworthy performance of processes by circulating the certificates
- 7. Integration with OPC UA & AAS
- 8. Summary and conclusions



5th INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE



## **Process** Information model vs DCC

#### Semantics from DCC:

DCC for temperature calibration DCC for pressure calibration DCC for mass calibration



## **Process** Information model vs D-CoC

Umgebungsbedingungen und Einflussgrößen:
Erwisomental conditions and influence factors:

Temperaturbereich
Feuchtigkeit
Humidity

Mechanisch
Mechanisch
Mechanical
Elektromagnetisch
Elektromagnetisch
Elektromagnetisch
Elektromagnetisch
Elektromagnetisch
Elektromagnetisch
Elektromagnetisch
Ez (gemäte GIML D11)
EZ (according to OIML D11)

| Genauigkeitsklasse:             | 2,0                                                                       |
|---------------------------------|---------------------------------------------------------------------------|
| Durchflussbereich:              | Praxisgerecht innerhalb des Durchflussbe<br>reichs des Messwertaufnehmers |
| Kleinste Messmenge:             | 1 kg                                                                      |
| Maximaler Gasdruck:             | 875 bar                                                                   |
| Minimaler Gasdruck:             | 50                                                                        |
| Umgebungsbedingungen/Einflussgr | ößen:                                                                     |
| klimatisch:                     | -25°C bis 55°C (betauend)                                                 |
| mechanisch:                     | M2                                                                        |
| elektromagnet                   | tisch: E1                                                                 |


| Débit permanent Q3 (m³/h) Q3 Permanent flowrate (m³/h)      | 1,6 | 2,5   |
|-------------------------------------------------------------|-----|-------|
| Débit de surcharge Q4 (m³/h)<br>Q4 Overload flowrate (m³/h) | 2,0 | 3,125 |

Sheveleva, Tatyana & Foyer, Gisa & Knopf, Dorothea d-coc@ptb.de










## **Semantics for DCC** | Pressure, temperature, mass



**₽**PIB

5TH INTERNATIONAL DCC CONFERENCE



## **Semantics for DCC** | Registration of DCC onto Informaton Model of IDS

Registration of semantics into/onto domain-agnostic information model of IDS:

classes, objects, values

#### Information Model:

RDF Schema (RDFS)/OWL ontology, which defines fundamental concepts for describing actors, their interactions, the resources exchanged by them, trust and data usage restrictions.

This model introduces semantics into the IDS and works in conjunction with the FAIR principles.


RDF/turtle, RDF/XML, JSON\_LD, N-Triples

- DCC and D-CoC as metadata protocol (no connectors needed for a process driven with credentials)
- One information model = one process cycle (full process flow)
- DCC and D-CoC as a specific protocol for exchange of metadata along the process
- Harmonized (meta)data model to be adopted to DX, DCR, DCA




5th INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTUR

## **Semantics for DCC** | Registration of DCC onto Information Model of IDS






## Semantics for DCC | Registration of DCC onto Informaton Model of IDS




#### Semantics for D-CoC | Semantics-driven exchange and disclosure of certificates






5th INTERNATIONAL DCC CONFERENCE















## Federated digital ecosystems | Performance of processes by circulating the certificates





5tH INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTUR

## OPC UA & AAS | Integration of DCC and D-CoC with information model of OPC UA






5TH INTERNATIONAL DCC CONFERENCE




## OPC UA & AAS | Integration of DCC and D-CoC with information model of OPC UA





5<sub>TH</sub> INTERNATIONAL DCC CONFERENCE

## OPC UA & AAS Integration of DCC and D-CoC with information model of OPC UA





5TH INTERNATIONAL DCC CONFERENCE



### **DCC in trusted QI** | Summary and discussions

Constructed scenario:


- demonstrates how selective disclosure of confidential data contained in a DCC, DCoC, or DPP can be automatically and interoperable accessed using standardized information model of IDS
- such a protocol can be used, for example, to extract and publish information on **metrological traceability** of specific measurements, including their uncertainties performed across a chain of calibrations
- the protocol facilitates the **exchange of metadata, data, and certificates** related to, for instance, the calibration of weighting scales
- is directly applicable across industrial interfaces like OPC UA and AAS
- provides a framework for communication between **stakeholders** (such as **legal or accrediting authority**, **companies**, **warehouses**, **and calibration laboratories**) across different participants of data space
- can be successfully achieved if the proposed protocols operate within the European Metrology Cloud and its ecosystem, adhering to the rules of international data spaces.



5<sub>TH</sub> INTERNATIONAL DCC CONFERENCE







Back to "Table of Contents" above

#### 019 DCC in a Mass Calibration Software

Presenting author: Lisa Busser, Keysight Technologies Deutschland GmbH

Email: lisa.busser@maro.de

Additional authors:

## **Abstract**

ScalesNet is an already existing and well-known Software in mass metrology. As such a stakeholder the developing team of ScalesNet has been asked to be the industrial partner of PTB with the goal of embedding the DCC in an already existing software.

This presentation includes a short demonstration of the already existing possibilities to generate a DCC automatically with results of a real measurement that validates against the schema of PTP. On top of that, the current state of development regarding the import of the data out of the DCC and into the software will be shown. Finally, two suggestions for changing the schema will be made. These proposals aim to facilitate repeatability of development for other software developers and to improve the interoperability of DCCs from different software vendors. In addition, it is particularly important to be able to make the import of data from the DCC back into software as automated as possible. The general data fields into which continuous text is inserted are particularly problematic here. Because different spellings of a company name or address are not automatically recognized as the same by software. This lecture promotes a suggestion as to how these text fields can be handled in the import to make importing possible with as little intervention as possible, without having to resort to a globally central database or having many duplicates of the same entry in your own database.



#### Back to "Table of Contents" above

## Presentation of Lisa Busser



# DCC IN A MASS-CALLIBRATION SOFTWARE

Lisa Busser, MARO Elektronik




## Disclaimer

- Close partnership with PTB
- DCC checked by PTB










## **XML in PDF**

Foxit PDF Reader

Collection form with any number of attachments possible







## **XML in PDF**


Visual Studio Code

```
| Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Section | Sect
```



## XML in PDF

Visual Studio Code







## **XML** in PDF – Digital Signature

- Accreditation Symbol is applicable
- Sign PDF digitally Sis easy
- XML is an attachment to PDF and is therefore automatically signed when the PDF is signed




## XML in PDF – Digital Signature

- Accreditation Symbol is applicable
- Sign PDF digitally is easy
- XML is an attachment to PDF and is therefore automatically signed when the PDF is signed


The total size of the file is significantly smaller!



## UUID (<u>U</u>NIVERSALLY <u>U</u>NIQUE <u>ID</u>ENTIFIER)



## UUID (UNIVERSALLY UNIQUE IDENTIFIER)







## UUID (Universally Unique Identifier)

- Is the UUID really unique?
  - 128 Bit "number" = 2<sup>128</sup> possibilities
- Crossreference-Table
   (to handle multiple UUIDs of the same object)
- Risc of duplications negligible (MAC-Adress, system time)
- Different ways to generate UUIDs



## Contact



Back to "Table of Contents" above

## O20 Progress Towards Producing DCC with Standard Structured Measurand Information

Presenting author: Jens Leinhos

Email:

Additional authors: Mark Kuster, Colin Walker, Mike Schwartz, and Jonathan Harben

## **Abstract**

Advances in machine-actionable information and data for the quality infrastructure require further exploration. This includes the implementation of Digital Calibration Certificates (DCC)



by producing real-world measuring instrument examples to investigate potential use cases with consumers. The integration of units and quantity information from three primary sources — ISO 80000 quantities and units, DCC Reference Types, and NCSLI measurand taxonomy will be examined. The study will review the progress made so far and identify any gaps, outlining goals for future work. This investigation aims to provide a foundation for further research and development in the field.

Back to "Table of Contents" above Back to "Table of Contents" above

Day 2 (2025-02-26 Wednesday)
Parallel Session "Applications NMI II"
(Chair: Jariya Buajarern)

# Digital Calibration Certificates for Radioactivity and FAIRData Exchange

Presenting author: Henrike Fleischhack Email: henrike.fleischhack@ptb.de

Additional authors: Stefan Röttger stefan.roettger@ptb.de, Matthias Oliver Stein

matthias.stein@ptb.de (PTB)

## **Abstract**

The future is digital: Our increasingly interconnected world requires fast, accurate exchange of data and information. DCCs (digital, machine-interpretable calibration certificates) will play a large role in digitalizing our quality infrastructure, allowing for automated, streamlined communication between calibration laboratories, research institutes, and manufacturing facilities. Simultaneously, the scientific community needs machine-interpretable data formats to support the exchange of complex datasets and measurement results.

In this contribution, we present work towards implementing DCCs for radioactivity and related calibration quantities, following the XML -DCC pioneered by PTB. Calibration certificates must describe the calibration item (activity standard) in terms of its size, shape, and isotope composition, as well as the calibration quantities such as activity, activity concentration or uniformity. We will also discuss a proposed data exchange format based on the DCC schema that allows the addition of raw data such as count rates or gamma-ray spectra, and meta data such as observing times and experimental setup. This new data format will support the sharing of datasets following the FAIR principles (findable, accessible, interoperable, reusable), in particular improving the aspects of interoperability and reusability over existing data formats.



#### Back to "Table of Contents" above

## Presentation of Henrike Fleischhack





# Digital Calibration Certificates for radioactivity and FAIR data exchange

Henrike Fleischhack, Matthias Oliver Stein & Stefan Röttger 4<sup>th</sup> international DCC conference

## **Calibration Certificates in Radioactivity**







What's next?

DCC 4 ®



## Content




- 1. Introduction
  - Activity calibration -> DCC requirements
  - Existing formats for digital data exchange
- 2. Implementation of the DCC for activity standards
  - Nuclide nomenclature
  - Measurement results
- 3. Future extensions of the DCC format


2025-02-26 3 DCC 4 ⊕

## **Activity and the Becquerel**










## **Activity Standards**







## Requirements of a DCC for radioactivity



#### Calibration item (one per DCC):

Calibration results: Detected nuclides

uniformity, ...

Shape, size, macroscopic composition, ...

Typ:

Kenn-Nummer Serial number: Radionuklide:

AI-3231

#### 1. Juli 2016, 13:00 Uhr MEZ

| Nuklid<br>Nuclide | E/keV  | A/Bq  | A'/Bq | A*/Bq |
|-------------------|--------|-------|-------|-------|
| Fe-59             | 1099,2 | 0 (6) | 7     | 10    |
| Rb-84             | 881,61 | 1 (6) | 9     | 13    |
| Cs-134            | 604.7  | 2 (6) | 7     | 10    |

Associated with reference time and nuclide.

 Metadata: assumed half lives (with uncertainties); measurement methods.

Activity, activity per mass/volume, activity ratio,

Decision limit & limit of detection (for impurities).

2025-02-26

DCC 4 to

#### IEEE/ANSI N42.42 format



https://www.nist.gov/pml/radiation-physics/ansiieee-n4242-2020-version

"American National Standard Data Format for Radiation Detectors Used for Homeland Security"

- XML-based, with schema.
- Contains information on detectors, raw data & measurement results.
- Lacking in metrology information (e.g. units!)
- Partially extensible specific format for specific use case.

```
<NuclideAnalysisResults>
 <Nuclide>
 <NuclideIdentifiedIndicator>true
 <NuclideName>Cs-137</NuclideName>
 <NuclideIDConfidenceValue>0.99/NuclideIDConfidenceValue>
 <NuclideActivityValue>804.5<NuclideActivityValue>
 <NuclideActivityUncertaintyValue>0.5<NuclideActivityUncertaintyValue>
 </Nuclide>
</NuclideAnalysisResults>
```



#### **BIPM-KC** format



"Machine-readable data and metadata of international key comparisons in radionuclide metrology" (Coulon et al, <a href="https://doi.org/10.1088/1361-6501/accd0b">https://doi.org/10.1088/1361-6501/accd0b</a>, Michotte et al, <a href="https://doi.org/10.1088/0026-1394/61/1A/06005">https://doi.org/10.1088/0026-1394/61/1A/06005</a>)

```
<kc:measurement>

 xml or json embedded in pdf.

 <kc:referenceDate>2019-08-22T12:00:00Z</kc:referenceDate>
 <kc:solutionID>Tb-161IRA2019-1</kc:solutionID>
■ Supports metrology information 4P-PS-BP-CB-GR-CO 4P-PS-BP-CB-GR-CO/kc:methodID>4P-PS-BP-CB-GR-CO 4P-PS-BP-CB-GR-CO/kc:methodID>4P-PS-BP-CB-GR-CO
 <kc:description>4pi (PS) beta-gamma coincidence</kc:description>
 (e.g., D-SI).
 <kc:activity>

 Description of radioactive solution

 <dsi:value>61970</dsi:value>
<dsi:unit>\kilo\becquerel</dsi:unit>
 (InChi, SMILES)
 <dsi:expandedUnc>
 <dsi:uncertainty>0.582</dsi:uncertainty>

 Currently only supports

 <dsi:coverageFactor>1</dsi:coverageFactor>
 <dsi:coverageProbability>0.68</dsi:coverageProbability>
 </dsi:expandedUnc>
 single-nuclide measurements.
 </kc:measurementResult>
 <kc:relStdUncertTypeA>0.16</kc:relStdUncertTypeA>

 Under development.

 <kc:relStdUncertTypeB>0.56</kc:relStdUncertTypeB>
 </kc:activity>
 </kc:measurement>
2025-02-26
```

## Towards a DCC for radioactivity




- (PTB-)DCC: Flexible, powerful data format.
- Interoperability with other branches of metrology/QI.
- Mandated by PTB (as alternative to paper/pdf).
- Requires community conventions/good practices:
  - Nuclides.
  - Measurement results & quantities.
  - Calibration item.

2025-02-26 10 DCC 4 ⊕



#### **Nuclide definition**



- Isotope: defined by number of protons & number of neutrons: Tc-99, 99Tc, (Z=43, N=56), (Z=43, A=99), ...
- Some isotopes have several (metastable) isomers: Tc-99 vs Tc-99m
- Approaches for machine-readable nuclide ID:
  - Pick one format, e.g. Tc-99(m)
  - Refer to online database.
  - Fully define nuclear state (quantum numbers) in DCC.

2025-02-26 11 DCC 4 ⊕

## Suggestion: Nuclide definition







- Nuclides defined in <dcc:statement>
  - InChI (isotope)
  - J<sup>P</sup> quantum numbers (isomer)
- Specify nuclear data, e.g. half life.
- Reference via id/refld.





### **Suggestion: Activity measurements**



```
<dcc:result refId="Na-22">
 <dcc:name>
 <dcc:content lang="de">Aktivität von Na-
 <dcc:content lang="en">activity of Na-22</dcc:content>
 </dcc:name>
 <dcc:data>
 <dcc:quantity refType="basic_measuredValue"</pre>
 rad_activity" refId="Na-22">
 <si:real>
 <si:value>2.52</si:value>
 <si:unit>\mega\becquerel</si:unit>
 <si:dateTime>2022-08-01T13:00:00</si:dateTime> refld: Nuclide
 <si:expandedUnc>
 <si:uncertainty>0.06</si:uncertainty>
 <si:coverageFactor>2</si:coverageFactor>
 <si:coverageProbability>0.95
 </si:coverageProbability>
 </si:expandedUnc>
 </si:real>
 </dcc:quantity>
 </dcc:data>
</dcc:result>
```

#### Activity (Bq):

- <dcc:result> per nuclide.
- refType:
  - quantity (activity)
  - measured value vs limit of detection.
- Referenze time:
  - SI datetime?
  - Or metadata?

## Status of DCC for Radioactivity



#### Available refTypes and good practice examples for common situations:

- Activity calibration with single nuclide and mixtures.
- Check for impurities.
- Comparison with customer-provided measurements.

#### In progress:

- Detailed description of calibration item.
- Area sources with uniformity measurements.
- Calibration solutions sold by weight.
- Stylesheet to mimic look of current calibration certificates.

2025-02-26 DCC 4 &



### **Extension: FAIR data exchange**



- Data format for measurement results (e.g., spectra) and metadata.
- XML-based, compatible with DCC and D-SI.
- Use cases:
  - Digital validation of gamma-ray spectrometry software
  - Data exchange in research
  - Possible extension of DCCs (on request).




Image source: M.O. Stein, H.Fleischhack, S. Röttger: "TraGamma – A digital service for validating gamma-ray spectrometry analysis software", in: Measurement: Sensors 2024 (Proceedings of the IMEKO world congress 2024), 101471, in press, <a href="https://doi.org/10.1016/j.measen.2024.101471">https://doi.org/10.1016/j.measen.2024.101471</a>; licensed under <a href="https://doi.org/10.1016/j.measen.2024.101471">CC-BY-NC-ND</a>

2025-02-26 15 DCC 4 **⊕** 

# Thank you! Any questions?

Physikalisch-Technische Bundesanstalt Braunschweig and Berlin Bundesallee 100, 38116 Braunschweig Henrike Fleischhack: <a href="mailto:henrike.fleischhack@ptb.de">henrike.fleischhack@ptb.de</a>

Back to "Table of Contents" above

## Transport of DCC with cryptographic signature in PDF/A

Presenting author: Diego Coppa (INTI),

Email: dcoppa@inti.gob.ar

Additional authors: Siegfried Hackel siegfried.hackel@ptb.de, (PTB), Héctor Laiz

laiz@inti.gob.ar, (INTI)

Keywords: DCC, PDF-A, Machine-readable, Digitalization, Cryptographic signatures, DCC

transport



#### Abstract:

The DCC has been reported on in the context of many international conferences. The DCC format, which has been developed internationally with all parties involved (industry, laboratories, NMI, accreditors) under the coordination of PTB, is used and further developed by many. A comprehensive collection of quotations can be found in [1]. One question that has come up again and again over time is how DCC is transferred from the laboratory to the customer. Several methods have emerged, which have one thing in common: the DCC must be passed on with integrity and authenticity. This can be solved by using cryptographic signatures. There will probably also be presentations from industry at this conference on this topic. In some cases, there are government restrictions on the selection of cryptographic methods. For example, in some countries the algorithms are specified, and some of them are not implemented in PDF/A. In other countries, such as Argentina, only PDF/A is allowed for sending cryptographically signed documents from the government. This presentation will discuss how the DCC XML file can be transported using the PDF-A transport container. It will also be discussed how the human-readable content should be handled, which can then be seen in the PDF-A. In this context, PDF-A should not be viewed as a machine-readable Calibration Certificate. It is only with the DCC-XML in the PDF-A container that this composite becomes machine-executable or even machine-interpretable. [1] [2].

#### References:

- [1] Hackel, Siegfried, et al., Digitization versus Digitalization & Digital Transformation in the field of calibrations and their subsequent use (2024). Physikalisch-Technische Bundesanstalt (PTB). Available at: https://doi.org/10.7795/120.20240126
- [2] D. N. Coppa, et al., "Software for the Creation of Machine-Readable and Human-Friendly Reports" in 3rd International DCC Conference 2023-02-28 to 2023-03-02 Proceedings, p.18, 2023. doi: 10.7795/810.20230331. Available at: https://oar.ptb.de/resources/show/10.7795/810.20230315

Back to "Table of Contents" above

#### Presentation of Diego Coppa





# Transport of DCC with cryptographic signature in PDF/A

Authors: Diego Coppa (INTI), Siegfried Hackel (PTB), Hector Laiz (INTI)

February, Online, DCC Conference 2025






## Initially



- Joint lecture by INTI and PTB
- > Aim of presentation:
  - Demonstrating a way of DCC transport with PDF/A
- > Background:
  - In some cases it is necessary to transport the DCC in this way







- > The DCC has been created
- > ISO/IEC 17025:
  - Transport from CalLab to Customer
- How should that happen?
- Many answers...
- ➤ ISO/IEC 17025 does not address signatures
  - What about integrity and authenticity?
- Cryptographic methods are necessary





# What are the different types of signature?



Signature on the document: inbound signature



Signature on the envelope or transport container: outbound signature







Inbound vs. Outbound



## Inbound:

- Each one is robust
- > "DCC-Atom"

## Outbound:


Possibility to combine many documents





## **Inbound Signature**







## **Outbound Signature**



## > Some examples:







• • •






# Outbound Signature with PDF / A3





## Important:

- Digitization vs. Digitalization & Digital Transformation
- See https://doi.org/10.7795/120.20240126





### Please note





- Take care when creating human readable
  - Also important for the human-readable in the fourth ring of the DCC
  - Some reports are very extensive, e.g. in the high frequency
- What should be readable in PDF / A3 itself?
  - Note regarding DCC in the container
  - Any further information?











- Why? What? How?
- INTI's Traceability of documents
- Hashing Fundamentals: Uniqueness, Irreversibility, and Integrity
- PDF/A-3
- Authenticity and integrity for a disembeded XML





## Why? What? How?



Why does INTI have to ensure the traceability of an XML (DCC) without signing it?

INTI, Argentina's NMI, is only capable to sign Calibration Certificates in PDF/A-3. Since DCC-XML is digitally signed, its traceability must be ensured without direct signing.













What does INTI do to ensure traceability of XMLs that can be requested by machines or by human?

INTI links the XML to a traceable PDF/A-3 document.









What does INTI do to ensure traceability of XMLs that can be requested by machines or by human?

INTI links the XML to a traceable PDF/A-3 document.

How does INTI link an XML to a PDF/A-3 and provide traceability to an XML?

INTI uses PDF/A-3 and HASH properties to establish authenticity and integrity of the XML file.











Name: MassDcCxMl.xml HASH: SHA-256 HASH Value: 1f3a3732d1e9c5d9f 6f8bff07c9db8e56 c9f8e8c258d





# INTI's Traceability of documents




# INTI's Traceability of documents









#### ADVANTAGES

- Authenticity
- Integrity
- Non-repudiation

#### DISADVANTAGES


- Different typesScope of validation
- Increase of trust

# INTI's Traceability of documents












# INTI's Traceability of documents













## Hashing Fundamentals: Uniqueness, Irreversibility, and Integrity




### **Unique Hash Generation**





Random examples



### **Hash Irreversibility**








1f3a3732d1e9c5d9f6f8bff0 7c9db8e56c9f8e8c258d5a 4b60c2d8d7a2f3f173



### **Hash Irreversibility**








## **Hash Irreversibility**










### **Hash Integrity**








### **Hash Integrity**













## PDF/A-3

#### PDF, PDF/A, PDF/A-3





- 1. ISO standard 32000-1:2008 embed and longterm archiving
- 2. XMP as metadata of the PDF/A
- 3. Free programs/Python exist for displaying and editing PDF/A files.
- 4. The multi-part PDF/A standard offers great flexibility to users.
- 5. PDF/A files can be used on any platform.
- 6. PDF can act as a container (similar to ZIP)
- 7. XSL-FO can be used to create PDFs from XMLs

ISO 32000-1:2008. Document management - Portable document format – Part 1: PDF 1.7. International Organization for Standardization. https://www.youtube.com/watch?v=g-tJRSsZHyc










## How is a PDF/A-3 composed?










## How is a PDF/A-3 composed?












## How is a PDF/A-3 composed?










## How is a PDF/A-3 composed?
















# Authenticity and integrity for a disembeded XML

# Authenticity and integrity for a disembeded XML














# Authenticity and integrity for a disembeded XML










# Authenticity and integrity for a disembeded XML












# Authenticity and integrity for a disembeded XML










Boschung, Gregor & Wollensack, Michael & Zeier, Markus & Blaser, Cédric & Hof, Christian & Stathis, Manuel & Blattner, Peter & Stuker, Florian & Basic, Nina & Grasso Toro, Federico. (2021). PDF/A-3 solution for digital calibration certificates. Measurement: Sensors. 18. 100282. 10.1016/j.measen.2021.100282.

# Authenticity and integrity for a disembeded XML











#### **Conclusions**









- PDF/A-3 is a MACHINE READABLE and a HUMAN READABLE format.
- PDF/A-3 can act as a storage container (similar to ZIP).
- XSL-FO to assure no misinformation
- PDF/A-3 assures long term stability https://www.iso.org/standard/57229.html.







## Thanks!

dcoppa@inti.gob.ar Siegfried.Hackel@ptb.de laiz@inti.gob.ar

Back to "Table of Contents" above

### 023 A service to authenticate documents issued by MSL

Presenting author: Blair Hall; Measurement Standards Laboratory of New Zealand

Email: blair.hall@measurement.govt.nz

Additional authors: Joseph Borbely joseph.borbely@measurement.govt.nz; Measurement

Standards Laboratory of New Zealand

#### <u>Abstract</u>

This talk describes an online service that is being developed at the Measurement Standards Laboratory of New Zealand (MSL) to authenticate documents. The service will allow clients to check whether a digital document in their possession was issued by MSL and is currently valid.



Currently, MSL issues calibration reports either as PDF files or paper documents. PDF reports are sent to clients as email attachments, accompanied by a digital checksum to allow an integrity check of the file received.

The planned new web-based service will provide greater flexibility. By submitting a document's checksum to the service, clients can query its status at any time. The service will return one of three possible responses:

- Valid: The document was issued by MSL and is currently valid.
- o Cancelled: The document was issued but has since been cancelled and is no longer valid.
- o Unrecognised: There is no record of the document, indicating it is not valid.

To ensure confidentiality, the service relies on checksums and does not expose document contents. Additionally, it does not require user authentication, simplifying the verification process for both MSL and clients.

This new service will enable MSL to indicate a cancelled report status, which is difficult to communicate effectively under the existing system. It will allow clients to perform dynamic document verification, protecting against inadvertent corruption at any time after receiving a file. Furthermore, the service can authenticate other file types, such as spreadsheets. This capability will enable MSL to provide supplementary information alongside reports in formats suited to clients' needs.

Back to "Table of Contents" above

#### Presentation of Blair Hall






#### Why are we doing this?

- MSL issues electronic calibration and measurement reports (PDF)
- We currently provide a checksum with each report
- We wish to offer a better service to our customers



#### What will this project deliver?

- An online verification service for electronic documents issued by MSL
- Clients can verify the integrity of documents at any time
- MSL can change the status of documents
- Long-term solution with low cost of ownership for MSL







#### How does it work?

Anyone can query a cloud-based service—no authentication needed



#### Provide checksums to the service and receive a response

58463ec72964de1bf98f4ff67eee3aa2 fc4c68f094e3647c2c6ab8694544bcdd 72b302bf297a228a75730123efef7c41



}







"58463ec72964de1bf98f4ff67eee3aa2": "valid"
"fc4c68f094e3647c2c6ab8694544bcdd": "cancelled"

"72b302bf297a228a75730123efef7c41": "unrecognised"



## Private (MSL)



### Public (Cloud)

Anyone can query (no login required)



Pick a single cloud provider...







### Private (MSL)

MSL maintains three local text files: 'valid.txt', 'cancelled.txt', 'log.txt', which are safe behind the corporate firewall

Two files: 'valid.txt' and 'cancelled.txt', containing only checksums, are uploaded to a cloud server (using an authenticated—access key—connection)

No confidential information is stored in the cloud













#### Local (MSL) command line tool







#### valid.txt

- · valid checksums
- · one checksum per line

#### cancelled.txt

- · cancelled checksums
- one checksum per line

#### log.txt



Stores information about each operation

2024-11-28T12:22:37 [ valid ] [ci\j.borbely ] Appended 8cd5a16fe0274922f62bc4a5df37bdbe to "nas\valid.txt" 2024-11-28T12:22:37 [ cancel ] [ci\j.borbely ] Moved 227bc609651f929e367c3b2b79e09d5b from "nas\valid.txt" to "nas\cancelled.txt" 2024-11-28T12:22:37 [rollback] [ci\j.borbely ] Moved 227bc609651f929e367c3b2b79e09d5b from "nas\cancelled.txt" to "nas\valid.txt"

#### Query

https://kapua.measurement.govt.nz/checksum?9d38a9a45d24640cc33b583507465e09

```
checksum?9d38a9a45d24640cc33b583507465e09
Pretty-print

 "9d38a9a45d24640cc33b583507465e09": "valid"
}
```

Or with the command line tool at MSL

```
PS C:\Users\j.borbely> checksum query 9d38a9a45d24640cc33b583507465e09 8cd5a16fe0274922f62bc4a5df37bdbe kiwi
 "9d38a9a45d24640cc33b583507465e09": "valid",
 4:
 "8cd5a16fe0274922f62bc4a5df37bdbe": "cancelled",
 "kiwi": "unrecognised"
```



#### Final remarks

- Accepted in-principle by the MSL Chief Metrologist and MSL Quality Manager
- Currently a beta-version is being evaluated by MSL's Quality Council (responsible for the quality system)
- A more detailed description of the idea, with a proof-ofconcept working example in Python, is available: <a href="https://apmp-dxfg.github.io/dxfg-checksum-tut">https://apmp-dxfg.github.io/dxfg-checksum-tut</a>
- We are happy to get feedback about this project.





Back to "Table of Contents" above

# O24 Progressive Shift towards Digitalization in TUBITAK UME Mass Laboratory

Presenting author: Beste Korutlu Email: <a href="mailto:beste.korutlu@tubitak.gov.tr">beste.korutlu@tubitak.gov.tr</a>

Additional authors:

### **Abstract**

In this rapidly evolving digital era, the adoption of Findable, Accessible, Interoperable and Reusable (FAIR) Data Principles is inevitable in metrological data and metadata. Digital



calibration certificates (DCC) and digital comparison reports emerged as the outcome of the comprehensive work within this period. The concept of this paper is built on progressive work towards a synergic co-operation of mass metrology and digitalization.

In mass metrology, the calibration of the test weight is typically performed by substitution weighing with reference standard of higher accuracy class. The most commonly used weighing cycle is Reference-Test-Reference (or in-short ABBA-cycle where "A" represents weighing the reference standard and "B" represents weighing the test weight), thereby eliminating the linear drift. The ABBA-cycle is repeated at least by the minimum number specified in the relevant international recommendations/standards/guidelines for the particular accuracy class of the test weight. The average weighing difference between the test weight and the reference standard is used in finding the error in the mass of the test weight. During the calibration, the environmental conditions are recorded in order to take into account the air buoyancy effect in the weighing process.

TUBITAK UME Mass Laboratory has recently undergone an infrastructure renewal in weighing instruments in the range of 2 kg - 50 kg. The calibration reports by these instruments are in human readable Tanner Database format with .tdb extension. We reconstruct the calibration report into a machine-readable Extensible Markup Language format having .xml extension by developing precise functions in Python Programming Language. The calibration report is composed of two parts as administrative data and measurement results. Initially, the report is split into two so that the administrative data is stored for further producing a digital calibration certificate while the measurement results for determining the error in the mass of the test weight. The analysis of the data is performed using the calibration data in the measurement results and the relevant data of the reference standard.

On the onset of this work, we use GEMIMEG tool in order to generate a DCC for the reference standard with .xml extension. Once the reference standard data is in the machine-readable format, it will be possible to call it into the same environment with the measurement results and assign the mass error for the test weight without human-intervention. In this paper, we demonstrate how the mass error of a single test weight is determined with minimum involvement of the operator.



#### Back to "Table of Contents" above

### Presentation of Beste Korutlu

## **Progressive Shift Towards**

## DIGITALIZATION

### in TUBITAK UME Mass Laboratory

Assoc. Prof. Dr. Beste Korutlu Yılmaz Head of Mass Laboratory TUBITAK UME Türkiye


26.02.2025

#### 5TH INTERNATIONAL DCC CONFERENCE

DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

## **Current Calibration Flow**








### DIGITALIZED CALIBRATION FLOW





3








Mean value of weighing difference 

Net force on test − Net force on reference











#### **INGREDIENTS**











Mass Lab




### Libraries


















### Current DDC of the Reference












 $\boldsymbol{m_t} = \boldsymbol{m_r} + \boldsymbol{\rho_a}(\boldsymbol{V_t} - \boldsymbol{V_r}) + \overline{\Delta \boldsymbol{m}}$ 

 $V_t = 1248.1 \, \text{cm}^3$ 







12



## True Mass of the Test



#### $\boldsymbol{m_t} = \boldsymbol{m_r} + \boldsymbol{\rho_a}(\boldsymbol{V_t} - \boldsymbol{V_r}) + \overline{\Delta \boldsymbol{m}}$

m\_test\_true = vs[0] + rho\_a \* (vs\_test[2]-vs[2]) \* 1E-3 + mean\_diff\_value
print(f"True Mass of Test (g): {m\_test\_true}")

True Mass of Test (g): -0.001754999999999998

Root element: {https://ptb.de/dcc}digitalCalibrationCertificate True Mass Error of Test in Previous Certificate (g): -0.0013 Conventional Mass Error of Test in Previous Certificate (g): -0.001

Volume of Test in Previous Certificate (cm^3): 1248.1 Nominal Mass of Test (g): 10000.0

 $\Delta m_t = (-0.0018 \,\mathrm{g}) - (-0.0013 \,\mathrm{g}) = -0.0005 \,\mathrm{g}$ 


Is it smaller than the uncertainty?

COMING SOON

13

### FUTURE TASKS









Apply the concept with a volunteering customer for interoperability

Drag and Drop the Calibration Report

Drag and Drop Current DCC of Reference

Drag and Drop Previous DCC of Test

Click to Generate DCC of Tes

14





Back to "Table of Contents" above

# O25 Practical Applications of Digital Calibration Exchange (DCX) covering Certificates, Requests, and Templates

Presenting author: Søren Kynde; DFM – Danish National Metrology Institute

Email: srk@dfm.dk

Additional authors: David Balslev-Harder dbh@dfm.dk

### **Abstract**

This presentation provides a comprehensive exploration of the Digital Calibration Exchange (DCX) through practical examples and scenarios. We begin by examining four general scenarios of calibration, and how this creates a basis for simplifying the representation of calibration results in a FAIR maner. These scenarios will be demonstrated with XML examples, highlighting the versatility and adaptability of the DCX framework. A key focus will be on the PDF representation of XML data, demonstrating how similarities across these scenarios can be leveraged to create standardized and easily interpretable documents. The presentation will also delve into the commonalities observed across different metrology areas, showcasing how DCX facilitates a unified approach to calibration data management. By presenting examples of Digital Calibration Certificates (DCCs), Digital Calibration Requests (DCRs), and Digital Calibration Templates (DCTs), we will illustrate the practical implementation and benefits of DCX in real-world settings. Furthermore, we will outline the workflow involving templates, requests, and certificates, providing a clear understanding of how these components interact within the DCX framework. This workflow ensures a seamless and efficient process for managing calibration data, from initial request to final certification. In conclusion, the DCX framework offers a robust and standardized solution for the digital management of calibration data, enhancing accuracy, traceability, and interoperability across the metrology community. Outline of presentation

- Four general scenarios of calibration
- o PDF representation of xml certificates utilizing the similarities across the four scenarios
- o Similarities across different metrology areas



- o Workflow with templates, requests and certificates
- Examples of Digital Calibration Certificates (DCC's)
- Examples of Digital Calibration Requests (DCR's)
- Examples for Digital Calibration Templates (DCT's)
- o Solution available on github.com/TC-IM-1448/DCC-Tables
- o See other presentations for information on schema and software tools.

DFM's work is supported by funds from The Danish Agency for Institutions and Educational Grants.

Back to "Table of Contents" above

### Presentation of Søren Kynde



Digital calibration exchange: DCX

Practical applications with examples

Søren Kynde, DFM David Balslev-Harder, DFM 5'th DCC-conference 2025-02-26



### Motivation for DCX

- Explore impact on middleware architecture when schema is designed w. focus on:
- Use of references rather than emphasis on tree structure (PTB-DCC approach):
  i. Extend use of xs.ID and xs.IDREF and utilizing xs.key and xs.keyref to restrict pointing options
  - Introduce new attribute-names and restricted options to guide machine interpretability
  - avoid multiple re-entry of identical data
  - Represent data content in tables for easier integration with Excel and databases
    - and provide better overview in user interface experience.
  - $\begin{tabular}{ll} Emphasis on descriptive query expressions for finding and interpreting measurement data. \end{tabular}$
  - Same elements/structure for all calibration/metrology areas.
  - Define as much as possible in one Schema, to simplify interoperability and facilitate validation data machine interpretability and ability level to support automatic data extraction.
- DCC is more than just a data-file, it may act as a legal document.
  - Support human readable fields for all elements in the xml-file.
  - Support PDF-generation w. low level of middleware assumptions.
- Simplify software/middleware management
  - Use schema in middleware to adapt for received xml's used schema version.
  - Provide tools for Excel that support file reading across different metrology area's

### Motivation for DCX

- Explore impact on middleware architecture when schema is designed w. focus on:
  - Use of references rather than emphasis on tree structure (PTB-DCC approach):
    - Extend use of xs.ID and xs.IDREF and utilizing xs.key and xs.keyref to restrict pointing options
    - Introduce new attribute-names and restricted options to guide machine interpretability
    - avoid multiple re-entry of identical data
  - b) Represent data content in tables for easier integration with Excel and databases
    - and provide better overview in user interface experience.
  - Emphasis on descriptive query expressions for finding and interpreting measurement data.
  - Same elements/structure for all calibration/metrology areas.
  - Define as much as possible in one Schema, to simplify interoperability and facilitate validation data machine interpretability and ability level to support automatic data extraction.
- A calibration certificate is more than just a data-file, it may act as a legal document.
  - Support human readable fields for all elements in the xml-file.
  - Support PDF-generation w. low level of middleware assumptions.
- Simplify software/middleware management
  - Use schema in middleware to adapt for received xml's used schema version.
  - Provide tools for Excel that support file reading across different metrology area's



## A modular build of a calibration certificate

#### Calibration certificate

- · Administrative data
- Statements
- · Description of equipment
- Specification of equipment settings
- List of uniqure measurement configurations
- Results
  - Each result refers to a unique measurement configuration
- Annexes
  - Definitions of non SI-units
  - · List of embedded files
  - List of codeable elements

## Example

• Thanks to Dandiag for sharing this sample certificate





### Administrative data

#### Core data

- · Core data has a heading
- All other element in core data are fields
  - · A field has a value
  - All fields may have a heading
  - The heading can be stated in one or more languages



### Administrative data

#### Accreditation


- The dcx:accreditationType consists of a heading and a number of fields
- Information can be given in accordance with ISO17025

### Administrative data

#### Contacts and locations

- The contactAndLocationType consists of an address and contact info
- Contacts and locations are organized as a table with common headings







#### Used equipment

ID	Used equipment	Manufacturer	Product name	Product type	Customer ID	Serial number	DFM num- ber
item1	Pipette with 12 channels 50-1200 µm	Biohit	Picus	-	ID107	12008522	-
item2	Pipette tip (1200 µl)	Satorius	Optifit	-	-	8823987	-
ref1	Balance. resolution: 0.00001 g. calibrated:	Satorius	-	-	-	-	ID148

#### Settings

ID	-	description		value	unit
			parameter		
ch1	Channel	-	-	1	-
ch2	Channel	-	-	2	-
ch3	Channel	-	-	3	-
sin	Speed in	Same speed is used in all measure-		5	-
		ments	speed-in		
sout	speed out	Same speed is used in all measure-		5	-
		monte	enood out		

#### Measuring Systems Under Calibration

ID	Measuring system	Client equipment	References used	Setting ref.	Method ref.	Operational status.
ms1	Channel 1	item1 item2	ref1	ch1 si sout	in meth2	as found
ms2	Channel 2	item1	refl	ch2 si	n meth2	as found

#### Used equipment

ID	Used equipment	Manufacturer	Product name	Product type	Customer ID	Serial number	DFM num- ber
item1	Pipette with 12 channels 50-1200 µm	Biohit	Picus	-	ID107	12008522	-
item2	Pipette tip (1200 µl)	Satorius	Optifit	-	-	ss23987	-
ref1	Balance. resolution: 0.00001 g. calibrated: 2021-11-25.	Satorius	-	-	-	-	ID148

### Measurement results

	omene res							
REF	Pipette -	channel1						
calRes0	Measuring	System:ms1						
Cus- tomer Tags	Cali- bration point	Reference Value	Data-series used for calculation of repeatability and systematic error	Systematic error	Systematic er- ror (relative)	Repeat a- bility	Repeat a- bility (relative)	Excep- tion to ac- credi- tation
1	$\mu L$	$\mu L$	1	μL	%	$\mu L$	%	-
pt1	120	121.21	ser11	$1.21 \pm 0.35$	$1.01 \pm 0.29$	0.13	0.11	-
pt 2	600	600.3	ser12	$0.3 \pm 1.5$	$0.05 \pm 0.25$	0.24	0.04	-
- 1.9	1000	1100.1	1.0	00 1 04	0.00 1 0.0	0.46	0.04	200000


#### Settings

ID		description				valu	e unit	,
					parame	eter		
ch1	Channel	-			-	1	-	1
ch2	Channel	-			-	2	-	
ch3	Channel	-			-	3	-	-
sin	Speed in	Same speed	is used in all	me asure-		5	-	
		ments			speed-	-in		1
sout	speed out	Same speed	is used in all	me asure-		5	-	1
		ments			speed-	-out		1
								1
Measur	ing Systems Under Cal	ibration						1
ID	Measuring system	Client	References	Setting	Met	hod	Operational stat	tu I
		equipment	used	ref.	ref.			
ms1	Channel 1	item1	ref1	ch1 :	sin met	h2	as found	
		item2		sout				

REF	Series of	measurement	channel 1	REF	Series of	measuremen	t channel
ser11	Measuring	System:ms1		ser12	Measuring	System:ms1	
Cus- tomer Tags	Cali- bration Point	Reference value	Item indi- cation	Cus- tomer Tags	Cali- bration Point	Reference value	Item indi cation
-	$\mu L$	$\mu$ L	$\mu$ L	-	$\mu$ L	$\mu$ L	$\mu$ L
p1	120	121.3229	120	p1	600	600.5463	600
p2	120	121.1223	120	p2	600	600.466	600
p3	120	121.1724	120	р3	600	599.784	600
p4	120	121.1022	120	p4	600	600.5262	600
p5	120	121.2025	120	$p_5$	600	600.4058	600
p6	120	121.1825	120	p6	600	600.1049	600
p7	120	121.3129	120	p7	600	600.3457	600
p8	120	121.042	120	p8	600	600.3757	600
p9	120	121.1724	120	p9	600	600.7067	600
p10	120	121.5034	120	p10	600	600.1952	600




## ...multiple language support




REF ser11	Series of measurement channel 1 Measuring System:ms1			REF ser12	Series of measurement channel 1 Measuring System:ms1			
Kundens kode	Kalibre- ringspunkt	Referen- ceværdi	Emnets visning	Kundens kode	Kalibre- ringspunkt	Referen- ceværdi	Emnets visning	
	μL	μL	$\mu$ L	-	μL	μL	$\mu L$	
p1	120	121.3229	120	p1	600	600.5463	600	
p2	120	121.1223	120	p2	600	600.466	600	
р3	120	121.1724	120	р3	600	599.784	600	
10	***	*** ****	4.000		ann	nnn wann	000	

## Workflow






## Workflow



## Four types of calibration.





## A modular build of a calibration certificate

#### Calibration certificate

- •
- Results
  - Each result refers to a unique measurement configuration
  - Important columns are:
    - Reference value
    - · Indication value
    - Bias = indication value reference value
    - · Uncertainty of bias
    - ...
- ...

## Example of DCR (temperature)




## Example of DCR (temperature explicit)


## Example of DCR (With tolerances)



## Workflow




## Workflow



Released january 2025! https://github.com/TC-IM-1448/DCC-Tables



## Thank you!



Released january 2025! https://github.com/TC-IM-1448/DCC-Tables

Back to "Table of Contents" above

Day 3 (2025-02-27 Thursday)
Plenary Session
(Chair: Hector Laiz)

### P06 Motivation for using DCC

Presenting author: Martin Koval, CMI, Czech Metrology Institute

Email: <a href="mkoval@cmi.cz">mkoval@cmi.cz</a>
Additional authors: Nil

### **ABSTRACT**

The concept of DCC emerged in the field of digital metrology almost a decade ago. Despite its long history and many advances in this field, DCC has not yet gained wide acceptance by the metrology community. Over the years, several key steps have been taken in the promotion and implementation of DCC. Working Groups have been set up, allowing the coordination of activities at international level. Projects that have provided practical examples of DCC implementation. A number of conferences, trainings and workshops have been held and are planned to help disseminate information on DCC. There are even freely available software tools and case studies. DCCs off er a multitude of benefits including, better access to data, efficiency in process management (time savings, data analysis, flexibility) and many more.

From the many discussions in terms of motivation, we can identify three main key groups. Group 1 Users who want to benefit from the majority of DCC benefits. This group often manage hundreds to thousands of devices and need an effective digital solution. Group 2 Users who have a positive attitude towards DCC but are waiting for final solutions or a broader consensus in the metrology community. This group needs to work with calibration data, but for the time being makes do with the different mix of solutions available, whether it is classic PDF or paper. Group 3 are users who only need a calibration certificate for auditing purposes. Usually these are instruments that are in lower units and the value of the information from the calibration



certificate is not very meaningful to them. Hence, we can also deduce several remaining barriers to full implementation of DCC which include financial but also capacity constraints. Many organizations do not have the resources to create additional jobs or to maintain digital systems. Most importantly, they are concerned about price increases for IT-related services. Another major barrier is the conservative approach where practices in metrology prevent the adoption of innovative technologies where information on paper with a stamp is trusted more than data in digital form. Thus, the question remains how to effectively motivate users to adopt DCC.

One of the key problems is that society has lost its understanding of the importance of calibration and metrology in general. We are increasingly seeing metrology confused with meteorology, which is a sign that metrology has a weaker awareness. One of the first steps for motivation should be to raise awareness of metrology. Despite the fact that every person is surrounded by thousands of sensors every day, few are aware of the fact that the quality of services and products often depends on the correct calibration of instruments and sensors and, above all, the correct and efficient use of calibration data. Metrology is so often a background issue. Therefore, one of the necessary steps is to raise awareness of metrology, which will undoubtedly contribute to the growing curiosity and interest in what accurate and reliable data looks like. At the same time, a period is coming when the emphasis will again be on quality of services and products over quantity, where DCC can play a key role.

Back to "Table of Contents" above


Martin Koval

### **Presentation of Martin Koval**

**METROLOGY** 

INSTITUTE


### 5th INTERNATIONAL DCC CONFERENCE





- It is one of the fundamental pillars of digitalization in metrology.
- Brings many advantages over "still widely" deployed solutions.
- It can bring popularity to metrology, especially among the younger generation.





- The DCC concept has been around for several years.
- DCC is a topic at almost every conference that includes digitalization in metrology.
- Several working groups and projects dealing with DCC have been established around the world - most of them communicate with each other.
- There are implementations of DCC.
- There are guidance documents on how to get started with DCC.




• DCC is still not widely deployed.





What should bring "D" for "CC"

- · better access to data (streamline information),
- · machine readable,
- the possibility of implementation into internal systems,
- · the possibility of automatization of metrological activities,
- · with higher numbers of calibrations, easier and more effective management,
- "Green deal effect" elimination of papers, reduction of the load on resources (distribution, printing, etc.),
- · Life cycle monitoring of sensors/devices,
- DCC can make it easy to analyze calibration data and create digital twins that help increase efficiency and safety in the process industry,
- DCC can support digital transformation by enabling suppliers and laboratories to connect to a digital calibration system with centralized management easily,
- DCC should use a standardized approach to data entry, which facilitates the comparison and harmonization of data from different sources.
- DCC can enable preventive maintenance by alerting the need for instrument inspection instead of relying on fixed intervals, leading to a better risk-based approach to maintenance and calibration.
- DCC can increases the speed of traceability by replacing paper-based processes with easy digital search capabilities.
- DCC should be flexible so stakeholders can use their preferred calibration processes while still creating easily shareable and searchable digital certificates.
- · and others..
- · Instant access to calibration data.
- · Easier tracking and compliance.
- · No risk of losing paper documents.
- · Faster decision-making.
- · Secure and traceable storage.







- · Faster data processing.
- · Less manual work.
- · Automation of calibration workflows.
- · Reduced human errors.
- · Less storage space required.







- The DCC concept is (should be) ready to adapt to new technologies.
- · Scalable for different areas.
- Flexible for large and small organizations.
- Supports remote audits and inspections.







- USERS
- AUDITORS/REGULATORS
- PUBLISHERS



### 1 Group

- Manage thousands of devices.
- Require automation and integration.
- High data volume processing.
- Strong motivation for DCC adoption.







### 2 Group

- Interested in DCC, but hesitant.
- · Waiting for standardization.
- Use a mix of digital and paper.
- · Concerned about compatibility.
- · Need proven success cases.
- Open to gradual transition.






### 3 Group

- · Only need a certificate for audits.
- Use a small number of instruments.
- · No real benefit from digitalization.
- Prefer traditional paper format.
- · Concerned about IT-related costs.











- Each country should define its own interpretation of DCC.
- What are the differences in conducting an audit between CC and DCC?
- · Auditors have a weaker awareness of what DCC actually entails.
- · Set requirements for DCC.
- Legal acceptance of DCC is not guaranteed everywhere if digital certificates are not formally recognized, users will continue to prefer "paper/electronic" version of CC.








- Many Publisher are ready to start issuing DCC or are working on it.
- The rule is that if there is demand, the service will be provided.
- Not every country has a demand for DCC.



- High initial costs for DCC implementation.
- · Lack of IT infrastructure in small organizations.
- · Limited staff capacity.
- · Need for investment in training and education.
- Ongoing maintenance and support costs.










- · Strong preference for paper certificates.
- Trust in physical stamps & signatures.
- · Low level of trust in new technologies.
- · Resistance to change in established workflows.
- Lack of regulatory enforcement for DCC.



- · High initial investment in IT infrastructure.
- Cost of software licenses and updates.
- · Need for cybersecurity and data protection.
- Ongoing maintenance and technical support.
- · Uncertainty about cost-benefit ratio.









- · Uncertainty about legal acceptance of DCC.
- · Lack of clear guidelines for digital audits.
- · Resistance from auditors used to paper CC.
- Need for secure IT systems for DCC.



Main Barriers to DCC Adoption

How to pass an audit



- Many people confuse metrology with meteorology (never-ending story).
- Meteorology is more attractive than metrology to the general society.
- · Metrology often overlooked despite its impact.







- · Calibration ensures product quality and safety.
- · Impact on health, industry and daily life.
- · Unnoticed but essential for reliable measurements.
- · Lack of interest outside metrology community.
- · Low media coverage on metrology topic.





General public does not recognize calibration as crucial.



- · People already rely on digital tools daily.
- Younger generations knows just "online" data.
- · Consumers research product details before buying.
- · DCC can make calibration data more visible.
- Digital access increases transparency and trust.
- · Widely deployment of DCC can brings better interest in metrology.







#### Positive motivation:

- · There are many benefits.
- · Successful user stories.
- · Guidelines document on how to get started.

#### Forced by circumstances:

- · Young generation.
- Future technology development.





This work was partly funded by resources granted to the Czech Metrology Institute by the Ministry of Industry and Trade within the frame of internal research task IF2403601101 and UTR25E601102.



Ing. Martin Koval , Ph.D.
Czech Metrology Institute
Digitalization of metrology
Mobil: + 420 725 504 983
e-mail: mkoval@cmi.cz



Back to "Table of Contents" above Back to "Table of Contents" above

P07 Modeling Digital Calibration Certificates Using Object -Oriented Technologies to Address Heterogeneity and Harness FAIR Principles

Presenting author: Dr. Anjali Sharma; NPL, India

Email: anjali@nplindia.res.in


Additional authors: Niharika Bhatia



### **ABSTRACT**

The increasing digitization of calibration processes necessitates a unified and standardized approach to managing calibration data. This work explores the application of object-oriented technologies in the development of Digital Calibration Certificates (DCCs) to address the heterogeneity of calibration data formats and protocols. By employing a structured object-oriented model, the system resolves compatibility issues across diverse instruments, laboratories, and measurement standards, enabling seamless integration and data exchange. The proposed DCC model leverages key object-oriented concepts such as encapsulation, inheritance, and polymorphism to defi ne and manage calibration entities such as certificates, instruments, calibration methods, and environmental conditions. These objects are mapped to essential attributes in calibration systems, ensuring consistency and reusability across platforms. Through this approach, the model promotes interoperability, extensibility, and the ability to handle complex data relationships inherent to calibration activities.

### Presentation of Anjali Sharma





#### 5<sup>th</sup> INTERNATIONAL DCC CONFERENCE

Modelling Digital Calibration Certificates Using Object oriented Technologies To Address Heterogeneity & Harness FAIR Principles

#### Dr. Anjali Sharma

Sr Principal Scientist & Assoc. Prof. AcSIR anjali@nplindia.res.in



### Introduction



- Calibration ensures measurement reliability and traceability.
- DCCs aim to offer a standardized, traceable, and secure solution
- Future calibration certificates will be Digital Calibration Certificates (DCCs)
- DCCs will be successful if we achieve
  - Broad willingness of internal and external stakeholders to engage with DCCs
  - Interoperability



5th International DCC Conference , 27 February 2025, NPLI

### Introduction

- Different level of implementations.
- In 2023-24, 05 APMP NMIs started with the DCC interoperability study.
- The findings were discussed in DCC 4 conference, 2024
- Few APMP NMIs have conducted surveys and awareness programs to understand the APMP Stakeholder's viewpoints about DCC
- Major findings were presented in the BIPM FORUM-MD plenary meeting on 21st Feb 2025




5th International DCC Conference , 27 February 2025, NPLI



## **Schema Comparison Gap Analysis**





5th International DCC Conference , 27 February 2025, NPLI

## **Concerns & Strategy**



#### Stakeholder's Concerns

- \* Will it require continuous reengineering of code/systems developed for the implementation of DCC (w.r.t new Schema launched)?
- ❖ How DCC will integrate with the existing systems?
- ❖ How DCC will cater to the national and international
- What will be the impact on their business if they do not accept DCC?
- Why should we adopt DCC?

Its too early to invest into the infrastructure, human resources and organizational changes required for implementation of DCC.

#### Strategy

- Harmonized and standardized DCC schema
- Showcase direct benefits through early DCC adopter success stories/Lighthouse Projects
- Build trust by providing increased document security and better turnaround time.
- Incremental Approach: Provide bundle measurement data with the certificate on request.
- Solution should not require lot of changes at the receiving end
- Give them a business model which shows how its helpful
- NMIs can Push but Pull from stakeholders is critical for DCC to be a success
- · DCC is the future but needs more



## Why?

- Heterogeneous data formats
- Heterogeneous calibration schema formats, semantics, languages, laws & regulations, data formats etc.
- Manual errors and inefficiencies
- Lack of interoperability across platforms
- Difficulties in traceability and standardization



### What?



- A unified, machinereadable and interoperable format for calibration certificates.
- Traceable measurement data
- Semantics & data
- Modelling

5th International DCC Conference , 27 February 2025, NPLI

Ь

### Harmonization



- Forum-MD Task Group on DCC & DRMC Harmonization
- APMP-DXFG DCC Interoperability Task Group
- TC-IM 1448
- PTB-DKD

Schema Conversion and Validation

Semantic framework

**Definition of canonical terms** 

Identification of canonical terms

5th International DCC Conference , 27 February 2025, NPLI



## **Understand the Existing Schemas**



- . **Schema Mapping**: Analyze the existing schemas to understand the different formats, data types, and structures used in various calibration certificates.
- . Identify Key Elements: Identify the common elements (e.g., measurement data, calibration date, instrument details, etc) and the specific differences in how these elements are represented across the schemas.

5th International DCC Conference , 27 February 2025, NPLI

### Harmonization



- **Create a Canonical Schema**: Design and define in unabmigious way a unified or canonical schema that covers the essential components of all the source schemas.
  - . Identification of fields that are mandatory and mapped to ISO 17025
  - . Not mandatory fields mapped to ISO 17025
  - . Mandatory for organization itself, but it is not mandatory in 17025
  - · Fields required for digitalization
- Administrative Data:
- Measurement Results: Representation of measurement data, quantities, scales, units, traceability, and uncertainty.

5th International DCC Conference , 27 February 2025, NPLI



### How?




- Develop a scalable, modular, and ISO-compliant DCC system.
- Implement a Model-View-Controller (MVC) framework for system organization.
- Leverage object-oriented programming (OOP) for flexibility and maintainability.
- Ensure data security, integrity, and authenticity using cryptographic methods.
- Support interoperability and digital transformation in metrology.

5th International DCC Conference , 27 February 2025, NPLI

10

## **NPLI's Approach**



PROCEED AS PREVIOUSLY DONE





5th International DCC Conference , 27 February 2025, NPLI

11



### **DCC Generation Workflow**



- 1) User submits calibration request via web interface.
- 2) Calibration process is executed & relevant data is recorded.
- 3) DCC generated in machine-readable PDF/A-3 format.
- 4) SHA-256 checksum is applied for integrity verification.
- 5) Certificate securely stored & delivered to the customer.

5th International DCC Conference , 27 February 2025, NPLI

12

## **System Architecture Overview**



#### MODEL-VIEW-CONTROLLER (MVC) APPROACH

- Model: Defines data structures and business logic.
- ❖ View: Presents user interface via Flask.
- Controller: Manages communication between model and view.

#### **OPEN-SOURCE TECHNOLOGIES USED**

- **Python v3.11.7** \* ReactJS v18.3.1
- Pandas v2.1.4 **❖** Flask v3.0.3
- **❖** Miktex v24.4 (Windows) PostgreSQL v16.3













5th International DCC Conference, 27 February 2025, NPLI



## **Object Oriented Approach**



ENCAPSULATION POLYMORPHISM MODULARITY

#### **ADVANTAGES**

- •Improved Maintainability Easier debugging and enhancements.
- •Enhanced Collaboration Clear structure helps teamwork.
- •Better Code Efficiency Optimized through abstraction and reuse.

#### IMPLEMENTATION

- Used Encapsulation to bundle data and methods, improving modularity.
   Applied Inheritance to reuse code
- •Applied **Inheritance** to reuse code across related components.
- •Leveraged **Polymorphism** to design flexible and scalable systems.

#### **KEY FEATURES**

- Modularity: Components are selfcontained and easy to manage.
   Reusability: Common functionalities
- are inherited, reducing redundancy.
  •Scalability: New features can be integrated seamlessly.

5th International DCC Conference, 27 February 2025, NPL

1/


## **Class Diagrams-Key Entities**



- 1. CalibrationCertificate
- 2. DeviceUnderTest
- 3. CalibrationProcess
- 4. CalibrationMetrologists
- 5. CustomerDetails
- 6. TestMeasurement
- ${\bf 7.} \ \ Environmental Condition$
- 8. MeasurementUnits9. Uncertainty
- 10. Statements


- CalibrationCertificate → DeviceUnderTest (Association): A certificate is issued for the device under test.
- CalibrationCertificate → CalibrationMetrologists (Association): The metrologists involved in performing, checking, reviewing and signing the certificate.
- CalibrationCertificate → Customer (Association): The certificate is issued to a customer.
- CalibrationProcess → CalibrationMetrologists (Association): A technician is responsible for the process.
- CalibrationProcess → EnvironmentalCondition (Association): The environmental conditions are
- recorded during the calibration.
- CalibrationProcess → TestMeasurement (Aggregation): Multiple test measurements are part of the process.
- DeviceUnderTest → CalibrationProcess (Association): An instrument undergoes several calibration processes over time.





#### **CTBR Form**






Exploring ways to map the service categories to the BIPM SIDF

5th International DCC Conference , 27 February 2025, NPLI



# **CTBR Form**






5th International DCC Conference , 27 February 2025, NPLI

CSIR - NPL

#### 18

# **DCC Form**





5th International DCC Conference , 27 February 2025, NPLI

CSIR - NPL



# **Uploaded excel document**



Sr. No.	Energy applied (J) to UUC	Calculated (J) level	Measured Energy (J) level of UU	nergy correction (J) (Ecalculated-Emeasured	Uncertainty (J)	Coverage Factor (k)
1	10	9.5	9.6	(-) 0.1	± 0.1	2
2	20	19.3	19.4	(-) 0.1	± 0.1	2
3	50	48.2	48.4	(-) 0.2	± 0.2	2
4	100	97.1	97.4	(-) 0.3	± 0.3	2
5	150	146.1	146.4	(-) 0.3	± 0.5	2
6	270	259.5	260.3	(-) 0.8	± 0.9	2
Sr. No.	Discharge Time	Calculated Discharge Time (ms	Measured Discharge Time (ms)	Uncertainty (J)	Coverage Factor (k)	
1	charge Time (Positive Pulse Wi	6.6	6.4	± 0.1	2.17	
2	harge Time (Negative Pulse W	4.4	4.5	± 0.1	2	

1 S	r. No.	Defribillator Analyser Load Resistance (Ω)	Expected Value (Ω)	Measured Value (Ω)	Uncertainty (Ω)	Coverage Factor (k)
2	1	50	50	50.0179	± 0.00041	2.05
3						
4						

1	Sr. No.	Defribillator Analyser Load Resistance (Ω)	Expected Value (Ω)	Measured Value (Ω)	Uncertainty (Ω)	Coverage Factor (k)
2	1	50	50	51.0128	± 0.0005	2.11
2						

5th International DCC Conference , 27 February 2025, NPLI

20

# DCC (Via Excel)





# CSIR - National Physical Laboratory सी एस आई आर - राष्ट्रीय भौतिक प्रयोगशाला



Select Certificate Type Type of Certificate: Select an option Template: Choose File No file chosen Embed data in PDF: Add Graph in PDF:


CSIR - NPL

5th International DCC Conference , 27 February 2025, NPLI



# **Generated PDF A3**






5th International DCC Conference , 27 February 2025, NPLI

CSIR - NPL

22

# **Customer Welcome Page**





Welcome! You are logged in.

5th International DCC Conference , 27 February 2025, NPLI

CSIR - NPL



# **Summary**



An incremental approach for the modular and scalable DCC implementation at NPLI using MVC architecture and object-oriented principles.



Unique certificate identifiers & metadata indexing related to year, date, time, laboratory name and appended with a unique serial number.



Secure yet user-friendly digital access through PDF/A3 format with checksum error integration



measurement data is provided on demand to customers as a csv file embedded in PDF/A3 format.



standardized structure for future expansions.

5th International DCC Conference , 27 February 2025, NPLI

24

#### **DCC Value Drivers**



#### **GENERATION**

- · How easy is it to generate a Calibration Certificate
- · What level of effort is reliant on human to generate the Calibration Certificate

#### **■ SECURITY & TRUST**

- · How secure is the Calibration Certificate once it is generated
- How can the Certificate provide trust in the issuing actor
- · Can this security/trust be revoked

#### TRANSMISSION

- How are the Calibration Certificates transmitted
- Is the process secure and automatic
- · Can the transmission be reversed

#### MACHINE & HUMAN READABLE

- · Can a machine automatically read and access the data in a Calibration Certificate
- · Can a human read and understand the certificate and its implications
- Is the Calibration Certificate data accessible for secondary purposes i.e. analytics and insights

#### ©OPEN & STANDARDS

- · Is the Calibration Certificate technology open and does it prevent vendor/technology lock
- Does it facilitate standards either design or through configuration

#### (X) EASE OF ACCESS

- · Does the Calibration Certificate facilitate access for all users i.e. does not require new software or technology
- · Are the barriers to entry low to ensure adoption
- Are the skills easy to leam/procure

#### LONGEVITY & PRESERVATION

- · Does the creation and output of the Calibration Certificate allow for long term reference and preservation
- · Will the Calibration Certificates be supported and accessible by future technology

#### EXTENSIBILITY

- Does the underlying technology and Calibration Certificate design allow for flexibility to address jurisdiction differences (i.e. internationally compatible) and future industry changes
- Does the underlying design support growing into other certificates (investment reuse)
- Can we utilise middleware to extend the DCC solution to other systems within the organisation

5th International DCC Conference, 27 February 2025, NPLI



# Path Ahead NPL, India



- Recognises the need to evolve in small steps towards delivering digital products to our customers.
- Establish a strong IT infrastructure required for a secured and reliable system for generating electronic calibration certificates.
- Awareness programs for both internal and external stakeholders
- Path ahead for us is to conduct internal trials and evaluation on DCC and to find out the actual needs in industry (e.g. whether a PDF/A-3 report with embedded raw data is sufficient for the near future).
- · Create models for DCC

5th International DCC Conference , 27 February 2025, NPLI

26

## Acknowledgment



- R P Aloysius, NPLI
- Nidhi Singh, NPLI
- Girija Moona, NPLI
- Ved Varun, NPLI
- Paramita Guha, NPLI
- Srinivas Rao, NPLI
- Anshul Varshney, NPLI
- Niharika Bhatia
- Shubham Anand Jaiswal
- Moulik Sharma
- Hridvansh Gautam

- Martin Koval, CMI, Czech Republic,
- Shanna Schönhals, PTB, Germany,
- James Fedchak, NIST, USA,
- David Balslev-Harder, DFM, Denmark,
- Thomas Krah, PTB, Germany,
- Diego Coppa, INTI, Argentina,
- Julian Haller, Sartorius
- Blair Hall, MSL, New Zealand
- Cui Shan & Cui Yuxi , A\*STAR.NMC, Singapore
- Mark Ballico, NMI, Australia
- · Narin & Pariya Thongluang, NIM, Thailand
- Wan-Ho Cho, KRISS, Rep of Korea





Back to "Table of Contents" above

#### P08 DCC Sharing within a Government Agency

Presenting author: Sajiv Perera, Measurement Canada

Email: sajiv.perera@ised-isde.gc.ca

Additional authors: John Scott <u>john.scott2@ised-isde.gc.ca</u>, Daljit Dhaliwal daljit.dhaliwal@ised-isde.gc.ca, Ayad Harake ayad.harake@ised-isde.gc.ca

#### **ABSTRACT**

Measurement Canada [1] is the legal metrology agency responsible for trade measurement within the Government of Canada. Calibration of our primary standards is traced through and performed by the National Research Council of Canada, [2] and in some cases other designated laboratories as well. We provide calibration and certification services that are used by Measurement Canada's post market sector and which in turn provides services to Measurement Canada Authorized Service Providers (ASPs) [3].

In this presentation we will briefly show how our agency's size and narrow scope allows for rapid implementation and is ideal for early testing of machine-readable information and allows us to be agile in changes to the DCC standard. We will demonstrate our rationale for starting our digital certificate project with electricity measuring standards, as well as the wider context in how that information can be used in subsequent certifications. We will demonstrate how we transmit this information from our labs to our customers while maintaining traceability. We will highlight the potential increase in efficiency and improved quality which can be achieved by incorporating machine readable certification data within our department.

Finally, we will outline our goals for cooperating with other government organizations to receive and distribute FAIR [4] measurement data. This will demonstrate a use case for DCC that does not simply encompass a single agency or process, but how the DCC may be used across sectors and organizations.



#### References:

[1] Measurement Canada, Government of Canada,

URL: <a href="https://ised-isde.canada.ca/site/measurement-canada/">https://ised-isde.canada.ca/site/measurement-canada/</a>


- [2] National Research Council, Government of Canada, URL: https://nrc.canada.ca/en
- [3] "Authorized service providers", Measurement Canada, Government of Canada, URL: <a href="https://ised-isde.canada.ca/site/measurement-canada/en/authorized-service-providers">https://ised-isde.canada.ca/site/measurement-canada/en/authorized-service-providers</a>
- [4] "Roadmap for Open Science", Government of Canada, <a href="https://science.gc.ca/site/science/en/office-chief-science-advisor/open-science/roadmap-open-science">https://science.gc.ca/site/science/en/office-chief-science-advisor/open-science/roadmap-open-science</a>

Back to "Table of Contents" above

### Presentation of Sajiv Perera







# **Organisational Structure**

**Automated DCC Generation** 

Roadmap

Minister of Innovation, Science and Economic Development

~ 4000 🕮



Ministry of Innovation, Science and Economic Development

Small business and marketplace services branch





# **Traceability**



Primary standard for electrical power traceable to NRC


Incoming calibration information from the NRC still manual entry in our systems.

Our certificates generated by directly reading this information from database.

4

# **Modernisation**

- Calibration results still transmitted in a static format, not machine-readable.
- Opportunity for improvement through digital transformation of this traceability chain.





# **Project Scope**



Narrow scope of work allows for quick implementation and acceptance.



Large impact due to large volume of data transmitted.



Beneficial to increase efficiency of existing process.

6

# Calibration of an Electrical Standard

- Chosen due to:
  - Certification procedure already largely automated
  - · Large amount of repetitive data created
    - Low cost to calculate error values
  - Large set of correction values
    - Machine readable data saves input time and reduces transcription errors



version='1.0' encoding='utf-8'

http://www.w3.org/2001/XMLSch https://ptb.de/dcc https://ptb.de/ 3.3.0

7₹ xml

✓ [E] dcc:digitalCalibrationCertificate

⑧ xmlns:dcc

⑤ xmlns:si

> @ dcc:dccSoftware

@ xmlnsxsi




# **Automated DCC Generation**



> @ decorebata
> @ decorebata
> @ decorebata
> @ decorems

| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decorems
| decore

Roadmap





# Roadmap



- Further opportunities
  - DCC for Gas Mixtures
  - Acceptance of external DCC
- Timeline
  - Lab calibration-to-DCC process complete by spring 2025







Back to "Table of Contents" above

Day 3 (2025-02-27 Thursday)

Parallel Session: Schema & Semantics

(chair: Robert Hanisch)

O26 Augmenting the DCC Framework with the DSI Ontology: Enabling Semantic Web Integration in Calibration Data Management

Presenting author: Moritz Jordan Email: <a href="mailto:moritz.jordan@ptb.de">moritz.jordan@ptb.de</a>

Additional authors: Giacomo Lanza Giacomo.Lanza@ptb.de, Daniel Hutzschenreuter

Daniel.Hutzschenreuter@ptb.de

#### **ABSTRACT**

The Digital System of Units (DSI) Ontology represents a crucial step toward enabling seamless interoperability, reasoning, and traceability in metrology within the framework of Digital Calibration Certificates (DCCs). This presentation highlights the role of the DSI Ontology as a bridge between related standards and schemas, including the SI Reference Point, the DSI XML Schema, and the DCC Schema/Ontology. By aligning these frameworks, the ontology ensures consistency and coherence in the representation and exchange of calibration data.

Through its modular structure, the DSI Ontology formalizes essential concepts such as quantities, units, and measurement uncertainties, enabling not only smarter querying but also logical reasoning over calibration data. This reasoning capability facilitates the inference of implicit knowledge, validation of data consistency, and automation of workflows within metrology knowledge graphs. A core focus will be on showcasing its connections to related



ontologies and standards via an intuitive diagram, as well as demonstrating its use in realworld calibration scenarios.

The presentation will also delve into the ontology's adherence to international standards, its validation process, and tools supporting its development and application. Finally, we explore future directions for expanding the ontology to further facilitate automation, reasoning, and precision in calibration processes, driving innovation in the metrology domain.

# Presentation of Moritz Jordan



# The D-SI Ontology

**Enabling Semantic Web Integration into the DCC** 

Moritz Jordan, Giacomo Lanza, Daniel Hutzschenreuter, PTB



## Motivation for the D-SI Ontology



Why do we need a digital representation of SI?

- Make it more FAIR
- Increase interoperability by linking to other ontologies
- Format-agnostic semantic model of the Digital SI
  - -> More applications possible (API, Data Repositories...)



D-SI

D-SI



SI Brochure

D-SI XML Schema D-SI Update 2.2.0

**D-SI Ontology** 

05.05.2025

Moritz Jordan et al., PTB



## The Role of Ontologies in Metrology



- Benefits of using ontologies: interoperability, formal semantics, automated reasoning
- Reasoning can detect inconsistencies,
- BIPM SI Reference Point -> metrological perspective, not complete
- QUDT -> Community Project, application perspective





Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

Moritz Jordan et al., PTE

# What the D-SI Ontology Covers



- Quantity values (Real, Constant, Complex, etc...)
- Measurement uncertainty (Standard, Expanded, Multivariate...)
- Measurement units (D-SI, BIPM SI Ref. Framework)
- Quantity kind (BIPM SI Ref. Framework, QUDT)
- Fundamental constants as instances of Constant (i.e. Planck)

Physikalisch-Technische Bundesanstalt 
Braunschweig and Berlin


National Metrology Institute

Moritz Jordan et al., PTE




# **Key Components: QuantityValue**





# Key Components: MeasurementUncertainty PIB








# Alignment with BIPM SI and QUDT






Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

Moritz Jordan et al., F





# Validation & Challenges




- Python Mapping for converting XML D-SI data to DSIO instances
- Testing conversion with pytest for each instance of QuantityValue
- Reasoning with RDFS, OWLRL and HermiT
- Validating data instances with SHACL
- Limitations of OWL to map XML Simple Type with restrictions on xsd:double

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin 05 05 2025 National Metrology Institute

Moritz Jordan et al., PTE

**DSI Ontology Applications** 







## **Future Work & Roadmap**



- Expanding ontology coverage (List, RealXMLList, Hybrid..)
- Links to VIM, Links to chapters in GUM
- Publish D-SI Ontology (Documentation, Use Cases, Tools XML2DSIO)

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

11

#### **About**

Physikalisch-Technische Bundesanstalt Braunschweig and Berlin
Bundesallee 100
38116 Braunschweig
Moritz Jordan
Telefon: 0531 592-1245
E-Mail: moritz.jordan@ptb.de
www.ptb.de

Back to "Table of Contents" above

# 027 Changes in DCC Schema V.3.3.0

Presenting author: Muhammed-Ali Demir

Email: muhammed.demir@ptb.de

Additional authors: Et. al.

#### **ABSTRACT**

In this presentation we will show the latest changes in the DCC schema version V.3.3.0 with examples. The changes consist of changes to the documentation. Changes due to update of the referenced SI schema, consistency changes and additions to the schema.




#### Presentation of Muhammed-Ali Demir



# Changes in DCC Schema V.3.3.0

#### 5<sup>th</sup> DCC Conference

Muhammed-Ali Demir et al., WG1.24



#### New DCC schema V.3.3.0



- Release date: 29<sup>th</sup> of October 2024
- Source of changes:
  - DKD cross sectional group
  - DCC Core Team
- Let's have a closer look...

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# Changes in DCC Schema V.3.3.0



- Schema documentation
- D-SI Schema Update
- Consistency
- Additions


Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin 2025-02-27

National Metrology Institute

# **Schema documentation**



Updated xs:documentation-Tags



Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



#### D-SI Schema 1/5



Update from D-SI version 2.1.0 to 2.2.1

```
<xs:import namespace="https://ptb.de/si"
- schemaLocation="https://ptb.de/si/v2.1.0/SI_Format.xsd"/>
+ schemaLocation="https://ptb.de/si/v2.2.1/SI_Format.xsd"/>
```

- primitiveQuantityType/ quantityType now contain also: si:complexListXMLList
- Example: dcc:quantity

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

5

National Metrology Institute

# D-SI Schema 2/5



```
<xs:complexType name="quantityType">
 <xs:element name="name" type="dcc:textType" min0ccurs="0"/>
 <xs:element name="description" type="dcc:richContentType" min0ccurs="0"/>
 <xs:choice>
 <xs:element name="noQuantity" type="dcc:richContentType"/>
 <xs:element name="charsXMLList" type="dcc:charsXMLListType"/>
 <xs:element ref="si:real"/>
 <xs:element ref="si:list"/>
 <xs:element ref="si:hybrid"/>
 <xs:element ref="si:complex"/>
 <xs:element ref="si:constant"/>
 <xs:element ref="si:realListXMLList"/>
 <xs:element ref="si:complexListXMLList"/>
 </xs:choice>
<xs:element name="relativeUncertainty" type="dcc:relativeUncertaintyType" minOccurs="8" />
 <xs:element name="usedMethods" type="dcc:usedMethodListType" minOccurs="0"/>
<xs:element name="usedSoftware" type="dcc:softwareListType" minOccurs="0"/>
 <xs:element name="measuringEquipments" type="dcc:measuringEquipmentListType" minOccurs="8"/>
<xs:element name="influenceConditions" type="dcc:influenceConditionListType" minOccurs="8"/>
 <xs:element name="measurementMetaData" type="dcc:measurementMetaDataListType" minOccurs="0"/>
 </xs:sequence>
 <xs:attribute name="id" type="xs:ID" use="optional"/>
<xs:attribute name="refId" type="xs:IDREFS" use="optional"/>
 <xs:attribute name="refType" type="dcc:refTypesType" use="optional"/>
</xs:complexType>
```

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



#### D-SI Schema 3/5



```
- cartesian complex quantity -->
<si:complex>
 <si:label>amplifier signal</si:label>
 <si:valueReal>-10.310</si:valueReal>
 <si:valueImag>5.010</si:valueImag>
<si:unit>\volt</si:unit>
 <si:dateTime>2024-12-03T13:25:00Z</si:dat</pre>
 (si:complexListXMLList>
 <si:labelXMLList>amplifier signal</si:labelXMLList>
<!-- polar coordinate complex quantity -->
 <si:valueRealXMLList>-10.310 ...</si:valueRealXMLList>
<si:valueImagXMLList>5.010 ...</si:valueImagXMLList>
<si:unitXMLList>\volt ...</si:unitXMLList>
<si:complex>
 <si:label>amplifier signal</si:label>
 <si:valueMagnitude>11.463</si:valueMagnit
 <si:dateTimeXMLList>2024-12-03T13:25:00Z ...</si:dateTimeXMLList>
 <si:valuePhase>2.689</si:valuePhase>
 /si:complexListXMLList>
 <si:unit>\volt</si:unit>
 <si:unitPhase>\radian</si:unitPhase>
 polar coordinate complex list quantity -->
 <si:dateTime>2024-12-03T13:25:00Z</si:dat

⟨si:complexListXMLList>
</si:complex>
 <si:labelXMList>amplifier signal</si:labelXMLList>
 <si:valueMagnitudeXMList>11.463 ...</si:valueMagnitudeXMLList>
<si:valuePhaseXMList>2.689 ...</si:valuePhaseXMLList>
<si:unitXMList>\volt ...</si:unitXMLList>
 <si:unitPhaseXMList>\radian ...</si:unitPhaseXMLList>
 <si:dateTimeXMList>2024-12-03T13:25:00Z ...</si:dateTimeXMList>
 si:complexListXMLList
```

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

#### D-SI Schema 4/5



Update from D-SI version 2.1.0 to 2.2.1

```
<xs:import namespace="https://ptb.de/si"
- schemaLocation="https://ptb.de/si/v2.1.0/SI_Format.xsd"/>
+ schemaLocation="https://ptb.de/si/v2.2.1/SI_Format.xsd"/>
```

 Deprecated uncertrainty types -> use measurementUncertaintyUnivariate instead

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



#### D-SI Schema 5/5



```
<si:realListXMLList>
 <si:valueXMLList>8.072 0.089 0.107 -0.009 -0.084</si:valueXMLList>
 <si:unitXMLList>\kelvin</si:unitXMLList>
 expandedUncXMLList>
 ndedUncXMLList>
</si:realListXMLList>
<si:realListXMLList>
 <si:valueXMLList>0.072 0.089 0.107 -0.009 -0.084</si:valueXMLList>
 Possible children of
 <si:unitXMLList>\kelvin</si:unitXMLList>
<si:measurementUncertaintyUnivariateXMLList;</pre>
 si:measurementUncertaintyUnivariate
 <si:expandedMUXMLList>
 <si:valueExpandedMUXMLList>0.061</si:valueExpandedMUXMLList>
 standardMU
 <si:coverageFactorXMLList>2</si:coverageFactorXMLList>
<si:coverageProbabilityXMLList>0.95</si:coverageProbabilityXMLList>
 expandedMU
 <si:distributionXMLList>normal</si:distributionXMLList>
 coverageIntervalMU
 </si:expandedMUXMLList>
 mentUncertaintyUnivariateXMLList>
</si:realListXMLList>
```

# **Consistency 1/4**

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



National Metrology Institute

- Added dcc:link everywhere parallel to dcc:reference, in order to make it possible to specify links specifically e.g., in equipmentClassType
- Updated type from dcc:link to xs:anyURI instead of string (machine interpretability of the schema)

```
- <xs:element name="link" type="dcc:notEmptyStringType"/>
+ <xs:element name="link" type="xs:anyURI"/>

<dc:link>https://forcetechnology.com/en/events/2024/</dcc:link>
```

Physikalisch-Technische Bundesanstalt 
Braunschweig and Berlin



# **Consistency 2/4**



 contactType now has a refType (compare to contactNotStrictType)

```
<dcc:contact refType="...">
<xs:complexType name="contactType">
 <dcc:name
 <xs:sequence>
 <dcc:content>Muhammed-Ali Demir</dcc:content>
 <xs:element name="name" type="dcc:textType"/>
 </dcc:name>
 <xs:element name="eMail" type="dcc:notEmptyStringType" minOccur</pre>
 <dcc:eMail>muhammed.demir@ptb.de</dcc:eMail>
 <xs:element name="phone" type="dcc:notEmptyStringType" minOccur</pre>
 <dcc:location>
 <xs:element name="fax" type="dcc:notEmptyStringType" minOccurs:</pre>
 <dcc:city>Braunschweig</dcc:city>
 <xs:element name="link" type="xs:anyURI" min0ccurs="0"/>
 </dcc:location>
 <xs:element name="location" type="dcc:locationType"/>

<
 </xs:sequence>
 <xs:attribute name="id" type="xs:IO" use="optional"/>
 <xs:attribute name="refId" type="xs:IDREFS" use="optional"/>
 <xs:attribute name="refType" type="dcc:refTypesType" use="optional"/>
</xs:complexType>
```

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

# **PIB**

# Consistency 3/4

 Added attribute "refld" everywhere, where attribute "id" was defined e.g., in softwareType

```
<xs:complexType name="softwareType">
 <dcc:software refId="xyz"
 <xs:annotation>
 <xs:documentation>
 <dcc:name
 Information about a software including its name, version and a desc
 <dcc:content lang="en">Firmware abc</dcc:content>
 Information about the software, including its name, version, type,
 </dcc:name>
 <dcc:release>1.0</dcc:release>
 </xs:annotation>
 </dcc:software>
 <xs:sequence>
 <xs:attribute name="id" type="xs:ID" use="optional"/>
 <xs:attribute name="refId" type="xs:IDREFS" use="optional"/>
 <xs:attribute name="refType" type="dcc:refTypesType" use="optional"/>
 </xs:complexType>
```

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# **Consistency 4/4**



Added id & refld to itemListType

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin


13

National Metrology Institute

#### Additions 1/4



■ Items can now be nested to any depth, examples: PXI system with slots/sensors, MEMS — Inertial measurement unit



Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



#### Additions 2/4



 measuringEquipmentType now has installedSoftware to specify software for dcc:measuringEquipment



Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

15

#### Additions 3/4



It is now possible to place quantities in dcc:usedMethod so that quantities can now be specified in a machine interpretable manner


Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



#### Additions 4/4



 Now usedMethods, usedSoftware, measuringEquipments, influenceConditions and measurementMetaData can be placed in dcc:measurementResults



Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

17

National Metrology Institute

#### Information sources



- D-SI V.2.2.1: <u>link</u>
- D-SI example for si:complex: <u>link</u>
- DCC release page: <a href="https://gitlab.com/ptb/dcc/xsd-dcc/-/releases/v3.3.0">https://gitlab.com/ptb/dcc/xsd-dcc/-/releases/v3.3.0</a>
- DCC schema link: <a href="https://www.ptb.de/dcc/v3.3.0/dcc.xsd">https://www.ptb.de/dcc/v3.3.0/dcc.xsd</a>
- Documentation: <u>https://www.ptb.de/dcc/v3.3.0/autogenerated-docs/Doku</u> Oxygen 3.3.0.html
- Some XML fragments from the BAM temperature GP (Michael Melzer et al.) <u>link</u>

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



#### **Contributions**



DKD Cross sectional group & Muhammed-Ali Demir, Lutz Doering, Benjamin Gloger, Siegfried Hackel, Frank Härtig, Wiehke Heeren Daniel Hutzschenreuter, Justin Jagieniak, Moritz Jordan, Hans Koch, Thomas Krah, Shan Lin, Jan Loewe, Shanna Schönhals, Thorsten Schrader, Gamze Söylev-Öktem, Thomas Wiedenhöfer,

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

2025-02-27

National Metrology Institute

19

#### **About**



#### 028 An In-Depth Look at the DCX Schema

Presenting author: David Balslev-Harder; DFM - Danish National Metrology Institute

Email: dbh@dfm.dk

Additional authors: Søren Kynde srk@dfm.dk

#### **ABSTRACT**

The advent of Digital Calibration Certificates (DCC) marks a significant advancement in the field of metrology, offering enhanced accuracy, traceability, and interoperability. This presentation delves into the DCX Schema, a pivotal component in the implementation of DCCs. The DCX schema has been developed by DFM, with input from members of TC-IM



1448 WG1&2, aiming to investigate the use of reference pointers within the XML rather than the tree-structure emphasis applied in the PTB DCC-schema.

We begin with an overview of the DCX Schema's overall XSD structure, highlighting its robust framework designed to ensure data integrity and consistency. The schema's alignment with ISO 80000 standards is discussed, emphasizing its role in maintaining uniformity in the representation of physical quantities. Additionally, the integration of the BIPM-SI-Reference point within the schema is explored, underscoring its importance in providing a universally accepted reference for measurements.

A critical aspect of the DCX Schema is its ability to restrict entry-data, thereby preventing erroneous or non-compliant data entries. This presentation will outline the specific restrictions imposed by the schema, providing a comprehensive list of relevant constraints that ensure the accuracy and reliability of calibration data.

Furthermore, the schema's design includes human-readable fields on all elements, facilitating ease of interpretation and verification by end-users. This feature enhances the usability of DCCs, making them accessible to a broader audience, including those without extensive technical expertise.

In conclusion, the DCX Schema represents a significant step forward in the digitalization of calibration certificates, offering a standardized, reliable, and user-friendly solution for the metrology community.

#### **Outline of presentation**

- DCX Schema Overview
- o Incorporation of VIM, ISO 80000 and BIPM-SI-Reference point
- Data interpretation integrity using xsd restrictions
- Human-Readable Fields
- Solution available on github.com/TC-IM-1448/DCC-Tables
- See the implementation examples and user-interfaces in other presentations.

DFM's work is supported by funds from The Danish Agency for Institutions and Educational Grants.



# Presentation of David Balslev-Harder

#### An In-Depth Look at the DCX Schema

5<sup>th</sup> international DCC conference,

February 27th 2025

By David Balslev-Harder & Søren R. Kynde





DFM's incentive for DCC:


To support creation of FAIR DCC formats that are globally recognized/standardized, in order to support quality and efficiency in production companies, through automation.

Develop simple user interface software tool that can accept and modify DCC, DCR and Templates regardless of origin.


DFM receives funding by: Danish Agency for Institutions and Educational Grants


# Schema updates since last year

- 1. Introduced dcx: as namespace
- 2. DCR and Templates incorporated in the same schema.
- 3. Introduced human readable fields for all elements.
- 4. Simplified layout of administrative data
- 5. New default system for Quantities and Units based on ISO 80000 and BIPM SI-Digital-Framework unit representation introduced in schema including:
  - Incorporated in the data result tables.
  - 2. RegEx for validating BIPM-SI unit expression.
  - 3. Section for defining other quantities and units.
- Restriction on data type in columns depending on dataCategory introduced
- Simplified and improved query expressions for extracting measurement data and meta data.











# Administrative data • Administrative data • Statements • Description of equipment • Specification of equipment settings • List of uniquere measurement configurations • Results • Each result refers to a unique measurement configuration • Annexes • Definitions of non SI-units • List of embedded files • List of codeable elements dex.xsd (schema) \*\*xs:complexType name="digitalCalibrationExchangeType"> \*\*xs:complexType name="title" type="dex:transactionContentFileIdType"/> \*\*xs:element name="subtitle" type="dex:transactionContentFileIdType" minoccurs="0"/> \*\*xs:element name="subtitle" type="dex:transactionContentFileIdType" minoccurs="0"/> \*\*xs:element name="administrativeData" type="dex:administrativeDataType"/> \*\*xs:element name="administrativeDataType"/> \*\*xs:element name="calibrationExctangeType" minoccurs="0"/> \*\*xs:element name="administrativeDataType"/> \*\*xs:element name="calibrationExctangeType" minoccurs="0"/> \*\*xs:element name="calibrationExctangeType" minoccurs="0"/> \*\*xs:element name="calibrationExctangeType" minoccurs="0"/> \*\*xs:element name="calibrationExctangeType"/> \*\*xs:element name="calibrationExctangeType"//> \*\*xs:element name="calibrationExctangeType"//> \*\*xs:element name="calibrationExctangeType"// \*\*xs:element name="calibrationExctangeType"// \*\*xs:element name="calibrationExctangeType"// \*\*xs:element name="calibrationExctangeType"// \*\*xs:element name="calibrationExctangeType"// \*\*xs:element name="calibration



# mini\_DCC.xml (Document) c?ml version='1.0' encoding='utf-8'?> cdx:xdigitalCalibrationExchange xmlns:ddcx="https://dfm.dk" \_> cdx:xdigitalCalibrationExchange xmlns:ddcx="https://dcx| cdx:xdigitalCalibrationExchange xmlns:ddcx=https://dcx| cdx:xdigitalCalibrationExchange xmlns:ddcx=https://dcx| cdx:xdigitalCalibrationExchange xmlns:ddcx=https://dcx| cdx:xdigitalCalibrationExchange xmlns:ddcx=https://dcx| cdx:xdigitalCali

# Using SI-Digital-Framework repr. in schemas

- The SI-Digital-Framework unit expressions is well suited for incorporation in schemas (e.g. xsd) because it can be expressed as regular expressions:
- Suggestion: Standardize a regular expression for use in schemas.

	Unit expressions						
μ τς α • • •	Aparsing tool to provide the PID and semantic model for compound units.						
р с м	Non-ASCII characters use of or representation of perfix micro and units degree Celsius, ohm, degree, arcminute, assection.						
kg.mm2.ns-2 Parse Expression	μ *C Ω * * *						
	kg.mm2.ns-2	Parse Expression					



# Using SI-Digital-Framework repr. in schemas

- The SI-Digital-Framework unit expressions is well suited for incorporation in schemas (e.g. xsd) because it can be expressed as regular expressions:
- Suggestion: Standardize a regular expression for use in schemas.



dbh@dfm.dl

#### Quantities - ISO 80000

- A full discription of the data in the column requires definition of the quantity beeing measured and the unit used.
- The values of the quantity attribute relates to a single entry of the standards in the ISO 80000 series.

Item No.	Quantity			Unit
	Name Symbol Definition			
4-10	impulse		vector (ISO 80000-2) quantity describing the effect of force acting during a time interval:	N s kg m s <sup>-1</sup>
			$I = \int_{1}^{t_{2}} F dt$ $t_{1}$ where $F$ is force (item 4-9.1), $t$ is time (ISO 80000-3) and $[t_{1}, t_{2}]$ is considered time interval	
4-11	angular momentum	L	vector (ISO.80000-2) quantity described by the vector product: $L=\times p$ where $r$ is position vector (ISO.80000-3) with respect to the axis of rotation and $p$ is momentum (item 4-8)	kg m <sup>2</sup> s <sup>-1</sup>
4-12.1	moment of force	М	vector (ISO.80000-2) quantity described by the vector product: $M = r \times F$ where $r$ is position vector (ISO.80000-3) with respect to the axis of rotation and $F$ is force (item 4-9.1)	N m kg m² s-2
4-12.2	torque	T, MQ	quantity described by the scalar product:	Nm

dbh@dfm.dk



# Initial attempt to digitize ISO 80000

563 quantities from the 9 ISO 80000 norms mentioned earlier, was transferred to str format:

#### "4-31|volume flow rate|m3.s-1"

Unit was converted to BIPM SIdigital-Framework representation.

The full list can be included within schema's e.g. xsd.

dbh@dfm.dl

```
cxisinpletype name—quantity)ppe">
cxismoniation

cxisoniamintation

cxisoniamintation

the quantity pape is restricting the value of the quantity attribute of the columntype.

for meta data columns the attribute is used to describe the data.

for meta data columns the attribute is used to describe the data.

for meta data columns the attribute is used to match the data column(s) that is excited with the meta data.

for quantities defined is restricted to strings that represent entries from the ISO 80000 series.

Titen number] | [quantity nume] | Imati'

frample:

"pulled in the string contains information on Iten number, quantity nume and unit is as follows

"[Iten number] | [quantity nume] | Imati'

frample:

"pulled in the string contains information on Iten number, quantity nume and unit is as follows

"[Iten number] | [quantity nume] | Imati'

frample:

"pulled in the string contains information on Iten number, quantity nume and unit is as follows

"[Iten number] | [quantity nume] | Imati'

frample:

"pulled in the string contains information on Iten number, quantity nume and unit is as follows

"[Iten number] | [quantity nume] | Imati'

frample:

"pulled in the string contains information on Iten number, quantity nume and unit is as follows

"[Iten number] | [quantity nume] | Imati'

Iten number] | [quantity nu
```

# 

Columns: quantities and unit units



```
//dex:calibrationResult[@measurementConfigRe!="ms1"]/
 [@scope="reference"][@quantity="3-4|volume|m3"][@unit="cm3"]/dex:value/*[@idx="1"]/text() = 2516

dcx.xsd (schema)

cxs:complex/type name="columnType">
cxs:clement name="columnType">
cdex:column scope=indication* dataCategorylef="-" quantity="4-1|mass|kg" unit="kg">
cdex:column scope=indication* dataCategoryl
```



# Path forward

- With PTB investigate how the DCX requirements can be merged/harmonized with DCC, TemaTres
- 2. In TC-IM 1448: Harmonize semantics and structures, international perspectiv
- 3. Standardized regular expression for BIPM Digital SI unit expression
- 4. Standardized schema for ISO 80000 through ISO TC12.
- 5. Incorporate NCSLI metrology Taxonomy into settingCodes.
- 6. Pilot study and Full process cost analysis (Template, DCR, DCC, extraction)
  - 1. Use software tools presented yesterday.
  - 2. Cost of deployment, maintenance, versatility, for both clients and Service Providers.

dbh@dfm.dl

DFM's incentive for DCC:

To support creation of FAIR DCC formats that are globally recognized/standardized, in order to support quality and efficiency in production companies, through automation.

Develop simple user interface software tool that can accept and modify DCC, DCR and Templates regardless of origin.

**Danish collaboration partners:** 

DANDIAG: Tove Søgaard, Rene, Charlotte Rugaard

Novo Nordisk: Heidi Foldal, Jan Laursen, Johan Schroll-Flaischer

FORCE Technology: Ahmed Khan Leghari, Andreas Lyndrup Jensen, ...

Teknologisk Institut: Jonas Vind, Peter Friis Østergaard, Jan Nielsen, ...

DANAK: Erik Øhlenschlæger

Contact: David Balslev-Harder (dbh@dfm.dk)

PTE

Members in TC-IM 1448 WG1 and WG2:

DFM, Denmark:
BEV, Austria:
INTI, Argentina:
IPQ, Portugal:
NPL, UK:
NPL, UK:
NPL, UK:
David Balslev-Harder (Chair)
Zoltan Zelenka
Diego Nahuel Coppa,
Carlos Pires
NPL, UK:
Michael Chrubasik, Ian Smith

IPLI, India: Anjali Sharma

PTB, Germany: Shanna Schönhalls, Daniel Hutzschenreuter, Gamze Soeylev-oektem, Clifford Brown,

Wiebke Heeren.

SMD, Belgium: Quentin Baire, TI, Denmark: Jonas Vind TUBITAK, Turkey: Erkan Danaci

**Collegues from PTB:** 

Shanna Schönhalls, Thomas Krah Daniel Hutzschenreuter Muhammed-Ali Demir Gamze Soeylev-oektem Benedikt Seeger Clifford Brown Wiebke Heeren Siegfried Hackel Frank Härtig ...

Back to "Table of Contents" above

# 029 Compatible sub schemata of the DCC Schema

Presenting author: Muhammed-Ali Demir

Email: muhammed.demir@ptb.de

Additional authors: Et. al.

### Abstract:

A need has arisen for a smaller, stricter version of the DCC schema with reduced flexibility. Rather than creating an entirely incompatible schema, it is possible to develop a subschema of the DCC schema that is both smaller and more restrictive. If certain rules are followed during



the subschema's creation, DCCs generated using it will remain valid under the main DCC schema.

#### Presentation of Muhammed-Ali Demir



# Compatible sub schemata of the DCC Schema

5<sup>th</sup> DCC Conference

Muhammed-Ali Demir et al., WG1.24



### Reasons for sub schemas of the DCC



- Questions:
  - The main DCC schema is flexible -> complexity
  - Flexibility may not be what small laboratories need -> small resources
- Needs:
  - Simpler & less flexible schema may be requested
- Problems:
  - This may lead to different incompatible schemas
  - This incompatibilities already have happened before ...

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# The consequences of lacking uniformity





Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin 2025 02 27

National Metrology Institute

# **Solution**



- There is a solution to have a simpler sub schemata, derived from the DCC (main) schema
- When specific rules are followed, the DCC XML files generated from the sub-schema are automatically valid against the DCC main schema.
- DKD expert reports should also be considered for the generation of sub schemas e.g., see the presentation of: Itzel Dominguez, Carlos Galvan, Siegfried Hackel et al. from our colleagues of CENAM with the title of "Generating Mass DCC based on DKD expert report using a spreadsheet" from Tuesday at this conference

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# Rules for generating the sub schema 1/9



- Optional schema elements can be made mandatory\*
- Example: Manufacturer has to be specified in items

\* = not relevant for minimal schemas

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin 2025-02-27

National Metrology Institute

# Rules for generating the sub schema 2/9



- Optional schema elements can be deleted\*
- Example: Item elements should not have any sub items

\* = deleting some optional elements can make references with id/refld necessary

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# Rules for generating the sub schema 3/9



- Choices can be simplified
- Example: We only want to keep si:real and si:realListXMLList in dcc:quantity

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

# Rules for generating the sub schema 4/9



- Optional attributes can be made mandatory \*
- Example: every item has to have an id

\* = not relevant for minimal schemas

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# Rules for generating the sub schema 5/9



- Optional attributes can be deleted:
- Example: item elements should have no refld

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

## Rules for generating the sub schema 6/9



- Maximal cardinality of elements can be limited if they are unbounded
- Example: at max. 2 dcc:equipmentClass elements in dcc:item

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# Rules for generating the sub schema 7/9



- Minimal cardinality of elements can be set to 1 (or higher if maximal cardinality is unbounded)\*
- Example: at least one equipmentClass element has to be sepecified in itemType

\* = not relevant for minimal schemata

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

National Metrology Institute

# Rules for generating the sub schema 8/9



Don't delete mandatory elements from the main schema



Marie-Lan Nguyen, Public Domair

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin



# Rules for generating the sub schema 9/9



Don't introduce new elements, that don't exist in the main schema

Physikalisch-Technische Bundesanstalt ■ Braunschweig and Berlin

Marie-Lan Nguyen, Public Domain

National Metrology Institute

#### Information sources



 Itzel Dominguez et al. - Generating Mass DCC based on DKD expert report using a spreadsheet



#### **About**



Back to "Table of Contents" above

Day 3 (2025-02-27 Thursday)
Parallel Session "Key Aspects of DCC II"
(Chair: Mark Kuster)

# 030 Calibration and metrological traceability

Presenting author: Blair Hall; Measurement Standards Laboratory of New Zealand

Email: blair.hall@measurement.govt.nz

Additional authors: Christian Mueller-Schoell christian.mueller-schoell@mt.com, Mettler-

Toledo Int. Inc., Switzerland

### Abstract:

In 1980, Belanger considered four alternative explanations of the term metrological traceability [1]. In 1982, Nicholas and White, in their monograph Traceable Temperatures, proposed a variation of one of Belanger's suggestions [2]. Then the first edition of the Vocabulary of Metrology (VIM), published in 1984, included an entry for traceability based on another of Belanger's suggestions [3], Since then, subsequent editions of the VIM have elaborated on traceability. The current edition not only refers to three specialized terms—measurement result, calibration, and measurement uncertainty—but also includes eight explanatory notes. The VIM now defines the terms metrological traceability chain and metrological traceability to a measurement unit [4].

The current VIM definition is:

- o metrological traceability
- property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty



This highlights calibration and uncertainty; however, these concepts are absent in some of the earlier definitions. Are these elements essential?

Metrological traceability provides reliable information about the accuracy of a measurement, and we contend that its true value is realized when decisions are informed by measurements of real-world quantities. Calibration, as a metrological process, establishes some of the links in the traceability chain, but it is not sufficient on its own to complete the chain-measurement is also required.

This talk will explore metrological traceability through the lens of three key questions: why is traceability needed, what does traceability deliver, and how is traceability achieved? From this perspective, we frame the role of calibration - and, by extension, calibration reports - in the broader context of metrological traceability.

#### References:

- [1] Belanger, B. C., Traceability: an evolving concept, ASTM Standardization News, 8(1) pp 22—28,1980.
- [2] Nicholas, J. V, White, D. R., Traceable temperatures: an introductory guide to temperature measurement and calibration, 1982. (Now published by John Wiley & Sons)
- [3] BIPM, IEC, ISO, OIML, Vocabulary of metrology, Part 1, basic and general terms (international), 1984.
- [4] BIPM, IEC, IFCC, IUPAC, IUPAP, ISO, OIML, International Vocabulary of Metrology Basic and general concepts and associated terms, 3"^ ed., 2012.

#### Presentation of #030 BlairHall





#### Why do we need metrological traceability?

- Only metrologically traceable measurements can be relied on when decisions depend on realworld quantities.
- Traceable measurement results are accompanied by reliable information about their accuracy (uncertainty).
- The comparability of traceable measurement results—whether with other traceable results or with nominal quantity values—is fundamentally important.



Image: de.wikipedia.org

#### What does metrological traceability deliver?

$$Y = y - E_y$$

The measured value y is an estimate (an approximation) of Y—the measurand.

There is a residual measurement error,  $E_y$ , which is not known, so we don't know the measurand either.

A traceable measurement **result** reports y and a probabilistic description of  $E_y$  in the form of an uncertainty statement.

This reflects the **doubt** remaining about the value of the measurand **after** a measurement.





# Measuring is a dynamic process, but there is a definite outcome

The measured value, the measurement error, and the measurand are all considered **definite** values

However, the effects of influence-quantity errors accumulate, so the final error and measurand are **not known**.

The final result is determined by a 'snapshot' of definite values that arose during a particular measurement.

Known knowns, known unknowns, unknown knowns, and unknown unknowns.



Image: de.wikipedia.org

#### How does metrological traceability work?

A traceable measurement is carried out in a sequence of measuring stages.


All stages contribute to the final result—their errors contribute to the residual measurement error.

A **trace** of influence-quantity effects is used to **assess** the accuracy of a measurement result.

The trace begins with the realisation of primary standards and ends when a decision is made.

Reporting between stages forges the links in traceability chain.

Calibration reports have this function.





#### Why is calibration needed?

Measuring systems respond to a quantity of interest, but the response is never perfect.

Calibration characterises a system, so it can be used to estimate measurands

#### What does calibration deliver?

Calibration measures certain system parameters and characterises the system's response to the quantity of interest.

This provides information that is **necessary** for the system to be used to estimate an unknown value of a quantity of interest.



Image: de.wikipedia.org

#### How does calibration work?

- 1. Establish a relationship between reference standards for the quantity of interest (with nominal values) and (raw) system indications.
- Model the system, accounting for bias and response unpredictability, so that a raw response can be used to measure (estimate), with known accuracy, a quantity of interest.

Calibration involves measurement. So, calibration results are a 'snapshot' of what has happened with regards to time, location, environment etc.

Calibration establishes a link to the quantity of interest.



Image: de.wikipedia.org



#### What is the role of a calibration report (DCC)?

It provides information **necessary** to relate measuring system responses to estimates of the measured quantity.

It creates a quantifiable relation using references for that quantity

However, that is not **sufficient** to claim a calibrated system produces traceable measurement results.

#### We also need to:

- 1. quantify factors that influence system performance as used; and
- establish a relationship (model) between indications and measured quantities in the "as used" situation.

This model delivers the measured value and information about the accuracy (uncertainty) of this value as an estimate of the quantity.

#### A calibration report (DCC) is part of the final measurement model

The measuring system may be used in ways and in an environment that are different from where it was calibrated.

Using the system produces **some new**, definite, influence values, but the effects reported during earlier calibrations do not change

A simple example would be if you use a length gauge at a different temperature to the reference temperature in a calibration report.

E.g. the calibration value refers to 20 °C, but it is used at 22 °C and so the user must work with the data in a calibration certificate to obtain a valid traceable measurement at 22 °C.

Another example would be when a system is used to evaluate the difference between two measurements (e.g., uniformity of lighting level).



#### Final remarks

Metrological traceability gives information about the accuracy of a measurement result. This enables reliable **decision** making.

Calibration forges links along a traceability chain.

It establishes a relation between the response of a measuring system and metrological references of the measured quantity, ensuring that traceable measurements are **comparable**.

Information provided by calibration is **necessary** to claim the traceability of measurements using the system, but it is not **sufficient**: modelling of the system **as used** is also required.

Not all the "as used" properties of the system are known at the time of calibration.

The metrologist must develop a suitable model—a calibration report alone is not sufficient.



Back to "Table of Contents" above

# 031 Traceability chains using the DCC

Presenting author: Christian Mueller-Schoell, Mettler-Toledo Int. Inc., Switzerland

Email: <a href="mailto:christian.mueller-schoell@mt.com">christian.mueller-schoell@mt.com</a>

Additional authors:

### Abstract:

It is often claimed that a link (or even a digital hyperlink) in a DCC to the DCC of the higher-level reference standard used would be useful, or the prerequisite, or even the only prerequisite



to prove metrological traceability [1], [2], [3]. This is a false concept. It is based on a very simplified conceptual model of metrological traceability.

Metrological traceability is a property of a measurement (or calibration) result. The measurement result is the outcome of a measuring process, which, in turn (and obviously) has more inputs than just the calibration result of the higher level reference standard from the from the time of its calibration.

There are only a few sources [4], [5] where we can find a list of the necessary contributors to metrological traceability, these are (compiled, re-sorted, and re-worded by the author):

- calibration delivering metrologically traceable calibration results to establish the metrological traceability of all measuring instruments of the measuring system
- defined and reviewed calibration intervals for the measuring instruments (=the calibration program) serving both to establish the metrological traceability by each calibrations, and to maintain the metrological traceability during service life after a calibration,
- o a documented, controlled, proven, accepted measurement procedure,
- o documented measurement uncertainty,
- o competence (not only of individuals, but of the organization in the sense of ISO 17025),
- established measurement process assurance.


From this list, it is obvious that in order to deliver proof of traceability, we need much more than just a link to a higher-level DCC. This complete list is also in principle reflected in ISO 17025 and therefore, a measurement result gained under the scope of a laboratory accreditation according to ISO 17025, can be considered metrologically traceable. However, cancelling just one of the components of the list will breach metrological traceability. In the context of calibrations and DCCs, particular attention must be paid to the second bullet point: Even for all the calibrated equipment used in the process, it is not enough to refer to the metrological properties established at the point of calibration: We need to consider the metrological properties of the equipment at the point and the time of use, after the calibration has taken place. Determining these (as all other points except the first one), is only possible to be done by the user of the equipment and consequently can never be part of a DCC issued by the calibration laboratory.

The reason that this misconception seems to be popular, is probably rooted in an unfortunate phrasing of the definition of metrological traceability in the VIM [6] which puts the focus on calibrations only.

- [1] Oleg Korzhak, Vitalii Malisevych, Taras Kepeshchuk, Denys Serediuk, Volodymyr Hulyk, Digitalization, creation, and implementation of digital calibration certificates. Measurement: Sensors, 2024, 101467, ISSN 2665-9174, https://doi.Org/10.1016/j.measen.2024.101467
- [2] Hackel, S.; Schönhals, S.; Doering, L.; Engel, T.; Baumfalk, R. The Digital Calibration Certificate (DCC) for an End-to-End Digital Quality Infrastructure for Industry 4.0. Sci 2023, 5, 11. https://doi.Org/10.3390/sci5010011
- [3] Thomas Engel, Alexandru Popescu, GEMIMEG -Digitalization of calibration processes in metrology. Measurement: Sensors, 2025, 101474, ISSN 2665-9174, https://doi.Org/10.1016/j.measen.2024.101474.
- [4] NISTIR 6969 Selected Laboratory and Measurement Practices and Procedures to Support Basic Mass Calibrations (2019 Ed), GMP-13
- [5] ILAC, Newton, Australia, ILAC P-10:2002 (sic!)
- [6] BIPM, Paris, VIMA, https://jcgm.bipm.org/vim/en/index.html



# Presentation of Christian Mueller-Schoell



#### What is Metrological Traceability?

METTLER TOLEDO | 2

#### Metrological Traceability:

- Is the property of a measurement (or calibration) result (not of an instrument)
- Is the result/effect of a measurement (or calibration) process
- Is the linking of the result to "a reference" (e.g. an SI unit)
- Is expressed by a result value and the measurement uncertainty
- "L = 1.03 m, where U = 0.14 m (95 %)"
- The uncertainty qualifies the result. It expresses that, with a certain likelihood, the interval given by the uncertainty, embraces the true value of the measurand.



#### What is a metrologically traceable Calibration?

METTLER TOLEDO | 3

- A special kind of measurement process
- Is executed on a measuring instrument
- Establishes metrologically traceable results, at the point of calibration
- "Establishes a relation for obtaining a measurement result from an indication"

#### Attention:

- The results of this process are established at the time and location of the calibration (is a snapshot)
- Rarely does the calibration provider know under which circumstances and how long the calibrated instrument will be used
- The instrument will change its behaviour (performance) after the snapshot (time, location, temperature, gravity, air density, ...)
- The user of the instrument is responsible for taking into account any changes after the calibration

#### Modelling the behaviour after calibration

METTLER TOLEDO |4

 The user of the instrument is responsible for taking into account any changes after the calibration

This can be done in various ways:

- A) assume the calibration values stay constant and increase the uncertainty to account for any changes (e.g. weights according to OIML R111, ...)
- B) model the calibration value to account for any changes and increase the uncertainty by the modelling uncertainty (e.g. high accuracy resistors, Pt -Ir standards after nettoyage-lavage, ...)
- Note, that these values are dependent on the length of the calibration interval, chosen by
- We see: The values from the calibration certificate are not the ones being used during the service life of the instrument



#### What do we need apart from a calibrated instrument?

METTLER TOLEDO | 5


The values from the calibration certificate are not the ones being used during service life...

What is releveant during the service life?

- Prerequisites for a process delivering metrological traceability (similar to ILAC P-10:2002 (sic!))
  - Metrologically traceable calibration of the instrument  $\underline{s}$  (establish the metrological traceability by a calibration) = the calibration certificate, the DCC
  - Modelling the behaviour of the instrument after the calibration
  - defined and reviewed calibration intervals for the measuring instrument  $\underline{s}$  used (=the calibration program),
  - a documented controlled, proven procedure,
  - documented measurement uncertainty,
  - competence,
  - established measurement process assurance.
- 1. The snapshot values gained at calibration are meaningless during service life
- 2. The calibration provider can impossibly deliver the "traceability information" for the service life.

#### The metrologically traceable measurement Process

METTLER TOLEDO | 6






#### The metrologically traceable measurement Process

METTLER TOLEDO | 7

• The metrologically traceable Process is in fact :



#### Conclusions:

- Linking a calibration value to the values of a calibrated reference ALONE is meaningless and useless
- > It conveys a false understanding of how "metrological traceability" is built
- > Traceability is not a simple "chain"

METTLER TOLEDO |8

# Thank you!

Back to "Table of Contents" above

Applying FAIR Principles to Quality Infrastructure Data 032 for Economic Analysis - Approaching Challenges in Data Availability, Accessibility, and Interoperability

Presenting author: Ulrich Harmes-Liedtke

Email: uhl@mesopartner.com

Additional authors: Ann-Sara Ramkissoon Annsara. Ramkissoon@unifi.it



#### Abstract:

Painting a complete picture of the true value of Quality Infrastructure (QI) remains an insurmountable challenge since a significant portion of its value is derived from its contribution to public goods, such as the support it provides in the health, environment, social, and safety sectors, which are difficult to quantify. Easier to quantify is the value of the trade-based economic benefits of QI, such as imports, exports. Gross Domestic Product, etc. However, researchers are hindered by disparate data sources, varying levels of accessibility, and a lack of standardized frameworks that hamper effective data sharing and utilisation. The integration of FAIR (Findability, Accessibility, Interoperability, and Reusability) principles into the management and dissemination of QI data presents a transformative opportunity to address the critical data gaps that currently exist. This paper explores how applying FAIR principles can improve existing QI data repositories and guide the collection and curation of QI data in the coming years as we transition to a digital QI ecosystem.

The FAIR framework aligns with the requirements for reliable, interoperable datasets that underpin the digital transformation of QI services. It offers a roadmap for improving metadata standards, ensuring data traceability, and facilitating the harmonisation of datasets across organisations and borders.

We analyse existing QI data repositories, such as the BIPM Key Comparison Data Base (KCDB), the ISO Survey and initiatives like IAF CertSearch and composite indexes like the Global Quality Infrastructure Index (GQII) and the Quality Infrastructure for Development (QI4SD) Index, highlighting their strengths and limitations in terms of compliance with FAIR principles. Particular emphasis is given to the challenges of metadata standardization, the need for machine-readable formats, and the lack of timeseries data for longitudinal analyses.

In addition, we explore the different QI-digital initiatives, such as digital calibration certificates, smart standards, eAttestation and the digital product passport, applying FAIR principles in our recommendations on collecting and building datasets that facilitate economic analysis.

The paper proposes actionable solutions, including adopting linked data technologies, enhancing cross-institutional collaboration, and developing open-access tools for QI data integration. By showcasing the potential of FAIR-aligned practices, we advocate for a future where QI data is accessible and reliable as we continue our efforts to quantify the value of QI.

**Keywords**: FAIR principles. Quality Infrastructure, QI-Digital, Digital Calibration Certificates, Data Interoperability, Smart standards, composite indexes



# Presentation of Ulrich Harmes-Liedtke



# INTRODUCTION



# Objective:

To establish a causal relationship between QI and development indicators and quantify the benefits of QI.



# Why is this important?

To justify investment in QI development by governments and policymakers. To encourage firms to use Qi services.



# **CHALLENGE**

Gaps in QI data prevent the application of various econometric approaches.

National and international organizations collect data for their own purposes, primarily for documentation and traceability.

Statistical data is a byproduct but is not specifically structured for statistical analysis.

Issues exist with data interoperability, transparency and accessibility.



# FAIR PRINCIPLES: ENHANCING QI DATA COLLECTION AND CURATION FOR FUTURE RESEARCH



# **FAIR Principles**

Findability Accessibility Interoperability Reusability



# Why they matter?

Enables reliable, interoperable datasets.

Enhances data traceability and standardization across different systems.

Facilitates harmonization of Ql and economic datasets.



#### 1.: Administrative Data DCCs: A DATA SOURCE FOR FUTURE 2.: Results of the calibration **QI RESEARCH** • Large amounts of data are international transmitted in DCCs. not regulated recognized Tried and proven 4.: Document Administrative data and calibration results can serve as strong indicators of metrological competency and provide valuable insights for Can be signed cryptographically economic analysis.

## INDICATORS OF METROLOGICAL COMPETENCY FOR ECONOMIC RESEARCH



Source: Hackel, S., Schönhals, S., Doering, L., Engel, T., & Baumfalk, R. (2023). The Digital Colibration Certificate (DCC) for an End-to-End Digital Quality Infrastructure for Industry 4.0. Sci, 5(1), 11. https://doi.org/10.3390/sci5010011

- Total number of calibrations
- Number of accredited calibration labs
- · Level of measurement uncertainty
  - Required data disaggregation:
    - CountryYear

    - Metrology area
    - Measurement instrument
    - Economic sector



# KEY CONSIDERATIONS FOR EXTRACTING, CURATING, AND STORING ECONOMICALLY RELEVANT DATA FROM DCCs



How can economically relevant data be extracted and aggregated from DCCs while maintaining data privacy?



Which institutions or governing bodies could oversee the collection and curation of this data?



Which platforms can be used to store this data securely while ensuring easy access for researchers?



# FUTURE RESEARCH OPPORTUNITIES

Correspondence tables for metrology areas and economic sectors



# FUTURE RESEARCH CONSIDERATIONS


- How can the economic impact and effectiveness of Digital Calibration Certificates (DCCs) be assessed and measured?
- What metrics are being measured, and how frequently are they monitored?
- What data is needed to exploit different econometric approaches to examine the impact of DCCs?



#### **CONCLUSIONS**

- The introduction and implementation of DDCs present an opportunity for QI data collection for economic research, as well as new topics for research.
- When designing the data platform and system, it is essential to consider the statistical use of the collected data, ensuring that it is interoperable.





Back to "Table of Contents" above

# O33 A new data structure for improved scientific data reporting in a FAIR compliant manner

Presenting author: Jon Gabirondo-López; Department of Physics, University of the Basque

Country UPV/EHU

Email: ion.gabirondol@ehu.eus

Additional authors: Inigo Gonzalez de Arrieta (inigo.gonzalezdearrieta@ehu.eus); Inigo Arredondo (inigo.arredondo@ehu.eus); Gabriel A. Lopez (gabrielaleiandro.lopez@ehu.eus); Klaus Anhalt (Klaus.Anhalt@ptb.de); Josu M. Igartua (ion.igartua@ehu.eus)

### Abstract:

Scientific papers often lack effective methods for representing experimental results. Data is typically not provided in machine-readable formats, which hinders its accessibility and reusability. Additionally, the reported experimental information may omit crucial elements such as uncertainty values or details regarding data processing, making it difficult to reproduce results and conduct comparative studies. This issue is particularly acute when comparing measurement results obtained using widely differing methods, as is the case for the authors when determining the thermophysical properties of materials at very high temperatures.

To address these issues, we have developed PER (Framework for Experimental Results), which allows reporting data in a way that ensures its reusability and reproducibility. This framework consists of three main components: a data structure that determines how results are stored, a set of scripts to reproduce the results from the given raw data, and an open database or repository that will store both the data files and the scripts. The first step in developing FER is to define a data structure that ensures compliance with the FAIR principles of the data-reporting process. The presentation of this data structure forms the core of this contribution. Given the emphasis on uncertainty propagation at the core of FER, this data structure has been designed to be consistent with current practices in metrology, adopting a rigorous VIM-compliant nomenclature and allowing measurement uncertainty propagation following all the possibilities allowed by the GUM. Even though the data structure itself is



format-agnostic, we present a use-case using JSON as a working example (accessible at https://github.com/jongablop/fer) and developed a Python library to work with it: ferpy (available at https://github.com/jongablop/ferpy).

Widespread adoption of FER and the creation of domain-specific databases or repositories will ensure complete compliance with the FAIR principles and a more robust and transparent communication of scientific data.

Back to "Table of Contents" above

### Presentation of Jon Gabirondo-López

A new data structure for improved scientific data reporting in a FAIR-compliant manner

Jon Gabirondo-López<sup>1</sup>, I. González de Arrieta<sup>1</sup>, I. Arredondo<sup>2</sup> G. A. López<sup>1</sup>, K. Anhalt<sup>3</sup>, J. M. Igartua<sup>1</sup>

Physics Department, University of the Basque Country UPV/EHU, Spain

<sup>2</sup>Department of Electricity and Electronics, UPV/EHU, Spain

<sup>3</sup>Physikalisch-Technische Bundesanstalt (PTB), Berlin, Germany



ZIENTZIA ETA TEKNOLOGIA FAKULTATEA FAGULTAD DE CIENCIA Y TECNOLOGÍA



#### 1. Background

1. The necessity to standardize emissivity measurements.



### 1. Background

- 1. The necessity to standardize emissivity measurements.
- 2. The **integral upgrade of the HAIRL** emissometer during 2021-2025.

1

#### 2. The framework

#### fer:

Framework for Experimental Results




#### 2. The framework

#### fer:

#### Framework for Experimental Results

Three main elements:




2

#### 2. The framework

#### fer:

Framework for Experimental Results

Three main elements:





#### 2. The framework

#### fer:

#### Framework for Experimental Results

#### Three main elements:

- Data files: to store the results, the metadata and the raw data.
- **Scripts**: written in any open programming language.
- Open databases and repositories.


3

#### 2. The framework

#### fer:

#### Framework for Experimental Results

Three main elements:





#### 3. The data files

- FAIR principles
- Good reporting practicesby IUPAC
- Open Data
- VIM
- GUM

fer data files

5

#### 3. The data files

• fer data files are organized using key-values.



# 3. The data files

- fer data files are organized using key-values.
- They are not limited to a single file format.

6

#### 3. The data files

- fer data files are organized using key-values.
- They are not limited to a single file format.
- The data file consists of **three main** data structures and **five auxiliary** data structures.



#### 4. The data structure

#### Measurement:

- id
- changelog
- description
- correct
- measurands
- state
- results
- correlations
- source

7

#### 4. The data structure

#### Measurement:

- id
- changelog
- description
- correct
- measurands
- state
- results ← [ QuantityValues, QuantityValues, … ]
- correlations
- source



#### QuantityValues:

- id
- name
- description
- changelog
- quantities
- symbols
- units
- values
- standard\_uncertainties
- coverages
- probability\_density\_function
- correlation\_indices

8

#### 4. The data structure

#### Measurement:

- id
- changelog
- description
- correct
- measurands
- state
- results ← [ QuantityValues, QuantityValues, … ]
- correlations
- source



#### Measurement:

- id
- changelog
- description
- correct
- measurands
- state
- results ← [ QuantityValues, QuantityValues, … ]
- correlations
- source ← Source


9

#### 4. The data structure

#### Source:

- id
- name
- description
- model
- influence\_quantities
- input\_quantities
- correlations






11

#### 4. The data structure

#### Source:


- id
- name
- description
- model
- influence\_quantities
- input\_quantities  $\leftarrow$  [ Measurement, Measurement, ... ]
- correlations





13

#### 4. The data structure





#### Working example:

A thermocouple is used to perform a temperature measurement.

A voltmeter is used to measure the electromotive force between its wires, and the Seebeck coefficient is used to retrieve the temperature of the measuring junction.

15


#### 4. The data structure

#### Measurement

- -id
- -changelog
- -description Temperature measurement
- -correct
- -state
- -results
- -correlations
- -measurands
- -source




#### Final results



17


#### 4. The data structure




18



#### 4. The data structure



#### 4. The data structure





#### Measurement

- -id
- -changelog
- -description Measurement of the electromotive force
- -correct
- -state
- -results
- -correlations
- -measurands
- -source

20

#### 4. The data structure

#### Measurement


- -id
- -changelog
- -description Seebeck coefficient derived from the calibration curve
- -correct
- -state
- -results
- -correlations
- -measurands
- -source




#### 

21

#### 4. The data structure







23

#### 4. The data structure

# State: ChangelogEntry:

- name
- description
- quantity\_value

- timestamp
- user
- description

# ProbabilityDensity Correlations: Coverage: Function:

- 1
- name parameters
- values
- quantities
- correlation\_matrix
- method
- intervals
- probabilities
- degrees\_of\_freedom
- method



#### 5. Some implementations



https://github.com/jongablop/fer

JSON schema of the data structure.

Examples.

25

#### 5. Some implementations



https://github.com/jongablop/fer

JSON schema of the data structure.

Examples.



https://github.com/jongablop/ferpy

Python implementation of the data structure.

Easy to integrate.



#### 6. Conclusions

 The proposed data structure aligns with international standards and can be applied across various fields.

26

#### 6. Conclusions

- The proposed data structure aligns with international standards and can be applied across various fields.
- Its Python-based implementation allows **easy integration**.



#### 6. Conclusions

- The proposed data structure aligns with international standards and can be applied across various fields.
- Its Python-based implementation allows easy integration.
- It serves as the basis for open databases.

26

# A new data structure for improved scientific data reporting in a FAIR-compliant manner

<u>Jon Gabirondo-López</u><sup>1</sup>, I. González de Arrieta<sup>1</sup>, I. Arredondo<sup>2</sup> G. A. López<sup>1</sup>, K. Anhalt<sup>3</sup>, J. M. Igartua<sup>1</sup>

jon.gabirondol@ehu.eus







Back to "Table of Contents" above

Day 3 (2025-02-27 Thursday)
Parallel Session "DCR & DRMC"

(Chair: Alexis Valqui)

#### 035 Implementation of Digital Calibration Request, DCR

Presenting author: Heidi Foldal; Novo Nordisk A/S, Novo Alle 1, 2880 Bagsvaerd, Denmark

Email: hefo@novonordisk.com

Additional authors:



#### **Abstract**:

The project team at Novo Nordisk A/S is conducting pilots to implement the Digital Calibration Request (DCR) for digital calibration certificates from suppliers. The company's calibration process is managed through a centralised database where equipment information and calibration requirements are stored. The requirements transferred to the DCR XML files include calibration points, adjustment limits, maximum calibration uncertainty, and calibration methods for example the orientation of measuring equipment during calibration. The project team and IT developers have access to this data, which can be extracted from the database via middleware and used to generate a DCR XML file. Certain necessary data, such as contact information for the employee requesting the calibration, return address for the equipment, and whether the calibration is on-site or at a laboratory is not available in the database and must be manually entered into the middleware during the ordering process. Additionally, standard text that appears in all requests sent to suppliers, such as accreditation requirements according to ISO 17025 or Quality Agreement with Novo Nordisk A/S, is included. Excel is currently being utilised as the middleware tool for the pilot phase. Upon further development of the DCC (Digital Calibration Certificate) and DCR processes, full integration with the database is envisioned. Currently, the project team is mapping data to be implemented in the middleware. It is noted that not all data stored in the database is machine-readable or in a standardised format. This initiative serves an input for future standardisation and digitalisation of Novo Nordisk's work processes. For the execution of pilot projects, Novo Nordisk is collaborating with DFM, the NMI of Denmark, as well as various large and small calibration suppliers.

Back to "Table of Contents" above

#### Presentation of Heidi Foldal





# Implementation of Digital Calibration Request, DCR

Heidi Foldal, The 5 <sup>th</sup> DCC international conference 2025.02.27



2025.02.27 Implementation of Digital Calibration Request, DCR

#### Novo Nordisk Corporate Metrology Process

#### Corporate Metrology Process:

- Ensures reliable and traceable measurements for critical processes
- · Strategic goals:
  - · Remove manual transfer of data
  - · Have FAIR and relevant data for decision -making
- Pilots for implementation of Digital Calibration Certificates (DCC) and Digital Calibration Request (DCR)



2025.02.27 Implementation of Digital Calibration Request, DCR Novo Nordis

# Novo Nordisk Calibration Programme



#### **High level**

- · Define what to measure
- · Identify the requirements for the measurand
- Calibration requirements for the measuring equipment is based on the requirements for the measurand
- Calibration intervals are specified based on a risk assessment assessing impact, probability and mitigation
- If measuring equipment does not comply with requirements Deviations are created to evaluate the use in the entire period back to the last passed calibration
- Outsourcing:
  - Execution of calibration can be outsourced
  - Responsibility of result evaluation cannot be outsourced
  - Conformity statements in calibration certificates are not wanted
- Calibration certificates must be delivered to Novo Nordisk



Calibration Data Flow

- Corporate IT system manages calibration program
  - Centralized database used to records results from Novo Nordisk and supplier calibrations
  - Manual entry of calibration data into the system
  - IT system evaluates if a calibration is pass or fail
  - Performance monitored using dashboards


# Paper and PDF Calibration certificates Manual entry w/ 2nd person review (double swivel chair) DCCs Automatic transfer (Future) DCCs Predictive Metrology DCCs Predictive Metrology DCCs Simply being stored in a database does not guarantee that the data is FAIR.

#### Future goals

- · Predict equipment failure
- Optimize calibration program for efficiency and avoid over

-processing and under -processing

# Digital calibration process - focus on DCR





2025.02.27 Implementation of Digital Calibration Request, DCR Novo

# Why focus on the DCR

- The process commences with a calibration request rather than the issuance of a calibration certificate.
- It's crucial to emphasize that the entire calibration process holds significance, not just the ultimate deliverable, the certificate.
- Suppliers are also eager to digitize and automate, which is a prerequisite for NN to receive DCC.



2025.02.27

Implementation of Digital Calibration Request, DCI

#### What is a DCR

 A DCR is a Digital Calibration Request in an XML format file sent from the customer to the supplier.

#### In the NN pilot project:

- Plan to use an empty DCC and complete it with the necessary information, such as:
  - · Administrative data
  - · Calibration points
  - · Requirements Calibration Method







2025.02.27 Implementation of Digital Calibration Request, DCR Novo Nord

# Pilot project

• The purpose of the current pilot project is to evaluate whether we can successfully pass a DCR through a chain of three stakeholders




# Input sources for the DCR






# Input sources for the DCR




# Input sources for the DCR





2025.02.27 Implementation of Digital Calibration Request, DCR Novo Non

#### Input sources for the DCR



• No conformity statements in the DCC

2025.02.27 Implementation of Digital Calibration Request, DCR

Novo Nordisk

Validation

#### DCR middleware tool

#### Pilot phase:

- An Excel application is used as a pilot tool to generate the DCR
  - Data from ADA database pulled automatedly using the calibration work order
  - Manual data is typed in to a table



Input



2025.02.27 Implementation of Digital Calibration Request, DCR Novo Nordia

#### DCR XML structure

· In creation of a DCR, the structure of the DCC is followed for relevant content:



5 2025.02.27 Implementation of Digital Calibration Request, DCR Novo Nordisk

#### DCR XML structure

- "1" = mandatory
- "0" = not mandatory

```
Person creating the DCR

(1- succeptable provides a specification of the file content as selectable in the dox:transactionContent of the file content as selectable in the dox:transactionCon
```



imponitriamoni o Ergan Generation (100pcs), 5011

# Passing the DCR to Teknologisk Institut

· In the pilot NN will pass the DCR to TI by mail





2025.02.27

molementation of Digital Calibration Request DCF

#### **WHY**

- 20.000 supplier calibrations a year

#### Current status:

- Calibration certificates are analogue and not machine readable or interoperable
  - In average it takes 20 minutes to transfer data from the calibration certificate to the IT system including second person review
- Communication from Novo Nordisk to supplier is manual
  - In average it takes 25 minutes to transfer data from IT system to a letter
  - 5% of all revised calibration certificates are revised due to errors in NN calibration request



- · Trending on calibration data
  - Transparent data insight via the dashboard above
  - Machine Learning under development to trend on data including metadata, and classify between equipment exhibiting systematic drift or not





Back to "Table of Contents" above

# O36 Benefits of developing DCR with a focus on machine readability and automation.

Presenting author: Martin Cradock Østerlund, Teknologisk Institut, Denmark

Email: mard@teknologisk.dk

Additional authors:

#### Abstract:

As with the DCC, the development of a standard for a Digital Calibration Request (DCR) has huge implications for the transfer of calibration data and calibration requests. Moving away from the current standard of paper-based requests allows for automatic validation of calibration parameters, smoother integration into software and quality systems, and allows machines to parse data into a readable format, thus minimizing human error.

Additionally, a common DCR standard has massive implications for the development of fully automated calibration processes. By integrating DCR with calibration software, calibration laboratories have the potential to format a given DCR directly into instructions for calibration software to perform. Coupled with a standard for the DCC, this enables full automation of all steps in the calibration process, including validation of requests, performing the calibration, sending of calibration results and validating the results for the customer.

In this presentation, we discuss our visions for a DCR standard. We discuss what a DCR is, and what sort of information it should include. We will discuss the potential for developing a DCR standard with input from industry partners and how DCR can be integrated into fully automated software solutions. Along with this, we will discuss potential pitfalls that should be considered when developing DCR, with regards to automation and validation.



# Presentation of Martin Cradock Østerlund



# DCC conference 2025

Benefits of developing DCR with a focus on machine readability and automation

# **Topics**

- About me
- What is DCR?
- What are the benefits of DCR?
- Integration into automated software processes
- Steps in parsing a DCR
- Developing DCR with input from industry partners



#### **About me**




#### **Martin Cradock Østerlund**

- Background in physics
- Works with digitalization of DTI's calibration laboratories

Danish Technological Institute

#### What is DCR?

- Digital Calibration Request
- System-to-system communication




Page 456 of 487



#### What are the benefits of DCR?


- Requests can be received faster
- Reduces human errors
- Combined with DCC a shorter round trip time



Danish Technological Institute

# Integration into automated software processes


- Machine readability
- From DCR to software instructions





# Steps in parsing a DCR - Software data

• Characteristics of "good" software data



Danish Technological Institute

# **Example - Calibration of a thermometer**

"The customer instrument is a Platinum Resistance Thermometer, with a nominal resistance at  $0^{\circ}\text{C}$  of  $100\Omega$ "

"The customer instrument is a thermocouple of Type-K.

• Depending on the different types and subtypes, the output changes

```
types = 'thermometer' || 'anemometer' || 'barometer'
thermometerSubTypes = 'prt' || 'thermocouple' || 'thermistor'
```

• Very flexible and also very structured



# **Steps in parsing a DCR - Parsing**

- Taking something with one structure and converting it to our "good" software data.
- · What to consider

#### Input

- What is the input?
- How complex is it?
- What information do we expect it to contain?



#### Output

- · What is the output?
- How structured does it need to be?
- What information do we need to extract?

Danish Technological Institute

# **Steps in parsing a DCR - Parsing**



- Input is ambiguous and unstructured
- Many difficult operations in order to structure properly
- What if string is altered slightly?
- → Parser becomes very complex



# **Steps in parsing a DCR - Parsing**

# An easier example Class Instrument { string Type = 'thermometer'; string Kind = 'pt-100'; } dut: { type: 'thermometer'; parameters: {subType: 'prt'; nominalResistance: 100; }; }

- Input is much more structured
- Still a bit ofwork to do, but it's much more manageable
- → Parser becomes much simpler

#### Key takeaways:

- Keep inputs simple andstrucutred
- Obey our good principles on data structure

Danish Technological Institute

# **Steps in parsing DCR - The DCR**

- Written in XML
- Data storage with tags
- Machine readable
- Structured and with predefined rules









# **Taking inspiration from the DCC**

# 

https://www.gemimeg.ptb.de/gemimeg-tool/#/

Danish Technological Institute

# **Taking inspiration from the DCC**

#### The bad:

• refType and contents are free-text fields...

```
<dcc:measuringEquipments>
<dcc:measuringEquipment refType="Resistance Thermometer made of highly enriched Platinum">
<dcc:measuringEquipment refType="Resistance Thermometer made of highly enriched Platinum">
<dcc:measuringEquipment refType="Resistance Thermometer made of highly enriched Platinum">
<dcc:content lang="de">Platinum Resistance Thermometer mit einem Nennwiderstand von 100 Ohm bei 0 Grad Celsius.</dcc:content>
<dcc:content lang="en">Platinum Resistance Thermometer with a nominal resistance of 100Ohms at 0 degrees celsius.</dcc:content>
</dcc:name></dcc:name></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dcc.uname></dccc.uname></dcc.uname><
```

• Way too many edge cases for parser



# A more sound approach



#### Challenges:

- Less open than the DCC
- Many interested parties, with each their opinions on the format
- Structure must be applied to ALL parts of the DCR!

Danish Technological Institute

# Developing DCR with input from industry partners

- DCC/DCR interpreter tool
- We take feedback from invested industry partners, like Novo Nordisk
- Collaboration and sharing our findings with PTB.







#### **Conclusions**

- If the DCR is to be used for automation purposes, it requires a set of strict and stable rules.
- This allows for simpler parsing into software data, which is easier to maintain.
- XML is well suited for these purposes
- Compared to the DCC, a DCR needs further refinement and structure.
- Requires the format to be less open.











Acknowledgement This study was supported by a grant from Danish Technological Institutes performance contract 2025-2028, entered with the Danish Agency for Higher Education and Science, under The Ministry of Higher Education and Science DenmarkCollaborators are greatly acknowledged. Danish Technological Institute

### Thank you for listening

Martin Cradock Østerlund
Email: mard@teknologisk.dk

Phone number: +45 72202739



Danish Technological Institute

Back to "Table of Contents" above

# 037 The Digital Reference Material Document Derived from the DCC

Presenting author: Michael Melzer Email: michael.melzer@bam.de

Additional authors: Silke Richter, Mehran Monavari, Johannes van de Kreeke (all from Bundesanstalt für Materialforschung und -prüfung (BAM), Unter den Eichen 87, 12205

Berlin)



#### Abstract:

The DCC enables automatization in the metrological traceability of measurements and the dissemination and processing of data from calibration services. Many calibration and measurement procedures, however, rely on certified reference materials (CRMs) to obtain metrological traceability, especially in the broad field of chemical quantities and analytics. The metrological information of those reference materials is also provided in the form of certificates, and hence, their digitization bears the same automatization potential for quality assurance and traceability as the DCC itself.

Although fairly similar, a reference material certificate contains some different information, and has a variable structure and proprietary requirements compared to a calibration certificate, as it relies on the dedicated ISO 33401 standard (formally known as ISO Guide 31) rather than the ISO/IEC 17025. Accordingly, the established DCC-schema does not fully fit the purpose of machine-readable reference material certificates. At the German Federal Institute for Materials Research and Testing (BAM), we started to adapt the DCC-schema in order to derive a dedicated schema for CRM certificates. For the seamless digital dissemination of its certified values, including expanded uncertainties, this schema also utilizes the Digital-SI and thus provides a direct link to the BIPM's digital reference point for a global harmonization of metrological quantities. Furthermore, it should also be suitable for non-certified reference materials to form a common data structure for digital reference material documents (DRMD) in general. As a validated XML report, a DRMD is also compatible with established digital security and verification infrastructure, such as electronic seals and the novel digital accreditation symbol.

In this presentation we will highlight some of the essential differences between reference material documents and calibration certificates, but also point out their similarities, which well justify the derivation of the DRMD-Schema from the DCC. In addition, the first related tools, e.g. a DRMD generator app, the DRMD availability in the COMAR reference material database and an LLM-supported certificate conversion algorithm will be presented briefly.

This work was supported in part by national QI-Digital initiative funded by the German Federal Ministry for Economic Affairs and Climate Action.



#### Presentation of Michael Melzer



Sicherheit in Technik und Chemie

5<sup>th</sup> International DCC Conference

# THE DIGITAL RM DOCUMENT DERIVED FROM THE DCC

Michael Melzer, Mehran Monavari, Silke Richter, Johannes van de Kreeke, Claudia Koch

www.bam.de

# Our mission: safety in technology and chemistry



















The Digital RM Certificate Derived from the DCC



#### **BAM as Reference Material Producer**



- > >100 years of RM production
- ➤ ISO 17034 accredited as RM producer
- Webshop with more than 400 products (webshop.bam.de)
- Recent RM-developments in battery and micro-/nano-materials research




The Digital RM Certificate Derived from the DCC

3

#### The DCC as a Template for Other QI -Documents







#### **DRMD: Two Birds with One Stone**



#### Digital Reference Material Document



#### (Digital) Reference Material Certificate

- → Contains certified RM -values
- → May also contain uncertified RM-values
- → Contains more mandatory details

# (Digital) Product Information Sheet

- → Is similar to RM-certificate
- → Contains no certified RM -values
- → Some details are optional
- → Must not be called "certificate"

The Digital RM Certificate Derived from the DCC


5

BAM

#### DCC vs. DRMD

Uncertainty

# Measurement Uncertainty




 $\mathbf{U}_{\mathrm{DCC}} = \mathbf{U}_{\mathrm{cal}}$   $\mathbf{U}_{\mathrm{ref}}$ 

U<sub>res</sub> book co

U<sub>stab</sub> Long-term behavior to be considered by the DCC user Calibration is valid only for the calibrated item

#### <u>Uncertainty</u>



measurement uncertainty

**U**<sub>stab</sub> uncertainty

of stability

U<sub>hom</sub>
 uncertainty
 of homogenity

https://si-digital-framework.org/

➤ untilRevoked

➤ approvalDate




#### **S** BAM DCC vs. DRMD Dissemination **Issuing Relevant dates** > performanceDate **DCC Kalibration Lab** Customer > receiptDate DCC issuer DCC holder **>** issueDate **DCC** contains customer information **Publishing** Validity **DRMD** is not ▶ validityDate **RM** producer **DRMD** DRMD issuer 99999 customer ➤ timeAfterDispatch

DRMD


RM shop

specific

**Customer** 








#### Automize the monkey work: Databases







The Digital RM Certificate Derived from the DCC



#### **DRMD Generator App**








The Digital RM Certificate Derived from the DCC


11

### Automize the monkey work: Traceability










XML | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 | 01010 |

The Digital RM Certificate Derived from the DCC




### ML-training towards automated spectrum analysis ≥ BAM





- 50 Reference materials
- 80 % reference spectral data for training model
- 20 % reference spectral data for validation




The Digital RM Certificate Derived from the DCC

13

#### Outlook: AI-based DRMD generator





The Digital RM Certificate Derived from the DCC



#### **Acknowledgements and Contact**









Bashar Albakri



The Digital RM Certificate Derived from the DCC

15

Back to "Table of Contents" above

# O38 The Digital NIST: Update on the Digital Transformation of Reference Materials

Presenting author: Dinis Camara, NIST, USA E-mail address: william.camara@nist.gov

Additional authors:


### **Abstract**

At the beginning of 2022, NIST embarked on a pilot project to produce a few examples digital calibration reports and certificates of analysis for the purpose of assessing the scope and challenges of digital transformation in these measurement services. After completion of the pilot project, it was determined that a Digital Reference Material Certificate (DRMC) will require a new model for the data and schema. NIST identified the modelling of Reference Material Data and creation of a database to store the values contained in current certificates as a first step towards digital transformation of Reference Materials. To dive deeper into modelling the types of values, NIST and CENAM collaborated throughout the year. Utilizing the vast scope of over 1,100 NIST Standard Reference Materials (SRM) available, a sample of more than 130 different SRM were reviewed to determine the relationship between the data. The analysis led to a structure able to accommodate the wide variety of values, multipart materials, serialized materials, traceability statements, and other information related to measured values currently expressed in Reference Material Certificates. This talk will give an update on the progress and transition from prototype to production.

Back to "Table of Contents" above



### Presentation of Dinis Camara



# Acknowledgements NIST

#### **Digital NIST Team:**

Angela Lee
Benjamin Long
Robert Hansich
Catherine Rimmer
James Fedchak
Steven Choquette
Catherine Cooksey
Katya Delak
Itzel Domínguez Mendoza (CENAM)

Scientists, Statisticians, and other Subject Matter Experts at NIST

Members of the SIM Digital Reference Material Certificate Subworking Group



### What Goes in a DRMC NIST

#### **Current Draft of Schema**

- Organization Identification
- Reference Material Identification
- Material Information
  - Unit Description
  - Usage Information
  - Measurement Information
  - Additional Information
- Context
- Provenance
- References

#### **Based On:**

ISO 17034

ISO Guide 31/ISO 33401:2024

**NIST Quality Manual** 

**NIST Policy** 

**Customer Input** 

10

## Material Information NIST

- Information about the material and its properties
  - Unit Description
  - Usage Information
  - Measurement Information
  - Additional Information



### **Material Information**

Expiration of Certification: The certification of SRM 2940a is valid, within the measurement uncertainty specified, until 01 September 2028, provided the SRM is handled and stored in accordance with the instructions given here (see "Instructions for Use"). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

#### **Measurement Information**

- Period of Validity
- Traceability
- Measurand
- Values
- Units
- Uncertainties
- Type of value (Certified vs. Non-Certified)
- **Measurement Procedures**

16

## **NIST SRM Overview**

#### >1100 SRM Products



- Ferrous Metals
- Nonferrous Metals
- Microanalysis
- High Purity Materials
- Health and Industrial Hygiene · Radioactivity
- Inorganics
- Primary Gas Mixtures
- Fossil and Alternative Fuels
- Organics
- Food and Agriculture
- Ceramics and Glasses
- Cement
- Engine Wear Materials
- Forensics

- Ion Activity
- Polymeric Properties
- Thermodynamic Properties
- · Optical Properties
- Electrical Properties
- · Metrology, Liquids and Glasses
- · X-Ray Diffraction
- Sizing
- Geological Materials and Ores · Surface Finish
  - · Fire Research
  - · Nanomaterials
  - Miscellaneous Performance **Engineering Materials**

4




# Types of Values NIST



Sequences

Formulas



RM 8671 Heavy Chain Amino Acid Sequence

QVTLRESGPALVKPTQTLTLTCTFSGFSLSTAGMSVGWIRQPPGKALEWLADIWWDDKKHYNPSLKDRLTIS
KDTSKRQVVLKVTHMDPADTATYCARDWIFNFFDVWGGGTTVFVSS ASTKGESVFPLAPSSKSTSGGTA
ALGCLMVSYPFSPUTVSMSGALTSGVHTFPATU,SSGLSTSSSVYTPSSSLGTGTYCKNWHKPSHYDTO
KRV FRSCOKTHTCFPCP APELLGGESVFLFPFKPKDTMISKTEKVTCVVVDVSHEDZEVKFNNYTODG
KVHANTTRPREQINSTKYVSVJVLVLGWUNKKKYKKVSVNKALDAFIEKTISKAKAGGPKEKGVTLPFS
RENTTRQVSLTCLVKGFFFSDIAVSMSSGPENNYKTTPFVLDSDGSFFLYSKLTVDKSKWQGNVFSCS
VHHALTHRYKGLSLEPPK

RM 8671 Light Chain Amino Acid Sequence

DIQMTQSPSTLSASVGDRVTITCSASSRVGYMHWYQQKPGKAPKLLIYDTSKLASGVPSRFSGSGSGTEFTLT ISSLGPDDFATYICTGGSGYPFTFGGGTKVEIK RTVAAESVFIFPFSDEQLKSGTASVVCLLNNFYFREAKV QMKVDHALQSGGSGSVTFTGGGTKVEIKSGTAFTGKAPEKRKVAGEVTHGGLSSFVYKSFRRGEC

Figure A1. Primary amino acid sequence for RM 8671 with variable fragment antigen-binding (Fab) region in normal font, constant Fab region underlined, hinge region in italics, and Fc region in bold.

 $\lambda = -1.97313 \times 10^{-3} + 1.99227 \times 10^{-5} \rho + 1.07923 \times 10^{-4} T \pm 1.0 \% (k = 2)$ 

8

### Data vs. Format NIST

Table 1. Certified Mass Fraction Value for Total Extractable Fat in SRM 2387

Mass Fraction (g/100 g)

Table 2. Certified Mass Fraction Values for Elements in SRM 2387

 Mass Facture law for the first law for the first law factor with law factor law facto

Table 3. Certified Mass Fraction Values for Tocopherols in SRM 2387

RM 8671 Heavy Chain Amino Acid Sequence

QVTLRESGPALVKPTQTLTLTCTFSGFSLSTAGMSVGHIRQPPGKALEWLADIWMDDKKHYNPSLKDRLTIS
KDTSKNQVVLKVTNMDPADTATYYCARDMIFNFYFDVWGQGTTVTVSS ASTKGPSVFPLAFSSKSTSGGTA
ALGCLUKDVFPEPVTVSNNSGALTSGVHIFPAVLQSSGLISLSSVVTVPSSSJGTQTYICHVMHKFSNTKVD
KRV PFKSCHTHTCPEPC ABELLGGSFVFLFPPKPOTTAHISRTEFVTCVVVDVSHEBDEFVKRNYVD
EVHNAKTKPREEQYNSTIKVVSVLTVLHQDWLNGKSYKCKVSNKALDAPIEKTISKAKGQPREPQYTTLPBS
REEMTKGVQSTACLVKGFYSSIAVEWESNGQPENYKTTPPVLDSDGSFFLYSKLTVDKSRWQQGNVFSCS
VMHEALHNHYTQKSLSLSPGK

RM 8671 Light Chain Amino Acid Sequence

DIQMTQSPSTLSASVGDRVTITCSASSRVGYMHWYQQKPGKAPKLLIYDTSKLASGVPSRFSGSGSGTEFTLT ISSLQPDDFATYYCFQGSGYPTFGGGTKVEIK RTVAAPSVFIFPPSDEQLKSGTASVVCLLNNFYPRBAKV QMKVDNALQSGNSQESVTEQDSKDSTYSLSSTLTLSKADYEKHKVYACEVTHQGLSSPVTKSFNRGEC

Figure A1. Primary amino acid sequence for RM 8671 with variable fragment antigen-binding (Fab) region in normal font, constant Fab region underlined, hinge region in italics, and Fc region in bold.


Table 2. Reference Values for SRM 1d Limestone, Argillaceous

Constituent	Mass Fraction	Expanded Uncertainty
	(%)	(%)
C	11:50	0.25 <sup>b</sup>
TiO <sub>2</sub>	0.0306	0.0065 <sup>b</sup>
Cr <sub>2</sub> O <sub>3</sub>	0.0012	0.0002 <sup>8</sup>
BaO	0.0033	0.0011 <sup>b</sup>
Loss on Ignition (1000 °C)	41.57	0.06*

Preparation of Material: All chemicals used in the preparation of this SRM were of the highest purity available and were obtained from a commercial source. The amino acid solution was prepared by weighing the individual amino acids, concentrated HCl, and water and mixing until the amino acids were completely dissolved. The total mass of this solution was measured. The concentration of each amino acid was calculated using the measured density of the Di-mortl. HCl solution at 20°C of 1,00123 gmll. Corrections were made to the calculated amino acid concentrations based on purity as determined by LC with absorbance detection with confirmatory data provided by the manufacturer. The purity of each amino acid was also evaluated by elemental analysis at Galbrath Laboratories (Knowelli-TN). Moisture content for all




# Values and Their Data NIST



# Values and Uncertainties









## Values and Uncertainties

#### NST






# Values and Uncertainties


#### NST














# Multiple Value Types in One Certificate



# **SRM** Configuration


IDENTIFICATION LOT CERTIFICATE GROUPS MIXTURES FILE

SRM number: 2373 SRM Id: 493 SRM Batch:
SRM Name: Genomic DNA Standards for HER2 Measurements

ld Lot Name Nominal Short Title Component A - white SK-BR-3 SK-BR-3 100.00 SRM MDA-MB-231 SRM LOT 1 Component C - red MDA-MB-361 MDA-MB-361 100.00 SRM Component D - yellow MDA-MB-453 MDA-MB-453 100.00 SRM




#### **SRM** Configuration NST IDENTIFICATION LOT CERTIFICATE GROUPS MIXTURES FILE Id Part Id Title Short Title Unit Level 2 Level 2 VALUES SEQUENCES FORMULA # Type Sub Type Operator Value 2 Certified Unknown 2.89 $\mu$ g/L 2034-08-01 SRM Name: Trace Metals in Frozen Human Blood S BASIS CONVERSION SAMPLE SIZE TRACEABILITY CIENT OF VARIATION STATEMENTS METHODS PARTS REFERENCE TIME PROPERTIES REFERENCE Parts:





# احالا Database Creation

- Accommodates Values (Numbers), Sequences, and Formulas
- Accommodates NIST reported values and most of the additional values in certificates
- Created to integrate with current SRM IT applications and DBs
- Additional functionality added that will enhance other SRM applications
- Version 1



### Future Work NIST

- Complete the Documentation of the Model
- · Creation of usable XML Schema from modeling
- Expansion of other Measurement Information like Usage Information
- Best way to sign and secure certificates
  - Centralized Signatory
  - Blockchain
- Continued International Discussions





Back to "Table of Contents" above

### Closing Session & Statistical information:

# 5th International DCC Conference DCC for fair principles in the quality infrastructure

Martin Koval
CIPM - FORUM-MD-TG-H-DCC/DRMC (chair)
Closing session



#### **Final Numbers of the DCC Conference 2025**



45 presentations



760 registered participants



3 panel sessions 8 parallel sessions



2 open exchanges



5th INTERNATIONAL DCC CONFERENCE
DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE

#### A huge Thank You to the Conference Chairs

Session	Chairs		
DAY 1			
Plenary Session	Martin Koval		
Key Aspects of DCC I	Shanna Schönhals		
Applications NMII	Brett Hyland		
DAY 2			
Plenary Session	Hugo Gasca		
Tools & Sofware	Anjali Sharma		
Industry Applications	Carlos Galvan		
Applications NMIII	Jariya Buajarern		
DAY 3			
Plenary Session	Hector Laiz		
Schema & Semantics	Robert Hanisch		
Key Aspects of DCC II	Mark Kuster		
DCR & DRMC	Alexis Valqui		



5th INTERNATIONAL DCC CONFERENCE CC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE



#### A huge Thank You to the Conference Team

Oksana Baer
Franziska Decker
Muhammed Ali Demir
Lutz Doering
Wiebke Heeren
Justin Jagieniak
Moritz Jordan
Shan Lin
Jan Loewe



#### **Announcement of Special Issue in Elsevier Measurement: Digitalization**

- Based on the DCC Conference 2025
- Selected submissions will be invited to submit a paper for the journal
- Members of IPC invited to support review process/act as guest editors
- Open Access, with no publication fees
- Open Call will be issued in March/April

**5TH INTERNATIONAL DCC CONFERENCE** 



#### Hoping to meeting you soon at M4DConf 2025

- Special Session on Calibration and Comparison in a Digital Era
- · Do not hesitate to submit your work to this session





Calibration and Comparison in a Digital Era





5TH INTERNATIONAL DCC CONFERENCE


#### See you at the DCC Conference 2026


- · 6th International DCC Conference
- 24. 26. February 2026
- · Online and free of charge
- · Looking forward to meeting you again!



**5TH INTERNATIONAL DCC CONFERENCE** DCC FOR FAIR PRINCIPLES IN THE QUALITY INFRASTRUCTURE







Back to "Table of Contents" above